

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0030725\
 Data File : P0109735.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Mar 2025 21:12
 Operator : YP/AJ
 Sample : Q1517-01 10X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MOO-25-0062

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/10/2025
 Supervised By :Ankita Jodhani 03/10/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 08 00:07:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0022025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Feb 21 04:40:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.696	3.692	16377981	11327416	1.730	2.164m#
2) SA Decachlor...	8.752	8.704	12048956	4891817	1.401m	1.536

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO030725\
Data File : PO109735.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Mar 2025 21:12
Operator : YP/AJ
Sample : Q1517-01 10X
Misc :
ALS Vial : 32 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
MOO-25-0062


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/10/2025
Supervised By :Ankita Jodhani 03/10/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 08 00:07:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO022025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Feb 21 04:40:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.696 min
Delta R.T.: -0.002 min
Response: 16377981
Conc: 1.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
MOO-25-0062


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/10/2025
Supervised By :Ankita Jodhani 03/10/2025




#1 Tetrachloro-m-xylene

R.T.: 3.692 min
Delta R.T.: -0.002 min
Response: 11327416
Conc: 2.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.752 min
Delta R.T.: -0.005 min
Response: 12048956
Conc: 1.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.704 min
Delta R.T.: -0.003 min
Response: 4891817
Conc: 1.54 ng/ml