

Data Path : P:\HPCHEM1\ECD_0\Data\P0030818\
 Data File : P0042850.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Mar 2018 15:47
 Operator : SM/UA
 Sample : J1723-19MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SP-45MSD

Manual Integrations
 APPROVED

ugo
 3/9/2018 10:47:11 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 09 00:58:49 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 26 14:24:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.095	3.348	4398280	5787382	17.755	15.661
2) SA Decachlor...	9.470	8.126	2483667	3111503	9.827m	8.069
Target Compounds						
3) L1 AR-1016-1	4.958	4.162	1377556	1617548	205.121m	170.882
4) L1 AR-1016-2	5.299	4.598	1647178	1698214	183.939	207.250
5) L1 AR-1016-3	5.391	4.638	1619750	3069258	224.182	295.422 #
6) L1 AR-1016-4	5.814	4.800	1589947	2090571	223.792m	188.252
7) L1 AR-1016-5	6.047	5.138	2836951	4317258	394.085m	408.429m
31) L7 AR-1260-1	6.776	5.792	2960006	3635505	215.156	150.587m#
32) L7 AR-1260-2	7.025	5.979	4831482	5682113	264.135	169.151m#
33) L7 AR-1260-3	7.304	6.122	4994423	4239760	240.365	148.937m#
34) L7 AR-1260-4	7.381	6.294	2218066	5503916	149.858	160.993m
35) L7 AR-1260-5	7.903	6.825	5099878	6690192	150.867	98.857 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\PO030818\
Data File : PO042850.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Mar 2018 15:47
Operator : SM/UA
Sample : J1723-19MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 09 00:58:49 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.095 min
Delta R.T.: -0.007 min
Response: 4398280
Conc: 17.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#1 Tetrachloro-m-xylene

R.T.: 3.348 min
Delta R.T.: 0.000 min
Response: 5787382
Conc: 15.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.470 min
Delta R.T.: -0.014 min
Response: 2483667
Conc: 9.83 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.126 min
Delta R.T.: -0.011 min
Response: 3111503
Conc: 8.07 ng/ml




#3 AR-1016-1

R.T.: 4.958 min
Delta R.T.: -0.010 min
Response: 1377556
Conc: 205.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#3 AR-1016-1

R.T.: 4.162 min
Delta R.T.: -0.009 min
Response: 1617548
Conc: 170.88 ng/ml




#4 AR-1016-2

R.T.: 5.299 min
Delta R.T.: -0.011 min
Response: 1647178
Conc: 183.94 ng/ml



#4 AR-1016-2

R.T.: 4.598 min
Delta R.T.: -0.010 min
Response: 1698214
Conc: 207.25 ng/ml




#5 AR-1016-3

R.T.: 5.391 min
Delta R.T.: -0.013 min
Response: 1619750
Conc: 224.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#5 AR-1016-3

R.T.: 4.638 min
Delta R.T.: -0.008 min
Response: 3069258
Conc: 295.42 ng/ml




#6 AR-1016-4

R.T.: 5.814 min
Delta R.T.: -0.012 min
Response: 1589947
Conc: 223.79 ng/ml m



#6 AR-1016-4

R.T.: 4.800 min
Delta R.T.: -0.011 min
Response: 2090571
Conc: 188.25 ng/ml




#7 AR-1016-5

R.T.: 6.047 min
Delta R.T.: -0.011 min
Response: 2836951
Conc: 394.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#7 AR-1016-5

R.T.: 5.138 min
Delta R.T.: -0.010 min
Response: 4317258
Conc: 408.43 ng/ml m




#31 AR-1260-1

R.T.: 6.776 min
Delta R.T.: -0.014 min
Response: 2960006
Conc: 215.16 ng/ml



#31 AR-1260-1

R.T.: 5.792 min
Delta R.T.: -0.013 min
Response: 3635505
Conc: 150.59 ng/ml m




#32 AR-1260-2

R.T.: 7.025 min
Delta R.T.: -0.015 min
Response: 4831482
Conc: 264.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#32 AR-1260-2

R.T.: 5.979 min
Delta R.T.: -0.012 min
Response: 5682113
Conc: 169.15 ng/ml m




#33 AR-1260-3

R.T.: 7.304 min
Delta R.T.: -0.011 min
Response: 4994423
Conc: 240.36 ng/ml



#33 AR-1260-3

R.T.: 6.122 min
Delta R.T.: -0.013 min
Response: 4239760
Conc: 148.94 ng/ml m




#34 AR-1260-4

R.T.: 7.381 min
Delta R.T.: -0.013 min
Response: 2218066
Conc: 149.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
SP-45MSD


Manual Integrations
APPROVED

ugo
3/9/2018 10:47:11 AM




#34 AR-1260-4

R.T.: 6.294 min
Delta R.T.: -0.014 min
Response: 5503916
Conc: 160.99 ng/ml m



#35 AR-1260-5

R.T.: 7.903 min
Delta R.T.: -0.012 min
Response: 5099878
Conc: 150.87 ng/ml



#35 AR-1260-5

R.T.: 6.825 min
Delta R.T.: -0.012 min
Response: 6690192
Conc: 98.86 ng/ml