

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031119\
 Data File : P0054317.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2019 18:14
 Operator : SM/SJ
 Sample : K1691-05
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CL-02-030819-C

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 00:53:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 11 14:40:27 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.023	3.112	68309276	6691419	19.876	23.893
2) SA Decachlor...	9.332	7.673	20936974	3770184	16.622	13.400

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031119\
Data File : P0054317.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2019 18:14
Operator : SM/SJ
Sample : K1691-05
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-02-030819-C

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 00:53:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031119.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 11 14:40:27 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.023 min
Delta R.T.: 0.002 min
Response: 68309276
Conc: 19.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
CL-02-030819-C




#1 Tetrachloro-m-xylene

R.T.: 3.112 min
Delta R.T.: 0.003 min
Response: 6691419
Conc: 23.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.332 min
Delta R.T.: 0.002 min
Response: 20936974
Conc: 16.62 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.673 min
Delta R.T.: 0.000 min
Response: 3770184
Conc: 13.40 ng/ml