

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031119\
 Data File : P0054333.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2019 22:44
 Operator : SM/SJ
 Sample : K1837-03
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MR-MH-10-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 12 01:02:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031119.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 11 14:40:27 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.021	3.110	98018262	5994020	28.521	21.403
2) SA Decachlor...	9.328	7.671	17574938	4066785	13.953	14.454

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031119\
Data File : P0054333.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2019 22:44
Operator : SM/SJ
Sample : K1837-03
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MR-MH-10-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 12 01:02:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031119.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 11 14:40:27 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.021 min
Delta R.T.: 0.000 min
Response: 98018262
Conc: 28.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
MR-MH-10-COMP




#1 Tetrachloro-m-xylene

R.T.: 3.110 min
Delta R.T.: 0.001 min
Response: 5994020
Conc: 21.40 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.328 min
Delta R.T.: -0.002 min
Response: 17574938
Conc: 13.95 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.671 min
Delta R.T.: 0.000 min
Response: 4066785
Conc: 14.45 ng/ml