

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031120\
 Data File : P0067096.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Mar 2020 10:17
 Operator : DD\AJ
 Sample : AR1260ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 11 11:47:04 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 11 11:46:07 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.114	3.399	1535320	1937953	50.000	50.000
2) SA Decachlor...	9.574	8.274	1785734	2366700	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.260	4.450	576075	662074	500.000	500.000
4) L1 AR-1016-2	5.281	4.467	874604	1219726	500.000	500.000
5) L1 AR-1016-3	5.342	4.638	501917	559974	500.000	500.000
6) L1 AR-1016-4	5.439	4.681	431137	475251	500.000	500.000
7) L1 AR-1016-5	5.728	4.889	419174	585417	500.000	500.000
31) L7 AR-1260-1	6.837	5.903	861635	1152830	500.000	500.000
32) L7 AR-1260-2	7.092	6.091	1281781	1438725	500.000	500.000
33) L7 AR-1260-3	7.447	6.241	1077123	1321246	500.000	500.000
34) L7 AR-1260-4	7.670	6.705	999866	1169409	500.000	500.000
35) L7 AR-1260-5	7.976	6.949	2326676	2816327	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031120\
Data File : P0067096.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Mar 2020 10:17
Operator : DD\AJ
Sample : AR1260ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 11 11:47:04 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031120.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 11 11:46:07 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.114 min
Delta R.T.: 0.000 min
Response: 1535320
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.399 min
Delta R.T.: 0.000 min
Response: 1937953
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.574 min
Delta R.T.: 0.000 min
Response: 1785734
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.274 min
Delta R.T.: 0.000 min
Response: 2366700
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.260 min
Delta R.T.: 0.000 min
Response: 576075
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.450 min
Delta R.T.: 0.000 min
Response: 662074
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.281 min
Delta R.T.: 0.000 min
Response: 874604
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.467 min
Delta R.T.: 0.000 min
Response: 1219726
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.342 min
Delta R.T.: 0.000 min
Response: 501917
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.638 min
Delta R.T.: 0.000 min
Response: 559974
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.439 min
Delta R.T.: 0.000 min
Response: 431137
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.681 min
Delta R.T.: 0.000 min
Response: 475251
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.728 min
Delta R.T.: 0.000 min
Response: 419174
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 4.889 min
Delta R.T.: 0.000 min
Response: 585417
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.837 min
Delta R.T.: 0.000 min
Response: 861635
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 5.903 min
Delta R.T.: 0.000 min
Response: 1152830
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.092 min
Delta R.T.: 0.000 min
Response: 1281781
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.091 min
Delta R.T.: 0.000 min
Response: 1438725
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.447 min
Delta R.T.: 0.000 min
Response: 1077123
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.241 min
Delta R.T.: 0.000 min
Response: 1321246
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.670 min
Delta R.T.: 0.000 min
Response: 999866
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.705 min
Delta R.T.: 0.000 min
Response: 1169409
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.976 min
Delta R.T.: 0.000 min
Response: 2326676
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 6.949 min
Delta R.T.: 0.000 min
Response: 2816327
Conc: 500.00 ng/ml