

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031320\
 Data File : P0067212.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Mar 2020 13:59
 Operator : DD\AJ
 Sample : L1912-05
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 Z3-S1

Manual Integrations
 APPROVED

Sohil
 3/17/2020 2:02:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 13 16:54:15 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 11 15:01:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.111	3.398	721613	938217	23.085m	23.977
2) SA Decachlor...	9.572	8.273	581013	801555	15.995	16.462

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031320\
Data File : P0067212.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Mar 2020 13:59
Operator : DD\AJ
Sample : L1912-05
Misc :
ALS Vial : 12 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
Z3-S1


Manual Integrations
APPROVED

Sohil
3/17/2020 2:02:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 13 16:54:15 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031120.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 11 15:01:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.111 min
Delta R.T.: -0.001 min
Response: 721613
Conc: 23.09 ng/ml m

Instrument :
ECD_O
ClientSampleId :
Z3-S1


Manual Integrations
APPROVED

Sohil
3/17/2020 2:02:30 PM




#1 Tetrachloro-m-xylene

R.T.: 3.398 min
Delta R.T.: 0.001 min
Response: 938217
Conc: 23.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.572 min
Delta R.T.: 0.000 min
Response: 581013
Conc: 16.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.273 min
Delta R.T.: 0.000 min
Response: 801555
Conc: 16.46 ng/ml