

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031420\
 Data File : P0067249.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Mar 2020 11:27
 Operator : DD\AJ
 Sample : L1928-02
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 02-A-02-B-02-C

Manual Integrations
 APPROVED

Sohil
 3/16/2020 4:44:44 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 02:30:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Mar 11 15:01:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.108	3.396	1025388	796944	32.803	20.366 #
2)	SA Decachlor...	9.565	8.269	677496	1189740	18.652m	24.435m#
Target Compounds							
26)	L6 AR-1254-1	6.094	5.229	177228	178857	148.389m	86.767m#
27)	L6 AR-1254-2	6.309	5.375	311327	215330	161.958	117.175 #
28)	L6 AR-1254-3	6.669	5.770	264011	163836	131.061m	54.442m#
29)	L6 AR-1254-4	6.953	5.999	153284	123133	90.029	68.144
30)	L6 AR-1254-5	7.366	6.408	301221	299847	170.999	105.616 #
31)	L7 AR-1260-1	6.830	5.899	240190	244068	135.381	101.745
32)	L7 AR-1260-2	7.085	6.087	317829	343784	121.653	113.931
33)	L7 AR-1260-3	7.439	6.235	143320	373453	65.125	134.393 #
34)	L7 AR-1260-4	7.661	6.700	204875	137084	100.341	55.851 #
35)	L7 AR-1260-5	7.969	6.943	506360	339656	105.650	58.177 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031420\
Data File : PO067249.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Mar 2020 11:27
Operator : DD\AJ
Sample : L1928-02
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 02:30:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031120.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Mar 11 15:01:31 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.108 min
Delta R.T.: -0.004 min
Response: 1025388
Conc: 32.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#1 Tetrachloro-m-xylene

R.T.: 3.396 min
Delta R.T.: 0.000 min
Response: 796944
Conc: 20.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.565 min
Delta R.T.: -0.009 min
Response: 677496
Conc: 18.65 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.269 min
Delta R.T.: -0.004 min
Response: 1189740
Conc: 24.43 ng/ml m




#26 AR-1254-1

R.T.: 6.094 min
Delta R.T.: -0.003 min
Response: 177228
Conc: 148.39 ng/ml m

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#26 AR-1254-1

R.T.: 5.229 min
Delta R.T.: -0.002 min
Response: 178857
Conc: 86.77 ng/ml m




#27 AR-1254-2

R.T.: 6.309 min
Delta R.T.: -0.003 min
Response: 311327
Conc: 161.96 ng/ml



#27 AR-1254-2

R.T.: 5.375 min
Delta R.T.: -0.003 min
Response: 215330
Conc: 117.17 ng/ml




#28 AR-1254-3

R.T.: 6.669 min
Delta R.T.: -0.006 min
Response: 264011
Conc: 131.06 ng/ml m

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#28 AR-1254-3

R.T.: 5.770 min
Delta R.T.: -0.003 min
Response: 163836
Conc: 54.44 ng/ml m




#29 AR-1254-4

R.T.: 6.953 min
Delta R.T.: -0.005 min
Response: 153284
Conc: 90.03 ng/ml



#29 AR-1254-4

R.T.: 5.999 min
Delta R.T.: -0.002 min
Response: 123133
Conc: 68.14 ng/ml




#30 AR-1254-5

R.T.: 7.366 min
Delta R.T.: -0.005 min
Response: 301221
Conc: 171.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#30 AR-1254-5

R.T.: 6.408 min
Delta R.T.: -0.004 min
Response: 299847
Conc: 105.62 ng/ml




#31 AR-1260-1

R.T.: 6.830 min
Delta R.T.: -0.006 min
Response: 240190
Conc: 135.38 ng/ml



#31 AR-1260-1

R.T.: 5.899 min
Delta R.T.: -0.003 min
Response: 244068
Conc: 101.75 ng/ml




#32 AR-1260-2

R.T.: 7.085 min
Delta R.T.: -0.006 min
Response: 317829
Conc: 121.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#32 AR-1260-2

R.T.: 6.087 min
Delta R.T.: -0.003 min
Response: 343784
Conc: 113.93 ng/ml




#33 AR-1260-3

R.T.: 7.439 min
Delta R.T.: -0.007 min
Response: 143320
Conc: 65.12 ng/ml



#33 AR-1260-3

R.T.: 6.235 min
Delta R.T.: -0.004 min
Response: 373453
Conc: 134.39 ng/ml




#34 AR-1260-4

R.T.: 7.661 min
Delta R.T.: -0.009 min
Response: 204875
Conc: 100.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
02-A-02-B-02-C


Manual Integrations
APPROVED

Sohil
3/16/2020 4:44:44 PM




#34 AR-1260-4

R.T.: 6.700 min
Delta R.T.: -0.004 min
Response: 137084
Conc: 55.85 ng/ml



#35 AR-1260-5

R.T.: 7.969 min
Delta R.T.: -0.007 min
Response: 506360
Conc: 105.65 ng/ml



#35 AR-1260-5

R.T.: 6.943 min
Delta R.T.: -0.005 min
Response: 339656
Conc: 58.18 ng/ml