

Data Path : P:\HPCHEM1\ECD_0\Data\P0031518\
Data File : P0043060.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Mar 2018 14:31
Operator : SM/UA
Sample : J1754-09
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-01-031318-E

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 01:19:33 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.086	3.338	5056003	6714408	20.410m	18.170m
2) SA Decachlor...	9.456	8.114	3032046	2997534	11.997m	7.773 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0031518\
Data File : P0043060.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Mar 2018 14:31
Operator : SM/UA
Sample : J1754-09
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
CL-01-031318-E

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 01:19:33 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.086 min
Delta R.T.: -0.015 min
Response: 5056003
Conc: 20.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
CL-01-031318-E




#1 Tetrachloro-m-xylene

R.T.: 3.338 min
Delta R.T.: -0.010 min
Response: 6714408
Conc: 18.17 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.456 min
Delta R.T.: -0.029 min
Response: 3032046
Conc: 12.00 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.114 min
Delta R.T.: -0.023 min
Response: 2997534
Conc: 7.77 ng/ml