

Data Path : P:\HPCHEM1\ECD 0\Data\P0031618\
 Data File : P0043088.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Mar 2018 8:08
 Operator : SM/UA
 Sample : J1911-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-5A-COMP

Manual Integrations
 APPROVED

ugo
 3/16/2018 5:39:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 16 08:27:05 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Feb 26 14:24:10 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.089	3.333	4023601	6149757	16.243	16.642
2) SA Decachlor...	9.455	8.113	3326288	4628255	13.161m	12.002

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\P0031618\
Data File : P0043088.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2018 8:08
Operator : SM/UA
Sample : J1911-02
Misc :
ALS Vial : 4 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-5A-COMP


Manual Integrations
APPROVED

ugo
3/16/2018 5:39:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 16 08:27:05 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.089 min
Delta R.T.: -0.012 min
Response: 4023601
Conc: 16.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-5A-COMP


Manual Integrations
APPROVED

ugo
3/16/2018 5:39:04 PM




#1 Tetrachloro-m-xylene

R.T.: 3.333 min
Delta R.T.: -0.014 min
Response: 6149757
Conc: 16.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.455 min
Delta R.T.: -0.030 min
Response: 3326288
Conc: 13.16 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.113 min
Delta R.T.: -0.024 min
Response: 4628255
Conc: 12.00 ng/ml