

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0031620\
 Data File : P0067267.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 16 Mar 2020 14:29
 Operator : DD\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 17 02:29:17 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Mar 17 02:11:01 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.107	3.397	1121077	631726	50.000	50.000
2) SA Decachlor...	9.559	8.268	1524500	1442710	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.090	5.231	489142	545589	500.000	500.000
27) L6 AR-1254-2	6.306	5.377	804776	510598	500.000	500.000
28) L6 AR-1254-3	6.668	5.773	868257	777293	500.000	500.000
29) L6 AR-1254-4	6.950	6.000	777237	596824	500.000	500.000
30) L6 AR-1254-5	7.364	6.410	792285	729260	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO031620\
Data File : PO067267.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 16 Mar 2020 14:29
Operator : DD\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 17 02:29:17 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO031620.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Mar 17 02:11:01 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.107 min
Delta R.T.: 0.000 min
Response: 1121077
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.397 min
Delta R.T.: 0.000 min
Response: 631726
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.559 min
Delta R.T.: 0.000 min
Response: 1524500
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.268 min
Delta R.T.: 0.000 min
Response: 1442710
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 489142
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 545589
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.306 min
Delta R.T.: 0.000 min
Response: 804776
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.377 min
Delta R.T.: 0.000 min
Response: 510598
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.668 min
Delta R.T.: 0.000 min
Response: 868257
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 5.773 min
Delta R.T.: 0.000 min
Response: 777293
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 6.950 min
Delta R.T.: 0.000 min
Response: 777237
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.000 min
Delta R.T.: 0.000 min
Response: 596824
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.364 min
Delta R.T.: 0.000 min
Response: 792285
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.410 min
Delta R.T.: 0.000 min
Response: 729260
Conc: 500.00 ng/ml