

Data Path : P:\HPCHEM1\ECD 0\Data\P0032018\
Data File : P0043201.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Mar 2018 17:46
Operator : SM/UA
Sample : J1953-09
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
031418-JETT-E-U-M

Manual Integrations
APPROVED

ugo
3/22/2018 2:24:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 05:13:02 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.080	3.334	3648335	4746691	14.728m	12.845
2)	SA Decachlor...	9.442	8.106	3307147	4361532	13.086	11.310m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO032018\
Data File : PO043201.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Mar 2018 17:46
Operator : SM/UA
Sample : J1953-09
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
031418-JETT-E-U-M


Manual Integrations
APPROVED

ugo
3/22/2018 2:24:46 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 21 05:13:02 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO022118.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Feb 26 14:24:10 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.080 min
Delta R.T.: -0.022 min
Response: 3648335
Conc: 14.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
031418-JETT-E-U-M


Manual Integrations
APPROVED

ugo
3/22/2018 2:24:46 PM




#1 Tetrachloro-m-xylene

R.T.: 3.334 min
Delta R.T.: -0.014 min
Response: 4746691
Conc: 12.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.442 min
Delta R.T.: -0.042 min
Response: 3307147
Conc: 13.09 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.106 min
Delta R.T.: -0.031 min
Response: 4361532
Conc: 11.31 ng/ml m