

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032020\
 Data File : P0067363.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 20 Mar 2020 16:08
 Operator : DD\AJ
 Sample : L1762-18MS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TR-05-031920MS

Manual Integrations
 APPROVED

Sohil
 3/24/2020 2:49:55 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 20 17:00:12 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 16:36:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.940	3.952	625255	1041444	29.419m	29.791
2)	SA Decachlor...	10.926	9.248	687828	916954	23.939	23.508
Target Compounds							
3)	L1 AR-1016-1	6.260	5.209	254205	433463	280.926m	283.366m
4)	L1 AR-1016-2	6.284	5.229	309798	303809	248.648	150.592m#
5)	L1 AR-1016-3	6.352	5.420	210198	352224	268.346	315.603m
6)	L1 AR-1016-4	6.457	5.473	199631	331176	313.859	356.077m
7)	L1 AR-1016-5	6.774	5.702	203496	327165	314.945m	275.899
31)	L7 AR-1260-1	7.944	6.799	359359	620832	296.214	289.731
32)	L7 AR-1260-2	8.209	6.996	474885	739382	317.136	279.811m
33)	L7 AR-1260-3	8.572	7.150	298462	781430	258.892	320.201
34)	L7 AR-1260-4	8.802	7.634	338245	500084	251.856	235.600
35)	L7 AR-1260-5	9.125	7.881	662469	1137455	233.133	228.159

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032020\
Data File : P0067363.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 20 Mar 2020 16:08
Operator : DD\AJ
Sample : L1762-18MS
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleID :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 20 17:00:12 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 19 16:36:53 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.940 min
Delta R.T.: 0.000 min
Response: 625255
Conc: 29.42 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#1 Tetrachloro-m-xylene

R.T.: 3.952 min
Delta R.T.: 0.000 min
Response: 1041444
Conc: 29.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.926 min
Delta R.T.: 0.003 min
Response: 687828
Conc: 23.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.248 min
Delta R.T.: 0.000 min
Response: 916954
Conc: 23.51 ng/ml




#3 AR-1016-1

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 254205
Conc: 280.93 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#3 AR-1016-1

R.T.: 5.209 min
Delta R.T.: -0.001 min
Response: 433463
Conc: 283.37 ng/ml m




#4 AR-1016-2

R.T.: 6.284 min
Delta R.T.: 0.000 min
Response: 309798
Conc: 248.65 ng/ml



#4 AR-1016-2

R.T.: 5.229 min
Delta R.T.: -0.001 min
Response: 303809
Conc: 150.59 ng/ml m




#5 AR-1016-3

R.T.: 6.352 min
Delta R.T.: 0.002 min
Response: 210198
Conc: 268.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#5 AR-1016-3

R.T.: 5.420 min
Delta R.T.: -0.002 min
Response: 352224
Conc: 315.60 ng/ml m




#6 AR-1016-4

R.T.: 6.457 min
Delta R.T.: 0.000 min
Response: 199631
Conc: 313.86 ng/ml



#6 AR-1016-4

R.T.: 5.473 min
Delta R.T.: 0.000 min
Response: 331176
Conc: 356.08 ng/ml m




#7 AR-1016-5

R.T.: 6.774 min
Delta R.T.: 0.002 min
Response: 203496
Conc: 314.94 ng/ml m

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#7 AR-1016-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 327165
Conc: 275.90 ng/ml




#31 AR-1260-1

R.T.: 7.944 min
Delta R.T.: 0.001 min
Response: 359359
Conc: 296.21 ng/ml



#31 AR-1260-1

R.T.: 6.799 min
Delta R.T.: -0.001 min
Response: 620832
Conc: 289.73 ng/ml




#32 AR-1260-2

R.T.: 8.209 min
Delta R.T.: 0.000 min
Response: 474885
Conc: 317.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#32 AR-1260-2

R.T.: 6.996 min
Delta R.T.: -0.001 min
Response: 739382
Conc: 279.81 ng/ml m




#33 AR-1260-3

R.T.: 8.572 min
Delta R.T.: 0.000 min
Response: 298462
Conc: 258.89 ng/ml



#33 AR-1260-3

R.T.: 7.150 min
Delta R.T.: -0.001 min
Response: 781430
Conc: 320.20 ng/ml




#34 AR-1260-4

R.T.: 8.802 min
Delta R.T.: 0.000 min
Response: 338245
Conc: 251.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
TR-05-031920MS


Manual Integrations
APPROVED

Sohil
3/24/2020 2:49:55 PM




#34 AR-1260-4

R.T.: 7.634 min
Delta R.T.: 0.000 min
Response: 500084
Conc: 235.60 ng/ml



#35 AR-1260-5

R.T.: 9.125 min
Delta R.T.: 0.002 min
Response: 662469
Conc: 233.13 ng/ml



#35 AR-1260-5

R.T.: 7.881 min
Delta R.T.: 0.000 min
Response: 1137455
Conc: 228.16 ng/ml