

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0032318.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 23 05:22:35 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0043251.D 1000 =P0043249.D 750 =P0043250.D
 250 =P0043252.D 50 =P0043253.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.057	1.944	2.014	2.120	2.460	2.119	E5	9.48
2) SA	Decachlorobiphenyl	2.347	2.141	2.258	2.483	2.914	2.429	E5	12.30
3) L1	AR-1016-1	5.518	4.965	5.307	5.799	6.863	5.690	E3	12.70
4) L1	AR-1016-2	0.831	0.737	0.785	0.890	1.095	0.868	E4	16.05
5) L1	AR-1016-3	6.715	6.069	6.436	7.067	8.480	6.953	E3	13.36
6) L1	AR-1016-4	6.997	6.361	6.724	7.110	8.464	7.131	E3	11.20
7) L1	AR-1016-5	6.034	5.295	8.095	6.112	7.359	6.579	E3	17.12
8) L2	AR-1221-1	3.297					3.297	E3	0.00
9) L2	AR-1221-2	2.415					2.415	E3	0.00
10) L2	AR-1221-3	7.685					7.685	E3	0.00
11) L3	AR-1232-1	4.183					4.183	E3	0.00
12) L3	AR-1232-2	3.235					3.235	E3	0.00
13) L3	AR-1232-3	5.134					5.134	E3	0.00
14) L3	AR-1232-4	3.287					3.287	E3	0.00
15) L3	AR-1232-5	2.450					2.450	E3	0.00
16) L4	AR-1242-1	3.521					3.521	E3	0.00
17) L4	AR-1242-2	8.300					8.300	E3	0.00
18) L4	AR-1242-3	4.327					4.327	E3	0.00
19) L4	AR-1242-4	4.171					4.171	E3	0.00
20) L4	AR-1242-5	4.725					4.725	E3	0.00
21) L5	AR-1248-1	7.534					7.534	E3	0.00
22) L5	AR-1248-2	6.663					6.663	E3	0.00
23) L5	AR-1248-3	9.188					9.188	E3	0.00
24) L5	AR-1248-4	9.237					9.237	E3	0.00
25) L5	AR-1248-5	6.295					6.295	E3	0.00
26) L6	AR-1254-1	1.592					1.592	E4	0.00
27) L6	AR-1254-2	1.031					1.031	E4	0.00
28) L6	AR-1254-3	1.803					1.803	E4	0.00
29) L6	AR-1254-4	1.489					1.489	E4	0.00
30) L6	AR-1254-5	1.089					1.089	E4	0.00
31) L7	AR-1260-1	1.345	1.210	1.285	1.450	1.737	1.405	E4	14.59
32) L7	AR-1260-2	1.734	1.581	1.662	1.845	2.257	1.816	E4	14.60
33) L7	AR-1260-3	1.954	1.788	1.881	2.047	2.493	2.032	E4	13.49
34) L7	AR-1260-4	7.125	6.433	6.824	7.497	8.777	7.331	E3	12.25
35) L7	AR-1260-5	2.948	2.757	2.865	3.073	3.586	3.046	E4	10.62
36) L8	AR-1262-1	1.063					1.063	E4	0.00
37) L8	AR-1262-2	1.347					1.347	E4	0.00
38) L8	AR-1262-3	1.247					1.247	E4	0.00
39) L8	AR-1262-4	1.672					1.672	E4	0.00
40) L8	AR-1262-5	3.403					3.403	E4	0.00
41) L9	AR-1268-1	5.182					5.182	E4	0.00
42) L9	AR-1268-2	4.639					4.639	E4	0.00
43) L9	AR-1268-3	3.932					3.932	E4	0.00
44) L9	AR-1268-4	1.693					1.693	E4	0.00
45) L9	AR-1268-5	1.113					1.113	E5	0.00

Signal #2 Calibration Files

500 =P0043251.D 1000 =P0043249.D 750 =P0043250.D
 250 =P0043252.D 50 =P0043253.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.895	2.719	2.820	2.954	3.246	2.927	E5	6.79
2) SA	Decachlorobiphenyl	3.455	3.310	3.466	3.583	4.066	3.576	E5	8.12

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0032318.M
 Title : GC EXTRACTABLES
 Last Update : Fri Mar 23 05:22:35 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0043251.D 1000 =P0043249.D 750 =P0043250.D
 250 =P0043252.D 50 =P0043253.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	7.782	7.590	7.500	8.150	9.289	8.062	E3	9.05
4)	L1	AR-1016-2	7.173	6.617	6.871	7.496	8.046	7.241	E3	7.70
5)	L1	AR-1016-3	0.879	0.812	0.844	0.911	1.035	0.896	E4	9.59
6)	L1	AR-1016-4	0.956	0.884	0.918	0.999	1.093	0.970	E4	8.35
7)	L1	AR-1016-5	1.036	0.962	0.999	0.998	1.077	1.014	E4	4.32
8)	L2	AR-1221-1	4.310					4.310	E3	0.00
9)	L2	AR-1221-2	3.142					3.142	E3	0.00
10)	L2	AR-1221-3	1.011					1.011	E4	0.00
11)	L3	AR-1232-1	5.406					5.406	E3	0.00
12)	L3	AR-1232-2	1.141					1.141	E4	0.00
13)	L3	AR-1232-3	4.694					4.694	E3	0.00
14)	L3	AR-1232-4	3.504					3.504	E3	0.00
15)	L3	AR-1232-5	2.448					2.448	E3	0.00
16)	L4	AR-1242-1	5.025					5.025	E3	0.00
17)	L4	AR-1242-2	1.872					1.872	E4	0.00
18)	L4	AR-1242-3	5.738					5.738	E3	0.00
19)	L4	AR-1242-4	5.713					5.713	E3	0.00
20)	L4	AR-1242-5	6.210					6.210	E3	0.00
21)	L5	AR-1248-1	1.132					1.132	E4	0.00
22)	L5	AR-1248-2	1.096					1.096	E4	0.00
23)	L5	AR-1248-3	1.745					1.745	E4	0.00
24)	L5	AR-1248-4	1.424					1.424	E4	0.00
25)	L5	AR-1248-5	8.162					8.162	E3	0.00
26)	L6	AR-1254-1	2.148					2.148	E4	0.00
27)	L6	AR-1254-2	1.927					1.927	E4	0.00
28)	L6	AR-1254-3	3.204					3.204	E4	0.00
29)	L6	AR-1254-4	2.651					2.651	E4	0.00
30)	L6	AR-1254-5	1.483					1.483	E4	0.00
31)	L7	AR-1260-1	2.188	2.038	2.114	2.230	2.634	2.241	E4	10.35
32)	L7	AR-1260-2	3.024	2.813	2.909	3.214	4.200	3.232	E4	17.37
33)	L7	AR-1260-3	2.545	2.385	2.472	2.656	3.064	2.624	E4	10.10
34)	L7	AR-1260-4	2.982	2.830	2.911	3.103	3.516	3.068	E4	8.78
35)	L7	AR-1260-5	5.695	5.447	5.595	5.848	6.520	5.821	E4	7.17
36)	L8	AR-1262-1	2.296					2.296	E4	0.00
37)	L8	AR-1262-2	2.091					2.091	E4	0.00
38)	L8	AR-1262-3	3.719					3.719	E4	0.00
39)	L8	AR-1262-4	2.892					2.892	E4	0.00
40)	L8	AR-1262-5	6.374					6.374	E4	0.00
41)	L9	AR-1268-1	9.626					9.626	E4	0.00
42)	L9	AR-1268-2	8.704					8.704	E4	0.00
43)	L9	AR-1268-3	7.359					7.359	E4	0.00
44)	L9	AR-1268-4	2.950					2.950	E4	0.00
45)	L9	AR-1268-5	1.939					1.939	E5	0.00

(#) = Out of Range