

Data Path : P:\HPCHEM1\ECD_0\Data\P0032418\
 Data File : P0043269.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 23 Mar 2018 9:30
 Operator : SM/UA
 Sample : J1995-05
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PIT-A-0-10

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 23 09:51:48 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0032318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Mar 23 05:22:35 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.078	3.332	4253416	5702637	20.073	19.483
2) SA Decachlor...	9.436	8.102	4085597	5818751	16.822	16.271

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0032418\
Data File : P0043269.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 23 Mar 2018 9:30
Operator : SM/UA
Sample : J1995-05
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PIT-A-0-10

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 23 09:51:48 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0032318.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Mar 23 05:22:35 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.078 min
Delta R.T.: 0.000 min
Response: 4253416
Conc: 20.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
PIT-A-0-10




#1 Tetrachloro-m-xylene

R.T.: 3.332 min
Delta R.T.: 0.000 min
Response: 5702637
Conc: 19.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.436 min
Delta R.T.: 0.000 min
Response: 4085597
Conc: 16.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 5818751
Conc: 16.27 ng/ml