

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_0\Data\P0032419\  
 Data File : P0054533.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 24 Mar 2019 11:35  
 Operator : SM/SJ  
 Sample : AR1248ICV500  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 ECD\_0  
 ClientSampleId :  
 ICVPO032319AR1248

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Mar 25 05:23:00 2019  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_0\methods\P0032319.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Mon Mar 25 05:05:57 2019  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|                             | Compound        | RT#1  | RT#2  | Resp#1  | Resp#2  | ng/ml    | ng/ml      |
|-----------------------------|-----------------|-------|-------|---------|---------|----------|------------|
| -----                       |                 |       |       |         |         |          |            |
| System Monitoring Compounds |                 |       |       |         |         |          |            |
| 1)                          | SA Tetrachlo... | 4.313 | 3.627 | 1816493 | 1698248 | 51.992   | 51.250     |
| 2)                          | SA Decachlor... | 9.936 | 8.613 | 2226367 | 1483479 | 47.347   | 44.620     |
| Target Compounds            |                 |       |       |         |         |          |            |
| 3)                          | L1 AR-1016-1    | 5.475 | 4.708 | 559477  | 520493  | 333.695  | 326.459    |
| 4)                          | L1 AR-1016-2    | 5.497 | 4.727 | 661419  | 538477  | 285.102  | 252.457    |
| 5)                          | L1 AR-1016-3    | 5.559 | 4.903 | 352990  | 365497  | 239.934  | 300.413 #  |
| 6)                          | L1 AR-1016-4    | 5.655 | 4.942 | 378329  | 686418  | 311.770  | 696.944 #  |
| 7)                          | L1 AR-1016-5    | 5.950 | 5.156 | 907105  | 941639  | 709.769  | 730.441    |
| 8)                          | L2 AR-1221-1    | 0.000 | 3.839 | 0       | 10334   | N.D.     | 25.125 #   |
| 9)                          | L2 AR-1221-2    | 4.615 | 3.926 | 39757   | 39341   | 118.838  | 125.131    |
| 10)                         | L2 AR-1221-3    | 4.684 | 4.002 | 76788   | 70465   | 74.110   | 73.188     |
| 11)                         | L3 AR-1232-1    | 4.684 | 4.002 | 76788   | 70465   | 93.293   | 93.428     |
| 12)                         | L3 AR-1232-2    | 5.209 | 4.727 | 246674  | 538477  | 538.629  | 640.399    |
| 13)                         | L3 AR-1232-3    | 5.497 | 4.903 | 661419  | 365497  | 728.417  | 778.047    |
| 14)                         | L3 AR-1232-4    | 5.655 | 4.984 | 378329  | 715908  | 812.631  | 1708.592 # |
| 15)                         | L3 AR-1232-5    | 5.747 | 5.156 | 812122  | 941639  | 2158.276 | 2010.648   |
| 16)                         | L4 AR-1242-1    | 5.475 | 4.708 | 559477  | 520493  | 437.431  | 435.274    |
| 17)                         | L4 AR-1242-2    | 5.497 | 4.727 | 661419  | 538477  | 375.091  | 337.789    |
| 18)                         | L4 AR-1242-3    | 5.559 | 4.903 | 352990  | 365497  | 310.280  | 398.582 #  |
| 19)                         | L4 AR-1242-4    | 5.655 | 4.984 | 378329  | 715908  | 405.549  | 785.406 #  |
| 20)                         | L4 AR-1242-5    | 6.389 | 5.506 | 1030803 | 1289902 | 940.575  | 1082.197   |
| 21)                         | L5 AR-1248-1    | 5.475 | 4.708 | 559477  | 520493  | 422.726  | 438.252    |
| 22)                         | L5 AR-1248-2    | 5.747 | 4.942 | 812122  | 686418  | 422.690  | 425.421    |
| 23)                         | L5 AR-1248-3    | 5.950 | 4.984 | 907105  | 715908  | 427.166  | 429.163    |
| 24)                         | L5 AR-1248-4    | 6.351 | 5.156 | 1087729 | 941639  | 432.010  | 461.819    |
| 25)                         | L5 AR-1248-5    | 6.389 | 5.547 | 1030803 | 978363  | 428.197  | 490.902    |
| 26)                         | L6 AR-1254-1    | 6.323 | 5.506 | 968632  | 1289902 | 528.037  | 588.087    |
| 27)                         | L6 AR-1254-2    | 6.545 | 5.652 | 1078207 | 449935  | 372.951  | 231.180 #  |
| 28)                         | L6 AR-1254-3    | 6.906 | 6.055 | 634135  | 624932  | 196.148  | 192.215    |
| 29)                         | L6 AR-1254-4    | 7.191 | 6.283 | 426635  | 426331  | 168.169  | 210.503 #  |
| 30)                         | L6 AR-1254-5    | 7.607 | 6.699 | 106251  | 137220  | 37.614   | 45.670     |
| 31)                         | L7 AR-1260-1    | 7.069 | 6.201 | 89544   | 333755  | 36.431   | 143.048 #  |
| 32)                         | L7 AR-1260-2    | 7.323 | 6.369 | 113258  | 94712   | 37.398   | 33.205     |
| 33)                         | L7 AR-1260-3    | 7.680 | 6.524 | 30965   | 82048   | 13.185   | 30.759 #   |
| 34)                         | L7 AR-1260-4    | 7.896 | 6.995 | 51331   | 17661   | 18.729   | 7.890 #    |
| 35)                         | L7 AR-1260-5    | 8.216 | 7.234 | 33933   | 45365   | 6.298    | 8.950 #    |
| 36)                         | L8 AR-1262-1    | 7.680 | 6.768 | 30965   | 5615    | 8.342    | 1.699 #    |
| 37)                         | L8 AR-1262-2    | 8.216 | 7.234 | 33933   | 45365   | 4.989    | 8.188 #    |
| 38)                         | L8 AR-1262-3    | 0.000 | 7.518 | 0       | 13358   | N.D.     | 5.780 #    |
| 39)                         | L8 AR-1262-4    | 0.000 | 7.581 | 0       | 23608   | N.D.     | 5.754 #    |
| 40)                         | L8 AR-1262-5    | 0.000 | 8.073 | 0       | 24312   | N.D.     | 13.302 #   |
| 41)                         | L9 AR-1268-1    | 0.000 | 7.518 | 0       | 13358   | N.D.     | 2.155 #    |

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0032419\  
Data File : P0054533.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 24 Mar 2019 11:35  
Operator : SM/SJ  
Sample : AR1248ICV500  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 25 05:23:00 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0032319.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Mon Mar 25 05:05:57 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

|        | Compound  | RT#1  | RT#2  | Resp#1 | Resp#2 | ng/ml | ng/ml    |
|--------|-----------|-------|-------|--------|--------|-------|----------|
| 42) L9 | AR-1268-2 | 0.000 | 7.581 | 0      | 23608  | N.D.  | 4.233 #  |
| 43) L9 | AR-1268-3 | 0.000 | 7.790 | 0      | 7625   | N.D.  | 1.650 #  |
| 44) L9 | AR-1268-4 | 0.000 | 8.073 | 0      | 24312  | N.D.  | 12.019 # |
| 45) L9 | AR-1268-5 | 0.000 | 8.359 | 0      | 15689  | N.D.  | 1.293 #  |



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_O\Data\P0032419\  
Data File : P0054533.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 24 Mar 2019 11:35  
Operator : SM/SJ  
Sample : AR1248ICV500  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Mar 25 05:23:00 2019  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_O\methods\P0032319.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Mon Mar 25 05:05:57 2019  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.313 min  
Delta R.T.: 0.006 min  
Response: 1816493  
Conc: 51.99 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.627 min  
Delta R.T.: 0.002 min  
Response: 1698248  
Conc: 51.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.936 min  
Delta R.T.: 0.004 min  
Response: 2226367  
Conc: 47.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.613 min  
Delta R.T.: -0.002 min  
Response: 1483479  
Conc: 44.62 ng/ml



#3 AR-1016-1


R.T.: 5.475 min  
Delta R.T.: 0.005 min  
Response: 559477  
Conc: 333.70 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#3 AR-1016-1

R.T.: 4.708 min  
Delta R.T.: 0.001 min  
Response: 520493  
Conc: 326.46 ng/ml




#4 AR-1016-2

R.T.: 5.497 min  
Delta R.T.: 0.006 min  
Response: 661419  
Conc: 285.10 ng/ml



#4 AR-1016-2


R.T.: 4.727 min  
Delta R.T.: 0.001 min  
Response: 538477  
Conc: 252.46 ng/ml



#5 AR-1016-3


R.T.: 5.559 min  
Delta R.T.: 0.006 min  
Response: 352990  
Conc: 239.93 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#5 AR-1016-3

R.T.: 4.903 min  
Delta R.T.: 0.002 min  
Response: 365497  
Conc: 300.41 ng/ml




#6 AR-1016-4

R.T.: 5.655 min  
Delta R.T.: 0.004 min  
Response: 378329  
Conc: 311.77 ng/ml



#6 AR-1016-4


R.T.: 4.942 min  
Delta R.T.: 0.000 min  
Response: 686418  
Conc: 696.94 ng/ml



#7 AR-1016-5


R.T.: 5.950 min  
Delta R.T.: 0.005 min  
Response: 907105  
Conc: 709.77 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#7 AR-1016-5

R.T.: 5.156 min  
Delta R.T.: 0.002 min  
Response: 941639  
Conc: 730.44 ng/ml




#8 AR-1221-1

R.T.: 0.000 min  
Exp R.T. : 4.515 min  
Response: 0  
Conc: N.D.



#8 AR-1221-1


R.T.: 3.839 min  
Delta R.T.: 0.001 min  
Response: 10334  
Conc: 25.12 ng/ml



#9 AR-1221-2


R.T.: 4.615 min  
Delta R.T.: 0.014 min  
Response: 39757  
Conc: 118.84 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#9 AR-1221-2

R.T.: 3.926 min  
Delta R.T.: 0.003 min  
Response: 39341  
Conc: 125.13 ng/ml




#10 AR-1221-3

R.T.: 4.684 min  
Delta R.T.: 0.007 min  
Response: 76788  
Conc: 74.11 ng/ml



#10 AR-1221-3


R.T.: 4.002 min  
Delta R.T.: 0.003 min  
Response: 70465  
Conc: 73.19 ng/ml



#11 AR-1232-1


R.T.: 4.684 min  
Delta R.T.: 0.007 min  
Response: 76788  
Conc: 93.29 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#11 AR-1232-1

R.T.: 4.002 min  
Delta R.T.: 0.003 min  
Response: 70465  
Conc: 93.43 ng/ml




#12 AR-1232-2

R.T.: 5.209 min  
Delta R.T.: 0.007 min  
Response: 246674  
Conc: 538.63 ng/ml



#12 AR-1232-2


R.T.: 4.727 min  
Delta R.T.: 0.003 min  
Response: 538477  
Conc: 640.40 ng/ml



#13 AR-1232-3


R.T.: 5.497 min  
Delta R.T.: 0.007 min  
Response: 661419  
Conc: 728.42 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#13 AR-1232-3

R.T.: 4.903 min  
Delta R.T.: 0.004 min  
Response: 365497  
Conc: 778.05 ng/ml




#14 AR-1232-4

R.T.: 5.655 min  
Delta R.T.: 0.005 min  
Response: 378329  
Conc: 812.63 ng/ml



#14 AR-1232-4


R.T.: 4.984 min  
Delta R.T.: 0.003 min  
Response: 715908  
Conc: 1708.59 ng/ml



#15 AR-1232-5


R.T.: 5.747 min  
Delta R.T.: 0.007 min  
Response: 812122  
Conc: 2158.28 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#15 AR-1232-5

R.T.: 5.156 min  
Delta R.T.: 0.003 min  
Response: 941639  
Conc: 2010.65 ng/ml




#16 AR-1242-1

R.T.: 5.475 min  
Delta R.T.: 0.007 min  
Response: 559477  
Conc: 437.43 ng/ml



#16 AR-1242-1


R.T.: 4.708 min  
Delta R.T.: 0.003 min  
Response: 520493  
Conc: 435.27 ng/ml



#17 AR-1242-2


R.T.: 5.497 min  
Delta R.T.: 0.006 min  
Response: 661419  
Conc: 375.09 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#17 AR-1242-2

R.T.: 4.727 min  
Delta R.T.: 0.003 min  
Response: 538477  
Conc: 337.79 ng/ml




#18 AR-1242-3

R.T.: 5.559 min  
Delta R.T.: 0.007 min  
Response: 352990  
Conc: 310.28 ng/ml



#18 AR-1242-3


R.T.: 4.903 min  
Delta R.T.: 0.004 min  
Response: 365497  
Conc: 398.58 ng/ml



#19 AR-1242-4


R.T.: 5.655 min  
Delta R.T.: 0.006 min  
Response: 378329  
Conc: 405.55 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#19 AR-1242-4

R.T.: 4.984 min  
Delta R.T.: 0.003 min  
Response: 715908  
Conc: 785.41 ng/ml




#20 AR-1242-5

R.T.: 6.389 min  
Delta R.T.: 0.007 min  
Response: 1030803  
Conc: 940.58 ng/ml



#20 AR-1242-5


R.T.: 5.506 min  
Delta R.T.: 0.003 min  
Response: 1289902  
Conc: 1082.20 ng/ml



#21 AR-1248-1


R.T.: 5.475 min  
Delta R.T.: 0.008 min  
Response: 559477  
Conc: 422.73 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#21 AR-1248-1

R.T.: 4.708 min  
Delta R.T.: 0.003 min  
Response: 520493  
Conc: 438.25 ng/ml




#22 AR-1248-2

R.T.: 5.747 min  
Delta R.T.: 0.007 min  
Response: 812122  
Conc: 422.69 ng/ml



#22 AR-1248-2


R.T.: 4.942 min  
Delta R.T.: 0.003 min  
Response: 686418  
Conc: 425.42 ng/ml



#23 AR-1248-3


R.T.: 5.950 min  
Delta R.T.: 0.007 min  
Response: 907105  
Conc: 427.17 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#23 AR-1248-3

R.T.: 4.984 min  
Delta R.T.: 0.003 min  
Response: 715908  
Conc: 429.16 ng/ml




#24 AR-1248-4

R.T.: 6.351 min  
Delta R.T.: 0.007 min  
Response: 1087729  
Conc: 432.01 ng/ml



#24 AR-1248-4


R.T.: 5.156 min  
Delta R.T.: 0.003 min  
Response: 941639  
Conc: 461.82 ng/ml



#25 AR-1248-5


R.T.: 6.389 min  
Delta R.T.: 0.007 min  
Response: 1030803  
Conc: 428.20 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#25 AR-1248-5

R.T.: 5.547 min  
Delta R.T.: 0.003 min  
Response: 978363  
Conc: 490.90 ng/ml




#26 AR-1254-1

R.T.: 6.323 min  
Delta R.T.: 0.006 min  
Response: 968632  
Conc: 528.04 ng/ml



#26 AR-1254-1


R.T.: 5.506 min  
Delta R.T.: 0.002 min  
Response: 1289902  
Conc: 588.09 ng/ml



#27 AR-1254-2


R.T.: 6.545 min  
Delta R.T.: 0.010 min  
Response: 1078207  
Conc: 372.95 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#27 AR-1254-2

R.T.: 5.652 min  
Delta R.T.: 0.003 min  
Response: 449935  
Conc: 231.18 ng/ml




#28 AR-1254-3

R.T.: 6.906 min  
Delta R.T.: 0.006 min  
Response: 634135  
Conc: 196.15 ng/ml



#28 AR-1254-3


R.T.: 6.055 min  
Delta R.T.: 0.002 min  
Response: 624932  
Conc: 192.22 ng/ml



#29 AR-1254-4


R.T.: 7.191 min  
Delta R.T.: 0.006 min  
Response: 426635  
Conc: 168.17 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#29 AR-1254-4

R.T.: 6.283 min  
Delta R.T.: 0.003 min  
Response: 426331  
Conc: 210.50 ng/ml




#30 AR-1254-5

R.T.: 7.607 min  
Delta R.T.: 0.005 min  
Response: 106251  
Conc: 37.61 ng/ml



#30 AR-1254-5


R.T.: 6.699 min  
Delta R.T.: 0.002 min  
Response: 137220  
Conc: 45.67 ng/ml



#31 AR-1260-1


R.T.: 7.069 min  
Delta R.T.: 0.005 min  
Response: 89544  
Conc: 36.43 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#31 AR-1260-1

R.T.: 6.201 min  
Delta R.T.: 0.016 min  
Response: 333755  
Conc: 143.05 ng/ml




#32 AR-1260-2

R.T.: 7.323 min  
Delta R.T.: 0.002 min  
Response: 113258  
Conc: 37.40 ng/ml



#32 AR-1260-2


R.T.: 6.369 min  
Delta R.T.: 0.000 min  
Response: 94712  
Conc: 33.21 ng/ml



#33 AR-1260-3


R.T.: 7.680 min  
Delta R.T.: 0.001 min  
Response: 30965  
Conc: 13.18 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#33 AR-1260-3

R.T.: 6.524 min  
Delta R.T.: 0.000 min  
Response: 82048  
Conc: 30.76 ng/ml




#34 AR-1260-4

R.T.: 7.896 min  
Delta R.T.: -0.007 min  
Response: 51331  
Conc: 18.73 ng/ml



#34 AR-1260-4


R.T.: 6.995 min  
Delta R.T.: 0.000 min  
Response: 17661  
Conc: 7.89 ng/ml



#35 AR-1260-5


R.T.: 8.216 min  
Delta R.T.: 0.003 min  
Response: 33933  
Conc: 6.30 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#35 AR-1260-5

R.T.: 7.234 min  
Delta R.T.: 0.000 min  
Response: 45365  
Conc: 8.95 ng/ml




#36 AR-1262-1

R.T.: 7.680 min  
Delta R.T.: 0.003 min  
Response: 30965  
Conc: 8.34 ng/ml



#36 AR-1262-1


R.T.: 6.768 min  
Delta R.T.: 0.034 min  
Response: 5615  
Conc: 1.70 ng/ml



#37 AR-1262-2


R.T.: 8.216 min  
Delta R.T.: 0.005 min  
Response: 33933  
Conc: 4.99 ng/ml

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#37 AR-1262-2

R.T.: 7.234 min  
Delta R.T.: 0.000 min  
Response: 45365  
Conc: 8.19 ng/ml




#38 AR-1262-3

R.T.: 0.000 min  
Exp R.T. : 8.510 min  
Response: 0  
Conc: N.D.



#38 AR-1262-3


R.T.: 7.518 min  
Delta R.T.: 0.001 min  
Response: 13358  
Conc: 5.78 ng/ml



#39 AR-1262-4


R.T.: 0.000 min  
Exp R.T.: 8.595 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#39 AR-1262-4

R.T.: 7.581 min  
Delta R.T.: 0.001 min  
Response: 23608  
Conc: 5.75 ng/ml




#40 AR-1262-5

R.T.: 0.000 min  
Exp R.T.: 9.220 min  
Response: 0  
Conc: N.D.



#40 AR-1262-5


R.T.: 8.073 min  
Delta R.T.: 0.001 min  
Response: 24312  
Conc: 13.30 ng/ml



#41 AR-1268-1


R.T.: 0.000 min  
Exp R.T.: 8.508 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#41 AR-1268-1

R.T.: 7.518 min  
Delta R.T.: 0.001 min  
Response: 13358  
Conc: 2.16 ng/ml




#42 AR-1268-2

R.T.: 0.000 min  
Exp R.T.: 8.596 min  
Response: 0  
Conc: N.D.



#42 AR-1268-2


R.T.: 7.581 min  
Delta R.T.: 0.000 min  
Response: 23608  
Conc: 4.23 ng/ml



#43 AR-1268-3


R.T.: 0.000 min  
Exp R.T. : 8.814 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248




#43 AR-1268-3

R.T.: 7.790 min  
Delta R.T.: 0.001 min  
Response: 7625  
Conc: 1.65 ng/ml




#44 AR-1268-4

R.T.: 0.000 min  
Exp R.T. : 9.221 min  
Response: 0  
Conc: N.D.



#44 AR-1268-4


R.T.: 8.073 min  
Delta R.T.: 0.002 min  
Response: 24312  
Conc: 12.02 ng/ml



#45 AR-1268-5

R.T.: 0.000 min  
Exp R.T.: 9.611 min  
Response: 0  
Conc: N.D.

Instrument :  
ECD\_O  
ClientSampleId :  
ICVPO032319AR1248



#45 AR-1268-5

R.T.: 8.359 min  
Delta R.T.: 0.000 min  
Response: 15689  
Conc: 1.29 ng/ml