

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032419\
 Data File : P0054550.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Mar 2019 16:24
 Operator : SM/SJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Manual Integrations
 APPROVED

Sohil
 3/26/2019 1:41:11 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 05:32:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 05:05:57 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.310	3.627	1963792	1808263	56.208	54.570
2)	SA Decachlor...	9.932	8.612	2326246	1574316	49.471	47.352
Target Compounds							
21)	L5 AR-1248-1	5.471	4.707	590236	532671	445.966	448.506
22)	L5 AR-1248-2	5.743	4.942	853885	706157	444.427	437.654
23)	L5 AR-1248-3	5.946	4.984	957650	735478	450.968	440.894
24)	L5 AR-1248-4	6.348	5.155	1161116	920178	461.157	451.294m
25)	L5 AR-1248-5	6.385	5.546	1111850	999944	461.864m	501.731

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032419\
Data File : P0054550.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Mar 2019 16:24
Operator : SM/SJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Sohil
3/26/2019 1:41:11 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 05:32:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032319.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 05:05:57 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.310 min
Delta R.T.: 0.003 min
Response: 1963792
Conc: 56.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Sohil
3/26/2019 1:41:11 PM




#1 Tetrachloro-m-xylene

R.T.: 3.627 min
Delta R.T.: 0.002 min
Response: 1808263
Conc: 54.57 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.932 min
Delta R.T.: 0.000 min
Response: 2326246
Conc: 49.47 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.612 min
Delta R.T.: -0.002 min
Response: 1574316
Conc: 47.35 ng/ml




#21 AR-1248-1

R.T.: 5.471 min
Delta R.T.: 0.004 min
Response: 590236
Conc: 445.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Sohil
3/26/2019 1:41:11 PM




#21 AR-1248-1

R.T.: 4.707 min
Delta R.T.: 0.003 min
Response: 532671
Conc: 448.51 ng/ml




#22 AR-1248-2

R.T.: 5.743 min
Delta R.T.: 0.003 min
Response: 853885
Conc: 444.43 ng/ml



#22 AR-1248-2

R.T.: 4.942 min
Delta R.T.: 0.003 min
Response: 706157
Conc: 437.65 ng/ml




#23 AR-1248-3

R.T.: 5.946 min
Delta R.T.: 0.004 min
Response: 957650
Conc: 450.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500


Manual Integrations
APPROVED

Sohil
3/26/2019 1:41:11 PM




#23 AR-1248-3

R.T.: 4.984 min
Delta R.T.: 0.002 min
Response: 735478
Conc: 440.89 ng/ml




#24 AR-1248-4

R.T.: 6.348 min
Delta R.T.: 0.004 min
Response: 1161116
Conc: 461.16 ng/ml



#24 AR-1248-4

R.T.: 5.155 min
Delta R.T.: 0.002 min
Response: 920178
Conc: 451.29 ng/ml m




#25 AR-1248-5

R.T.: 6.385 min
Delta R.T.: 0.003 min
Response: 1111850
Conc: 461.86 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Manual Integrations
APPROVED

Sohil
3/26/2019 1:41:11 PM



#25 AR-1248-5

R.T.: 5.546 min
Delta R.T.: 0.002 min
Response: 999944
Conc: 501.73 ng/ml