

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032519\
 Data File : P0054575.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 25 Mar 2019 10:45
 Operator : SM/SJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 25 12:02:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 12:00:46 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.305	3.622	2662961	2396443	78.439	77.749
2)	SA Decachlor...	9.925	8.606	3518287	2160902	73.558	71.595
Target Compounds							
3)	L1 AR-1016-1	5.467	4.702	1291952	1115406	747.136	740.235
4)	L1 AR-1016-2	5.489	4.722	1815302	1531793	761.444	753.103
5)	L1 AR-1016-3	5.551	4.897	1129038	859067	745.907	737.369
6)	L1 AR-1016-4	5.648	4.937	954999	704129	749.378	729.060
7)	L1 AR-1016-5	5.941	5.150	972684	914204	739.174	730.186
31)	L7 AR-1260-1	7.061	6.178	1779261	1577426	732.819	725.194
32)	L7 AR-1260-2	7.317	6.365	2201182	1884136	739.502	731.559
33)	L7 AR-1260-3	7.675	6.519	1763768	1775990	744.602	722.328
34)	L7 AR-1260-4	7.900	6.989	1996299	1447191	735.136	711.136
35)	L7 AR-1260-5	8.210	7.228	4059532	3329025	759.176	728.108



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032519\
Data File : P0054575.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 25 Mar 2019 10:45
Operator : SM/SJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 25 12:02:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 12:00:46 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.305 min
Delta R.T.: 0.000 min
Response: 2662961
Conc: 78.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: 0.000 min
Response: 2396443
Conc: 77.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.925 min
Delta R.T.: -0.001 min
Response: 3518287
Conc: 73.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.606 min
Delta R.T.: 0.000 min
Response: 2160902
Conc: 71.60 ng/ml



#3 AR-1016-1


R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 1291952
Conc: 747.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.702 min
Delta R.T.: 0.001 min
Response: 1115406
Conc: 740.23 ng/ml




#4 AR-1016-2

R.T.: 5.489 min
Delta R.T.: 0.000 min
Response: 1815302
Conc: 761.44 ng/ml



#4 AR-1016-2


R.T.: 4.722 min
Delta R.T.: 0.000 min
Response: 1531793
Conc: 753.10 ng/ml



#5 AR-1016-3


R.T.: 5.551 min
Delta R.T.: 0.000 min
Response: 1129038
Conc: 745.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 859067
Conc: 737.37 ng/ml




#6 AR-1016-4

R.T.: 5.648 min
Delta R.T.: 0.000 min
Response: 954999
Conc: 749.38 ng/ml



#6 AR-1016-4


R.T.: 4.937 min
Delta R.T.: 0.000 min
Response: 704129
Conc: 729.06 ng/ml



#7 AR-1016-5


R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 972684
Conc: 739.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.150 min
Delta R.T.: 0.001 min
Response: 914204
Conc: 730.19 ng/ml




#31 AR-1260-1

R.T.: 7.061 min
Delta R.T.: 0.000 min
Response: 1779261
Conc: 732.82 ng/ml



#31 AR-1260-1


R.T.: 6.178 min
Delta R.T.: 0.000 min
Response: 1577426
Conc: 725.19 ng/ml



#32 AR-1260-2


R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 2201182
Conc: 739.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.365 min
Delta R.T.: 0.000 min
Response: 1884136
Conc: 731.56 ng/ml




#33 AR-1260-3

R.T.: 7.675 min
Delta R.T.: 0.000 min
Response: 1763768
Conc: 744.60 ng/ml



#33 AR-1260-3


R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 1775990
Conc: 722.33 ng/ml



#34 AR-1260-4


R.T.: 7.900 min
Delta R.T.: 0.000 min
Response: 1996299
Conc: 735.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 6.989 min
Delta R.T.: 0.000 min
Response: 1447191
Conc: 711.14 ng/ml



#35 AR-1260-5

R.T.: 8.210 min
Delta R.T.: 0.000 min
Response: 4059532
Conc: 759.18 ng/ml



#35 AR-1260-5

R.T.: 7.228 min
Delta R.T.: 0.000 min
Response: 3329025
Conc: 728.11 ng/ml