

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032620\
 Data File : P0067452.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Mar 2020 14:00
 Operator : DD\AJ
 Sample : PB127766BSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB127766BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 26 16:48:11 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 16:36:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.942	3.950	544800	846351	25.633	24.210
2) SA Decachlor...	10.926	9.243	612854	742602	21.329	19.038
Target Compounds						
3) L1 AR-1016-1	6.261	5.207	225388	355974	249.080	232.710
4) L1 AR-1016-2	6.285	5.228	310170	462936	248.946	229.468
5) L1 AR-1016-3	6.351	5.418	194998	246416	248.942	220.796
6) L1 AR-1016-4	6.458	5.471	159146	204526	250.208	219.905
7) L1 AR-1016-5	6.772	5.698	159233	258494	246.440	217.989
31) L7 AR-1260-1	7.944	6.795	297903	472564	245.556	220.537
32) L7 AR-1260-2	8.209	6.993	361659	577248	241.522	218.453
33) L7 AR-1260-3	8.572	7.147	224921	536437	195.101	219.812
34) L7 AR-1260-4	8.801	7.630	271736	385780	202.334	181.749
35) L7 AR-1260-5	9.125	7.877	539912	875332	190.003	175.581



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032620\
Data File : P0067452.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Mar 2020 14:00
Operator : DD\AJ
Sample : PB127766BSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB127766BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 26 16:48:11 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 19 16:36:53 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.942 min
Delta R.T.: 0.002 min
Response: 544800
Conc: 25.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#1 Tetrachloro-m-xylene

R.T.: 3.950 min
Delta R.T.: -0.001 min
Response: 846351
Conc: 24.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.926 min
Delta R.T.: 0.003 min
Response: 612854
Conc: 21.33 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.243 min
Delta R.T.: -0.005 min
Response: 742602
Conc: 19.04 ng/ml



#3 AR-1016-1


R.T.: 6.261 min
Delta R.T.: 0.001 min
Response: 225388
Conc: 249.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#3 AR-1016-1

R.T.: 5.207 min
Delta R.T.: -0.003 min
Response: 355974
Conc: 232.71 ng/ml




#4 AR-1016-2

R.T.: 6.285 min
Delta R.T.: 0.002 min
Response: 310170
Conc: 248.95 ng/ml



#4 AR-1016-2


R.T.: 5.228 min
Delta R.T.: -0.003 min
Response: 462936
Conc: 229.47 ng/ml



#5 AR-1016-3


R.T.: 6.351 min
Delta R.T.: 0.001 min
Response: 194998
Conc: 248.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#5 AR-1016-3

R.T.: 5.418 min
Delta R.T.: -0.004 min
Response: 246416
Conc: 220.80 ng/ml




#6 AR-1016-4

R.T.: 6.458 min
Delta R.T.: 0.000 min
Response: 159146
Conc: 250.21 ng/ml



#6 AR-1016-4


R.T.: 5.471 min
Delta R.T.: -0.003 min
Response: 204526
Conc: 219.90 ng/ml



#7 AR-1016-5


R.T.: 6.772 min
Delta R.T.: 0.001 min
Response: 159233
Conc: 246.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#7 AR-1016-5

R.T.: 5.698 min
Delta R.T.: -0.003 min
Response: 258494
Conc: 217.99 ng/ml




#31 AR-1260-1

R.T.: 7.944 min
Delta R.T.: 0.000 min
Response: 297903
Conc: 245.56 ng/ml



#31 AR-1260-1


R.T.: 6.795 min
Delta R.T.: -0.005 min
Response: 472564
Conc: 220.54 ng/ml



#32 AR-1260-2


R.T.: 8.209 min
Delta R.T.: 0.000 min
Response: 361659
Conc: 241.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#32 AR-1260-2

R.T.: 6.993 min
Delta R.T.: -0.005 min
Response: 577248
Conc: 218.45 ng/ml




#33 AR-1260-3

R.T.: 8.572 min
Delta R.T.: 0.000 min
Response: 224921
Conc: 195.10 ng/ml



#33 AR-1260-3


R.T.: 7.147 min
Delta R.T.: -0.005 min
Response: 536437
Conc: 219.81 ng/ml



#34 AR-1260-4


R.T.: 8.801 min
Delta R.T.: 0.000 min
Response: 271736
Conc: 202.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127766BSD




#34 AR-1260-4

R.T.: 7.630 min
Delta R.T.: -0.005 min
Response: 385780
Conc: 181.75 ng/ml



#35 AR-1260-5

R.T.: 9.125 min
Delta R.T.: 0.002 min
Response: 539912
Conc: 190.00 ng/ml



#35 AR-1260-5

R.T.: 7.877 min
Delta R.T.: -0.004 min
Response: 875332
Conc: 175.58 ng/ml