

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032719\
 Data File : P0054661.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2019 15:20
 Operator : SM/SJ
 Sample : PB118307BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB118307BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 00:47:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.305	3.620	500035	455462	14.408	14.272
2)	SA Decachlor...	9.925	8.598	768579	533047	15.591	16.737
Target Compounds							
3)	L1 AR-1016-1	5.467	4.699	286553	265311	159.246	167.770
4)	L1 AR-1016-2	5.490	4.718	393044	353541	158.251	164.995
5)	L1 AR-1016-3	5.551	4.894	257554	202878	163.708	167.914
6)	L1 AR-1016-4	5.648	4.934	211846	167570	162.574	165.824
7)	L1 AR-1016-5	5.941	5.146	223968	223632	167.860	171.577
8)	L2 AR-1221-1	0.000	3.832	0	19253	N.D.	49.519 #
9)	L2 AR-1221-2	4.600	3.918	32731	28509	96.519	95.097
10)	L2 AR-1221-3	4.675	3.994	126163	110609	121.454	119.995
11)	L3 AR-1232-1	4.675	3.994	126163	110609	146.537	147.500
12)	L3 AR-1232-2	5.201	4.718	192041	353541	392.610	410.567
13)	L3 AR-1232-3	5.490	4.894	393044	202878	402.172	418.681
14)	L3 AR-1232-4	5.648	4.975	211846	207975	413.881	473.314
15)	L3 AR-1232-5	5.739	5.146	196565	223632	472.516	459.809
16)	L4 AR-1242-1	5.467	4.699	286553	265311	206.258	216.102
17)	L4 AR-1242-2	5.490	4.718	393044	353541	205.409	213.334
18)	L4 AR-1242-3	5.551	4.894	257554	202878	207.796	214.157
19)	L4 AR-1242-4	5.648	4.975	211846	207975	207.785	217.603
20)	L4 AR-1242-5	6.381	5.497	33571	188307	27.810	153.108 #
21)	L5 AR-1248-1	5.467	4.699	286553	265311	286.645	302.254
22)	L5 AR-1248-2	5.739	4.934	196565	167570	135.347	138.654
23)	L5 AR-1248-3	5.941	4.975	223968	207975	140.229	166.674
24)	L5 AR-1248-4	6.342	5.146	35649	223632	18.977	146.839 #
25)	L5 AR-1248-5	6.381	5.538	33571	31770	18.725	21.310
26)	L6 AR-1254-1	6.316	5.497	179941	188307	99.240	86.163
27)	L6 AR-1254-2	6.533	5.642	202991	183885	71.107	94.837 #
28)	L6 AR-1254-3	6.898	6.060	117662	329493	37.416	102.784 #
29)	L6 AR-1254-4	7.182	6.272	76485	38124	31.265	19.471 #
30)	L6 AR-1254-5	7.599	6.688	707522	622380	263.971	216.502
31)	L7 AR-1260-1	7.061	6.174	487372	455337	193.968	198.139
32)	L7 AR-1260-2	7.317	6.360	585848	550602	188.768	200.018
33)	L7 AR-1260-3	7.675	6.515	378871	511199	154.366	198.300 #
34)	L7 AR-1260-4	7.899	6.985	466663	365223	168.713	171.614
35)	L7 AR-1260-5	8.208	7.223	840892	799650	151.550	166.966
36)	L8 AR-1262-1	7.675	6.725	378871	396689	113.704	133.343
37)	L8 AR-1262-2	8.208	7.223	840892	799650	141.937	163.156
38)	L8 AR-1262-3	8.521	7.506	542757	151427	131.012	73.742 #
39)	L8 AR-1262-4	8.573	7.569	289126	548304	89.736	151.376 #
40)	L8 AR-1262-5	9.217	8.062	206383	157446	94.400	97.557
41)	L9 AR-1268-1	8.521	7.506	542757	151427	76.257	26.781 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032719\
 Data File : P0054661.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2019 15:20
 Operator : SM/SJ
 Sample : PB118307BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB118307BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 00:47:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.573	7.569	289126	548304	44.128	106.022 #
43) L9	AR-1268-3	0.000	7.779	0	15180	N.D.	3.527 #
44) L9	AR-1268-4	9.217	8.062	206383	157446	86.505	84.929
45) L9	AR-1268-5	9.606	8.347	62280	50343	3.780	4.423



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032719\
Data File : P0054661.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Mar 2019 15:20
Operator : SM/SJ
Sample : PB118307BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB118307BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 00:47:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.305 min
Delta R.T.: 0.002 min
Response: 500035
Conc: 14.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: -0.001 min
Response: 455462
Conc: 14.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.925 min
Delta R.T.: 0.002 min
Response: 768579
Conc: 15.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.598 min
Delta R.T.: -0.006 min
Response: 533047
Conc: 16.74 ng/ml



#3 AR-1016-1


R.T.: 5.467 min
Delta R.T.: 0.002 min
Response: 286553
Conc: 159.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#3 AR-1016-1

R.T.: 4.699 min
Delta R.T.: -0.002 min
Response: 265311
Conc: 167.77 ng/ml




#4 AR-1016-2

R.T.: 5.490 min
Delta R.T.: 0.003 min
Response: 393044
Conc: 158.25 ng/ml



#4 AR-1016-2


R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 353541
Conc: 165.00 ng/ml



#5 AR-1016-3


R.T.: 5.551 min
Delta R.T.: 0.003 min
Response: 257554
Conc: 163.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#5 AR-1016-3

R.T.: 4.894 min
Delta R.T.: -0.002 min
Response: 202878
Conc: 167.91 ng/ml




#6 AR-1016-4

R.T.: 5.648 min
Delta R.T.: 0.001 min
Response: 211846
Conc: 162.57 ng/ml



#6 AR-1016-4


R.T.: 4.934 min
Delta R.T.: -0.002 min
Response: 167570
Conc: 165.82 ng/ml



#7 AR-1016-5


R.T.: 5.941 min
Delta R.T.: 0.002 min
Response: 223968
Conc: 167.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#7 AR-1016-5

R.T.: 5.146 min
Delta R.T.: -0.002 min
Response: 223632
Conc: 171.58 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.512 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.832 min
Delta R.T.: -0.003 min
Response: 19253
Conc: 49.52 ng/ml



#9 AR-1221-2


R.T.: 4.600 min
Delta R.T.: 0.002 min
Response: 32731
Conc: 96.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#9 AR-1221-2

R.T.: 3.918 min
Delta R.T.: -0.002 min
Response: 28509
Conc: 95.10 ng/ml




#10 AR-1221-3

R.T.: 4.675 min
Delta R.T.: 0.002 min
Response: 126163
Conc: 121.45 ng/ml



#10 AR-1221-3


R.T.: 3.994 min
Delta R.T.: -0.003 min
Response: 110609
Conc: 120.00 ng/ml



#11 AR-1232-1


R.T.: 4.675 min
Delta R.T.: 0.000 min
Response: 126163
Conc: 146.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#11 AR-1232-1

R.T.: 3.994 min
Delta R.T.: -0.003 min
Response: 110609
Conc: 147.50 ng/ml




#12 AR-1232-2

R.T.: 5.201 min
Delta R.T.: 0.001 min
Response: 192041
Conc: 392.61 ng/ml



#12 AR-1232-2


R.T.: 4.718 min
Delta R.T.: -0.003 min
Response: 353541
Conc: 410.57 ng/ml



#13 AR-1232-3


R.T.: 5.490 min
Delta R.T.: 0.001 min
Response: 393044
Conc: 402.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#13 AR-1232-3

R.T.: 4.894 min
Delta R.T.: -0.003 min
Response: 202878
Conc: 418.68 ng/ml




#14 AR-1232-4

R.T.: 5.648 min
Delta R.T.: 0.000 min
Response: 211846
Conc: 413.88 ng/ml



#14 AR-1232-4


R.T.: 4.975 min
Delta R.T.: -0.004 min
Response: 207975
Conc: 473.31 ng/ml



#15 AR-1232-5


R.T.: 5.739 min
Delta R.T.: 0.000 min
Response: 196565
Conc: 472.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#15 AR-1232-5

R.T.: 5.146 min
Delta R.T.: -0.004 min
Response: 223632
Conc: 459.81 ng/ml




#16 AR-1242-1

R.T.: 5.467 min
Delta R.T.: 0.002 min
Response: 286553
Conc: 206.26 ng/ml



#16 AR-1242-1


R.T.: 4.699 min
Delta R.T.: -0.002 min
Response: 265311
Conc: 216.10 ng/ml



#17 AR-1242-2


R.T.: 5.490 min
Delta R.T.: 0.003 min
Response: 393044
Conc: 205.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#17 AR-1242-2

R.T.: 4.718 min
Delta R.T.: -0.002 min
Response: 353541
Conc: 213.33 ng/ml




#18 AR-1242-3

R.T.: 5.551 min
Delta R.T.: 0.002 min
Response: 257554
Conc: 207.80 ng/ml



#18 AR-1242-3


R.T.: 4.894 min
Delta R.T.: -0.002 min
Response: 202878
Conc: 214.16 ng/ml



#19 AR-1242-4


R.T.: 5.648 min
Delta R.T.: 0.001 min
Response: 211846
Conc: 207.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: -0.002 min
Response: 207975
Conc: 217.60 ng/ml




#20 AR-1242-5

R.T.: 6.381 min
Delta R.T.: 0.002 min
Response: 33571
Conc: 27.81 ng/ml



#20 AR-1242-5


R.T.: 5.497 min
Delta R.T.: -0.001 min
Response: 188307
Conc: 153.11 ng/ml



#21 AR-1248-1


R.T.: 5.467 min
Delta R.T.: 0.001 min
Response: 286553
Conc: 286.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#21 AR-1248-1

R.T.: 4.699 min
Delta R.T.: -0.002 min
Response: 265311
Conc: 302.25 ng/ml




#22 AR-1248-2

R.T.: 5.739 min
Delta R.T.: 0.000 min
Response: 196565
Conc: 135.35 ng/ml



#22 AR-1248-2


R.T.: 4.934 min
Delta R.T.: -0.002 min
Response: 167570
Conc: 138.65 ng/ml



#23 AR-1248-3


R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 223968
Conc: 140.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#23 AR-1248-3

R.T.: 4.975 min
Delta R.T.: -0.002 min
Response: 207975
Conc: 166.67 ng/ml




#24 AR-1248-4

R.T.: 6.342 min
Delta R.T.: 0.000 min
Response: 35649
Conc: 18.98 ng/ml



#24 AR-1248-4


R.T.: 5.146 min
Delta R.T.: -0.003 min
Response: 223632
Conc: 146.84 ng/ml



#25 AR-1248-5


R.T.: 6.381 min
Delta R.T.: 0.000 min
Response: 33571
Conc: 18.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#25 AR-1248-5

R.T.: 5.538 min
Delta R.T.: -0.003 min
Response: 31770
Conc: 21.31 ng/ml




#26 AR-1254-1

R.T.: 6.316 min
Delta R.T.: 0.002 min
Response: 179941
Conc: 99.24 ng/ml



#26 AR-1254-1


R.T.: 5.497 min
Delta R.T.: -0.003 min
Response: 188307
Conc: 86.16 ng/ml



#27 AR-1254-2


R.T.: 6.533 min
Delta R.T.: 0.000 min
Response: 202991
Conc: 71.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#27 AR-1254-2

R.T.: 5.642 min
Delta R.T.: -0.004 min
Response: 183885
Conc: 94.84 ng/ml




#28 AR-1254-3

R.T.: 6.898 min
Delta R.T.: 0.001 min
Response: 117662
Conc: 37.42 ng/ml



#28 AR-1254-3


R.T.: 6.060 min
Delta R.T.: 0.011 min
Response: 329493
Conc: 102.78 ng/ml



#29 AR-1254-4


R.T.: 7.182 min
Delta R.T.: 0.000 min
Response: 76485
Conc: 31.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#29 AR-1254-4

R.T.: 6.272 min
Delta R.T.: -0.004 min
Response: 38124
Conc: 19.47 ng/ml




#30 AR-1254-5

R.T.: 7.599 min
Delta R.T.: 0.000 min
Response: 707522
Conc: 263.97 ng/ml



#30 AR-1254-5


R.T.: 6.688 min
Delta R.T.: -0.004 min
Response: 622380
Conc: 216.50 ng/ml



#31 AR-1260-1


R.T.: 7.061 min
Delta R.T.: 0.002 min
Response: 487372
Conc: 193.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#31 AR-1260-1

R.T.: 6.174 min
Delta R.T.: -0.003 min
Response: 455337
Conc: 198.14 ng/ml




#32 AR-1260-2

R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 585848
Conc: 188.77 ng/ml



#32 AR-1260-2


R.T.: 6.360 min
Delta R.T.: -0.003 min
Response: 550602
Conc: 200.02 ng/ml



#33 AR-1260-3


R.T.: 7.675 min
Delta R.T.: 0.001 min
Response: 378871
Conc: 154.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#33 AR-1260-3

R.T.: 6.515 min
Delta R.T.: -0.003 min
Response: 511199
Conc: 198.30 ng/ml




#34 AR-1260-4

R.T.: 7.899 min
Delta R.T.: 0.001 min
Response: 466663
Conc: 168.71 ng/ml



#34 AR-1260-4


R.T.: 6.985 min
Delta R.T.: -0.003 min
Response: 365223
Conc: 171.61 ng/ml



#35 AR-1260-5


R.T.: 8.208 min
Delta R.T.: 0.001 min
Response: 840892
Conc: 151.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#35 AR-1260-5

R.T.: 7.223 min
Delta R.T.: -0.004 min
Response: 799650
Conc: 166.97 ng/ml




#36 AR-1262-1

R.T.: 7.675 min
Delta R.T.: 0.000 min
Response: 378871
Conc: 113.70 ng/ml



#36 AR-1262-1


R.T.: 6.725 min
Delta R.T.: -0.005 min
Response: 396689
Conc: 133.34 ng/ml



#37 AR-1262-2


R.T.: 8.208 min
Delta R.T.: 0.000 min
Response: 840892
Conc: 141.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#37 AR-1262-2

R.T.: 7.223 min
Delta R.T.: -0.005 min
Response: 799650
Conc: 163.16 ng/ml




#38 AR-1262-3

R.T.: 8.521 min
Delta R.T.: 0.014 min
Response: 542757
Conc: 131.01 ng/ml



#38 AR-1262-3


R.T.: 7.506 min
Delta R.T.: -0.005 min
Response: 151427
Conc: 73.74 ng/ml



#39 AR-1262-4


R.T.: 8.573 min
Delta R.T.: -0.020 min
Response: 289126
Conc: 89.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#39 AR-1262-4

R.T.: 7.569 min
Delta R.T.: -0.006 min
Response: 548304
Conc: 151.38 ng/ml




#40 AR-1262-5

R.T.: 9.217 min
Delta R.T.: 0.000 min
Response: 206383
Conc: 94.40 ng/ml



#40 AR-1262-5


R.T.: 8.062 min
Delta R.T.: -0.005 min
Response: 157446
Conc: 97.56 ng/ml



#41 AR-1268-1


R.T.: 8.521 min
Delta R.T.: 0.017 min
Response: 542757
Conc: 76.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#41 AR-1268-1

R.T.: 7.506 min
Delta R.T.: -0.005 min
Response: 151427
Conc: 26.78 ng/ml




#42 AR-1268-2

R.T.: 8.573 min
Delta R.T.: -0.022 min
Response: 289126
Conc: 44.13 ng/ml



#42 AR-1268-2


R.T.: 7.569 min
Delta R.T.: -0.006 min
Response: 548304
Conc: 106.02 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 8.810 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PB118307BS




#43 AR-1268-3

R.T.: 7.779 min
Delta R.T.: -0.004 min
Response: 15180
Conc: 3.53 ng/ml




#44 AR-1268-4

R.T.: 9.217 min
Delta R.T.: 0.000 min
Response: 206383
Conc: 86.51 ng/ml



#44 AR-1268-4


R.T.: 8.062 min
Delta R.T.: -0.004 min
Response: 157446
Conc: 84.93 ng/ml



#45 AR-1268-5

R.T.: 9.606 min
Delta R.T.: -0.002 min
Response: 62280
Conc: 3.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118307BS



#45 AR-1268-5

R.T.: 8.347 min
Delta R.T.: -0.005 min
Response: 50343
Conc: 4.42 ng/ml