

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032719\
 Data File : P0054678.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2019 20:09
 Operator : SM/SJ
 Sample : K2103-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SHORE-RD-CLVRT-STLD-3H-A

Manual Integrations
 APPROVED

Sohil
 3/28/2019 4:07:59 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 00:55:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.303	3.619	501361	452425	14.446	14.177
2) SA Decachlor...	9.919	8.598	681424	490130	13.823m	15.390

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032719\
Data File : P0054678.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Mar 2019 20:09
Operator : SM/SJ
Sample : K2103-03
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SHORE-RD-CLVRT-STLD-3H-A


Manual Integrations
APPROVED

Sohil
3/28/2019 4:07:59 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 00:55:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.303 min
Delta R.T.: 0.000 min
Response: 501361
Conc: 14.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
SHORE-RD-CLVRT-STLD-3H-A


Manual Integrations
APPROVED

Sohil
3/28/2019 4:07:59 PM




#1 Tetrachloro-m-xylene

R.T.: 3.619 min
Delta R.T.: -0.001 min
Response: 452425
Conc: 14.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.919 min
Delta R.T.: -0.005 min
Response: 681424
Conc: 13.82 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.598 min
Delta R.T.: -0.006 min
Response: 490130
Conc: 15.39 ng/ml