

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0032720\
 Data File : P0067484.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2020 17:12
 Operator : DD\AJ
 Sample : PB127857BSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 PB127857BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 19:00:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 16:36:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.939	3.948	525953	762218	24.746	21.803
2)	SA Decachlor...	10.924	9.240	671745	835894	23.379	21.430
Target Compounds							
3)	L1 AR-1016-1	6.260	5.204	220846	329286	244.060	215.263
4)	L1 AR-1016-2	6.283	5.225	301420	429146	241.924	212.719
5)	L1 AR-1016-3	6.349	5.415	178518	234905	227.903	210.482
6)	L1 AR-1016-4	6.456	5.468	150620	191481	236.803	205.879
7)	L1 AR-1016-5	6.770	5.696	153458	241924	237.503	204.015
9)	L2 AR-1221-2	0.000	4.304	0	40834	N.D.	127.843 #
10)	L2 AR-1221-3	5.375	4.393	107920	159406	184.622	161.703
11)	L3 AR-1232-1	5.375	4.393	107920	159406	251.423	220.264
12)	L3 AR-1232-2	5.965	5.225	147685	429146	633.380	584.935
13)	L3 AR-1232-3	6.283	5.415	301420	234905	716.211	601.139
14)	L3 AR-1232-4	6.456	5.513	150620	233675	695.291	654.210
15)	L3 AR-1232-5	6.553	5.696	129763	241924	805.398	615.555
16)	L4 AR-1242-1	6.260	5.204	220846	329286	323.302	282.107
17)	L4 AR-1242-2	6.283	5.225	301420	429146	324.466	280.247
18)	L4 AR-1242-3	6.349	5.415	178518	234905	301.295	278.422
19)	L4 AR-1242-4	6.456	5.513	150620	233675	315.239	276.364
20)	L4 AR-1242-5	0.000	6.076	0	187280	N.D.	175.129 #
21)	L5 AR-1248-1	6.260	5.204	220846	329286	488.751	415.408
22)	L5 AR-1248-2	6.553	5.468	129763	191481	215.237	185.590
23)	L5 AR-1248-3	6.770	5.513	153458	233675	217.323	222.022
24)	L5 AR-1248-4	7.167	5.696	90528	241924	109.001	191.294 #
25)	L5 AR-1248-5	0.000	6.076	0	187280	N.D.	149.296 #
26)	L6 AR-1254-1	7.167	6.076	90528	187280	107.609	96.186
27)	L6 AR-1254-2	7.396	6.236	132297	191720	98.658	111.228
28)	L6 AR-1254-3	7.834	6.656	167138	96291	115.332	34.332 #
29)	L6 AR-1254-4	0.000	6.895	0	44124	N.D.	23.979 #
30)	L6 AR-1254-5	8.497	7.323	443631	680655	359.760	274.319
31)	L7 AR-1260-1	7.943	6.793	302824	469174	249.613	218.955
32)	L7 AR-1260-2	8.207	6.990	367556	566045	245.460	214.213
33)	L7 AR-1260-3	8.570	7.144	235105	530404	203.935	217.340
34)	L7 AR-1260-4	8.800	7.627	281000	401052	209.232	188.944
35)	L7 AR-1260-5	9.124	7.875	544295	917233	191.546	183.986
36)	L8 AR-1262-1	8.570	7.323	235105	680655	161.454	473.702 #
37)	L8 AR-1262-2	9.124	7.875	544295	917233	196.955	198.704
38)	L8 AR-1262-3	9.454	8.162	362149	196280	186.309	98.310 #
39)	L8 AR-1262-4	9.505	8.222	196210	693158	134.013	197.536 #
40)	L8 AR-1262-5	10.185	8.720	144544	229993	132.396	135.384
41)	L9 AR-1268-1	9.454	8.162	362149	196280	110.672	36.988 #
42)	L9 AR-1268-2	9.505	8.222	196210	693158	64.258	145.731 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032720\
 Data File : P0067484.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Mar 2020 17:12
 Operator : DD\AJ
 Sample : PB127857BSD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB127857BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 19:00:39 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Mar 19 16:36:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
44) L9	AR-1268-4	10.185	8.720	144544	229993	119.092	122.982
45) L9	AR-1268-5	0.000	8.998	0	82070	N.D.	6.796 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0032720\
Data File : P0067484.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Mar 2020 17:12
Operator : DD\AJ
Sample : PB127857BSD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB127857BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 19:00:39 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0031920.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Mar 19 16:36:53 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.939 min
Delta R.T.: -0.001 min
Response: 525953
Conc: 24.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#1 Tetrachloro-m-xylene

R.T.: 3.948 min
Delta R.T.: -0.004 min
Response: 762218
Conc: 21.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.924 min
Delta R.T.: 0.001 min
Response: 671745
Conc: 23.38 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.240 min
Delta R.T.: -0.009 min
Response: 835894
Conc: 21.43 ng/ml



#3 AR-1016-1


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 220846
Conc: 244.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#3 AR-1016-1

R.T.: 5.204 min
Delta R.T.: -0.006 min
Response: 329286
Conc: 215.26 ng/ml




#4 AR-1016-2

R.T.: 6.283 min
Delta R.T.: 0.000 min
Response: 301420
Conc: 241.92 ng/ml



#4 AR-1016-2


R.T.: 5.225 min
Delta R.T.: -0.005 min
Response: 429146
Conc: 212.72 ng/ml



#5 AR-1016-3


R.T.: 6.349 min
Delta R.T.: 0.000 min
Response: 178518
Conc: 227.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#5 AR-1016-3

R.T.: 5.415 min
Delta R.T.: -0.006 min
Response: 234905
Conc: 210.48 ng/ml




#6 AR-1016-4

R.T.: 6.456 min
Delta R.T.: -0.001 min
Response: 150620
Conc: 236.80 ng/ml



#6 AR-1016-4


R.T.: 5.468 min
Delta R.T.: -0.005 min
Response: 191481
Conc: 205.88 ng/ml



#7 AR-1016-5


R.T.: 6.770 min
Delta R.T.: 0.000 min
Response: 153458
Conc: 237.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#7 AR-1016-5

R.T.: 5.696 min
Delta R.T.: -0.006 min
Response: 241924
Conc: 204.02 ng/ml




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 5.290 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.304 min
Delta R.T.: -0.006 min
Response: 40834
Conc: 127.84 ng/ml



#10 AR-1221-3


R.T.: 5.375 min
Delta R.T.: 0.000 min
Response: 107920
Conc: 184.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#10 AR-1221-3

R.T.: 4.393 min
Delta R.T.: -0.006 min
Response: 159406
Conc: 161.70 ng/ml




#11 AR-1232-1

R.T.: 5.375 min
Delta R.T.: -0.001 min
Response: 107920
Conc: 251.42 ng/ml



#11 AR-1232-1


R.T.: 4.393 min
Delta R.T.: -0.006 min
Response: 159406
Conc: 220.26 ng/ml



#12 AR-1232-2


R.T.: 5.965 min
Delta R.T.: 0.000 min
Response: 147685
Conc: 633.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#12 AR-1232-2

R.T.: 5.225 min
Delta R.T.: -0.006 min
Response: 429146
Conc: 584.93 ng/ml




#13 AR-1232-3

R.T.: 6.283 min
Delta R.T.: 0.000 min
Response: 301420
Conc: 716.21 ng/ml



#13 AR-1232-3


R.T.: 5.415 min
Delta R.T.: -0.006 min
Response: 234905
Conc: 601.14 ng/ml



#14 AR-1232-4


R.T.: 6.456 min
Delta R.T.: 0.000 min
Response: 150620
Conc: 695.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#14 AR-1232-4

R.T.: 5.513 min
Delta R.T.: -0.006 min
Response: 233675
Conc: 654.21 ng/ml




#15 AR-1232-5

R.T.: 6.553 min
Delta R.T.: -0.001 min
Response: 129763
Conc: 805.40 ng/ml



#15 AR-1232-5


R.T.: 5.696 min
Delta R.T.: -0.006 min
Response: 241924
Conc: 615.55 ng/ml



#16 AR-1242-1


R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 220846
Conc: 323.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#16 AR-1242-1

R.T.: 5.204 min
Delta R.T.: -0.005 min
Response: 329286
Conc: 282.11 ng/ml




#17 AR-1242-2

R.T.: 6.283 min
Delta R.T.: -0.007 min
Response: 301420
Conc: 324.47 ng/ml



#17 AR-1242-2


R.T.: 5.225 min
Delta R.T.: -0.005 min
Response: 429146
Conc: 280.25 ng/ml



#18 AR-1242-3


R.T.: 6.349 min
Delta R.T.: 0.000 min
Response: 178518
Conc: 301.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#18 AR-1242-3

R.T.: 5.415 min
Delta R.T.: -0.005 min
Response: 234905
Conc: 278.42 ng/ml




#19 AR-1242-4

R.T.: 6.456 min
Delta R.T.: 0.000 min
Response: 150620
Conc: 315.24 ng/ml



#19 AR-1242-4


R.T.: 5.513 min
Delta R.T.: -0.005 min
Response: 233675
Conc: 276.36 ng/ml



#20 AR-1242-5


R.T.: 0.000 min
Exp R.T.: 7.237 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#20 AR-1242-5

R.T.: 6.076 min
Delta R.T.: -0.005 min
Response: 187280
Conc: 175.13 ng/ml




#21 AR-1248-1

R.T.: 6.260 min
Delta R.T.: 0.000 min
Response: 220846
Conc: 488.75 ng/ml



#21 AR-1248-1


R.T.: 5.204 min
Delta R.T.: -0.006 min
Response: 329286
Conc: 415.41 ng/ml



#22 AR-1248-2


R.T.: 6.553 min
Delta R.T.: -0.002 min
Response: 129763
Conc: 215.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#22 AR-1248-2

R.T.: 5.468 min
Delta R.T.: -0.006 min
Response: 191481
Conc: 185.59 ng/ml




#23 AR-1248-3

R.T.: 6.770 min
Delta R.T.: -0.001 min
Response: 153458
Conc: 217.32 ng/ml



#23 AR-1248-3


R.T.: 5.513 min
Delta R.T.: -0.006 min
Response: 233675
Conc: 222.02 ng/ml



#24 AR-1248-4


R.T.: 7.167 min
Delta R.T.: -0.031 min
Response: 90528
Conc: 109.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#24 AR-1248-4

R.T.: 5.696 min
Delta R.T.: -0.006 min
Response: 241924
Conc: 191.29 ng/ml




#25 AR-1248-5

R.T.: 0.000 min
Exp R.T. : 7.238 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.076 min
Delta R.T.: -0.048 min
Response: 187280
Conc: 149.30 ng/ml



#26 AR-1254-1


R.T.: 7.167 min
Delta R.T.: 0.000 min
Response: 90528
Conc: 107.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#26 AR-1254-1

R.T.: 6.076 min
Delta R.T.: -0.005 min
Response: 187280
Conc: 96.19 ng/ml




#27 AR-1254-2

R.T.: 7.396 min
Delta R.T.: 0.000 min
Response: 132297
Conc: 98.66 ng/ml



#27 AR-1254-2


R.T.: 6.236 min
Delta R.T.: -0.005 min
Response: 191720
Conc: 111.23 ng/ml



#28 AR-1254-3


R.T.: 7.834 min
Delta R.T.: 0.058 min
Response: 167138
Conc: 115.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#28 AR-1254-3

R.T.: 6.656 min
Delta R.T.: -0.005 min
Response: 96291
Conc: 34.33 ng/ml




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 8.071 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.895 min
Delta R.T.: -0.005 min
Response: 44124
Conc: 23.98 ng/ml



#30 AR-1254-5


R.T.: 8.497 min
Delta R.T.: -0.002 min
Response: 443631
Conc: 359.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#30 AR-1254-5

R.T.: 7.323 min
Delta R.T.: -0.006 min
Response: 680655
Conc: 274.32 ng/ml




#31 AR-1260-1

R.T.: 7.943 min
Delta R.T.: 0.000 min
Response: 302824
Conc: 249.61 ng/ml



#31 AR-1260-1


R.T.: 6.793 min
Delta R.T.: -0.007 min
Response: 469174
Conc: 218.95 ng/ml



#32 AR-1260-2


R.T.: 8.207 min
Delta R.T.: 0.000 min
Response: 367556
Conc: 245.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#32 AR-1260-2

R.T.: 6.990 min
Delta R.T.: -0.007 min
Response: 566045
Conc: 214.21 ng/ml




#33 AR-1260-3

R.T.: 8.570 min
Delta R.T.: 0.000 min
Response: 235105
Conc: 203.93 ng/ml



#33 AR-1260-3


R.T.: 7.144 min
Delta R.T.: -0.007 min
Response: 530404
Conc: 217.34 ng/ml



#34 AR-1260-4


R.T.: 8.800 min
Delta R.T.: 0.000 min
Response: 281000
Conc: 209.23 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#34 AR-1260-4

R.T.: 7.627 min
Delta R.T.: -0.007 min
Response: 401052
Conc: 188.94 ng/ml




#35 AR-1260-5

R.T.: 9.124 min
Delta R.T.: 0.000 min
Response: 544295
Conc: 191.55 ng/ml



#35 AR-1260-5


R.T.: 7.875 min
Delta R.T.: -0.006 min
Response: 917233
Conc: 183.99 ng/ml



#36 AR-1262-1


R.T.: 8.570 min
Delta R.T.: 0.000 min
Response: 235105
Conc: 161.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#36 AR-1262-1

R.T.: 7.323 min
Delta R.T.: -0.005 min
Response: 680655
Conc: 473.70 ng/ml




#37 AR-1262-2

R.T.: 9.124 min
Delta R.T.: 0.000 min
Response: 544295
Conc: 196.95 ng/ml



#37 AR-1262-2


R.T.: 7.875 min
Delta R.T.: -0.006 min
Response: 917233
Conc: 198.70 ng/ml



#38 AR-1262-3


R.T.: 9.454 min
Delta R.T.: 0.023 min
Response: 362149
Conc: 186.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#38 AR-1262-3

R.T.: 8.162 min
Delta R.T.: -0.006 min
Response: 196280
Conc: 98.31 ng/ml




#39 AR-1262-4

R.T.: 9.505 min
Delta R.T.: -0.018 min
Response: 196210
Conc: 134.01 ng/ml



#39 AR-1262-4


R.T.: 8.222 min
Delta R.T.: -0.008 min
Response: 693158
Conc: 197.54 ng/ml



#40 AR-1262-5


R.T.: 10.185 min
Delta R.T.: 0.002 min
Response: 144544
Conc: 132.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#40 AR-1262-5

R.T.: 8.720 min
Delta R.T.: -0.006 min
Response: 229993
Conc: 135.38 ng/ml




#41 AR-1268-1

R.T.: 9.454 min
Delta R.T.: 0.024 min
Response: 362149
Conc: 110.67 ng/ml



#41 AR-1268-1


R.T.: 8.162 min
Delta R.T.: -0.007 min
Response: 196280
Conc: 36.99 ng/ml



#42 AR-1268-2


R.T.: 9.505 min
Delta R.T.: -0.021 min
Response: 196210
Conc: 64.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB127857BSD




#42 AR-1268-2

R.T.: 8.222 min
Delta R.T.: -0.010 min
Response: 693158
Conc: 145.73 ng/ml




#44 AR-1268-4

R.T.: 10.185 min
Delta R.T.: 0.002 min
Response: 144544
Conc: 119.09 ng/ml



#44 AR-1268-4


R.T.: 8.720 min
Delta R.T.: -0.006 min
Response: 229993
Conc: 122.98 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 10.592 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
PB127857BSD



#45 AR-1268-5

R.T.: 8.998 min
Delta R.T.: -0.008 min
Response: 82070
Conc: 6.80 ng/ml