

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040420\
 Data File : P0067665.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Apr 2020 18:51
 Operator : DD\AJ
 Sample : PB128001BSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB128001BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 02:27:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.935	3.944	559204	813174	23.482	24.261
2)	SA Decachlor...	10.918	9.230	741025	848085	21.767	20.512
Target Compounds							
3)	L1 AR-1016-1	6.255	5.198	229973	342878	223.008	233.899
4)	L1 AR-1016-2	6.278	5.219	316949	458477	224.637	234.039
5)	L1 AR-1016-3	6.344	5.409	196227	251773	223.244	240.731
6)	L1 AR-1016-4	6.452	5.462	153190	202834	217.238	227.372
7)	L1 AR-1016-5	6.766	5.690	150405	254898	209.175	225.251
8)	L2 AR-1221-1	0.000	4.200	0	30170	N.D.	67.894 #
9)	L2 AR-1221-2	5.284	4.299	27913	44488	120.008	141.072
10)	L2 AR-1221-3	5.371	4.388	109149	166486	145.282	169.582
11)	L3 AR-1232-1	5.371	4.388	109149	166486	211.995	240.563
12)	L3 AR-1232-2	5.960	5.219	149024	458477	555.485	632.063
13)	L3 AR-1232-3	6.278	5.409	316949	251773	604.664	653.171
14)	L3 AR-1232-4	6.452	5.507	153190	245153	595.899	692.978
15)	L3 AR-1232-5	6.549	5.690	128301	254898	684.958	667.035
16)	L4 AR-1242-1	6.255	5.198	229973	342878	334.964	349.463
17)	L4 AR-1242-2	6.278	5.219	316949	458477	337.330	351.862
18)	L4 AR-1242-3	6.344	5.409	196227	251773	334.844	352.093
19)	L4 AR-1242-4	6.452	5.507	153190	245153	325.723	340.580
20)	L4 AR-1242-5	7.232	6.069	22892	204558	42.850	218.606 #
21)	L5 AR-1248-1	6.255	5.198	229973	342878	428.644	438.905
22)	L5 AR-1248-2	6.549	5.462	128301	202834	182.352	201.575
23)	L5 AR-1248-3	6.766	5.507	150405	245153	184.804	239.681 #
24)	L5 AR-1248-4	7.193	5.690	26116	254898	27.681	204.529 #
25)	L5 AR-1248-5	7.232	6.112	22892	32291	25.401	26.094
26)	L6 AR-1254-1	7.162	6.069	117634	204558	95.242	79.777
27)	L6 AR-1254-2	7.391	6.229	131322	199881	66.773	87.642 #
28)	L6 AR-1254-3	7.772	6.668	76910	354076	37.064	96.977 #
29)	L6 AR-1254-4	8.065	6.887	51554	45429	31.651	18.845 #
30)	L6 AR-1254-5	8.492	7.316	453811	708430	264.501	212.250
31)	L7 AR-1260-1	7.938	6.786	313216	491692	231.766	234.538
32)	L7 AR-1260-2	8.202	6.984	380600	599088	225.451	233.313
33)	L7 AR-1260-3	8.566	7.137	239914	547574	184.056	227.712
34)	L7 AR-1260-4	8.796	7.620	297310	410952	193.460	195.533
35)	L7 AR-1260-5	9.119	7.867	575494	974468	182.429	192.487
36)	L8 AR-1262-1	8.566	7.316	239914	708430	143.108	511.235 #
37)	L8 AR-1262-2	9.119	7.867	575494	974468	200.557	213.134
38)	L8 AR-1262-3	9.448	8.154	377816	206397	181.569	105.012 #
39)	L8 AR-1262-4	9.501	8.215	212542	700062	134.319	204.221 #
40)	L8 AR-1262-5	10.177	8.712	158670	221546	134.571	130.333
41)	L9 AR-1268-1	9.448	8.154	377816	206397	102.246	38.359 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040420\
 Data File : P0067665.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 Apr 2020 18:51
 Operator : DD\AJ
 Sample : PB128001BSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB128001BSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 04 02:27:29 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.501	8.215	212542	700062	61.323	141.784 #
43) L9	AR-1268-3	9.743	8.426	16278	18938	5.606	4.518
44) L9	AR-1268-4	10.177	8.712	158670	221546	114.297	114.501
45) L9	AR-1268-5	10.587	8.989	62972	75080	6.351	5.822



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040420\
Data File : PO067665.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 Apr 2020 18:51
Operator : DD\AJ
Sample : PB128001BSD
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB128001BSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 04 02:27:29 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.935 min
Delta R.T.: 0.000 min
Response: 559204
Conc: 23.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#1 Tetrachloro-m-xylene

R.T.: 3.944 min
Delta R.T.: 0.000 min
Response: 813174
Conc: 24.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.918 min
Delta R.T.: -0.001 min
Response: 741025
Conc: 21.77 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.230 min
Delta R.T.: -0.003 min
Response: 848085
Conc: 20.51 ng/ml



#3 AR-1016-1


R.T.: 6.255 min
Delta R.T.: -0.001 min
Response: 229973
Conc: 223.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#3 AR-1016-1

R.T.: 5.198 min
Delta R.T.: -0.002 min
Response: 342878
Conc: 233.90 ng/ml




#4 AR-1016-2

R.T.: 6.278 min
Delta R.T.: -0.001 min
Response: 316949
Conc: 224.64 ng/ml



#4 AR-1016-2


R.T.: 5.219 min
Delta R.T.: -0.002 min
Response: 458477
Conc: 234.04 ng/ml



#5 AR-1016-3


R.T.: 6.344 min
Delta R.T.: -0.002 min
Response: 196227
Conc: 223.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#5 AR-1016-3

R.T.: 5.409 min
Delta R.T.: 0.000 min
Response: 251773
Conc: 240.73 ng/ml




#6 AR-1016-4

R.T.: 6.452 min
Delta R.T.: 0.000 min
Response: 153190
Conc: 217.24 ng/ml



#6 AR-1016-4


R.T.: 5.462 min
Delta R.T.: -0.002 min
Response: 202834
Conc: 227.37 ng/ml



#7 AR-1016-5


R.T.: 6.766 min
Delta R.T.: -0.001 min
Response: 150405
Conc: 209.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#7 AR-1016-5

R.T.: 5.690 min
Delta R.T.: -0.002 min
Response: 254898
Conc: 225.25 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 5.187 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.200 min
Delta R.T.: -0.002 min
Response: 30170
Conc: 67.89 ng/ml



#9 AR-1221-2


R.T.: 5.284 min
Delta R.T.: -0.001 min
Response: 27913
Conc: 120.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#9 AR-1221-2

R.T.: 4.299 min
Delta R.T.: -0.002 min
Response: 44488
Conc: 141.07 ng/ml




#10 AR-1221-3

R.T.: 5.371 min
Delta R.T.: -0.002 min
Response: 109149
Conc: 145.28 ng/ml



#10 AR-1221-3


R.T.: 4.388 min
Delta R.T.: -0.001 min
Response: 166486
Conc: 169.58 ng/ml



#11 AR-1232-1


R.T.: 5.371 min
Delta R.T.: -0.001 min
Response: 109149
Conc: 211.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#11 AR-1232-1

R.T.: 4.388 min
Delta R.T.: -0.001 min
Response: 166486
Conc: 240.56 ng/ml




#12 AR-1232-2

R.T.: 5.960 min
Delta R.T.: -0.002 min
Response: 149024
Conc: 555.49 ng/ml



#12 AR-1232-2


R.T.: 5.219 min
Delta R.T.: -0.002 min
Response: 458477
Conc: 632.06 ng/ml



#13 AR-1232-3


R.T.: 6.278 min
Delta R.T.: -0.002 min
Response: 316949
Conc: 604.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#13 AR-1232-3

R.T.: 5.409 min
Delta R.T.: -0.002 min
Response: 251773
Conc: 653.17 ng/ml




#14 AR-1232-4

R.T.: 6.452 min
Delta R.T.: -0.001 min
Response: 153190
Conc: 595.90 ng/ml



#14 AR-1232-4


R.T.: 5.507 min
Delta R.T.: -0.002 min
Response: 245153
Conc: 692.98 ng/ml



#15 AR-1232-5


R.T.: 6.549 min
Delta R.T.: -0.002 min
Response: 128301
Conc: 684.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#15 AR-1232-5

R.T.: 5.690 min
Delta R.T.: -0.002 min
Response: 254898
Conc: 667.04 ng/ml




#16 AR-1242-1

R.T.: 6.255 min
Delta R.T.: -0.001 min
Response: 229973
Conc: 334.96 ng/ml



#16 AR-1242-1


R.T.: 5.198 min
Delta R.T.: -0.002 min
Response: 342878
Conc: 349.46 ng/ml



#17 AR-1242-2


R.T.: 6.278 min
Delta R.T.: -0.002 min
Response: 316949
Conc: 337.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#17 AR-1242-2

R.T.: 5.219 min
Delta R.T.: -0.002 min
Response: 458477
Conc: 351.86 ng/ml




#18 AR-1242-3

R.T.: 6.344 min
Delta R.T.: -0.001 min
Response: 196227
Conc: 334.84 ng/ml



#18 AR-1242-3


R.T.: 5.409 min
Delta R.T.: -0.002 min
Response: 251773
Conc: 352.09 ng/ml



#19 AR-1242-4


R.T.: 6.452 min
Delta R.T.: -0.001 min
Response: 153190
Conc: 325.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#19 AR-1242-4

R.T.: 5.507 min
Delta R.T.: -0.002 min
Response: 245153
Conc: 340.58 ng/ml




#20 AR-1242-5

R.T.: 7.232 min
Delta R.T.: -0.002 min
Response: 22892
Conc: 42.85 ng/ml



#20 AR-1242-5


R.T.: 6.069 min
Delta R.T.: -0.001 min
Response: 204558
Conc: 218.61 ng/ml



#21 AR-1248-1


R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 229973
Conc: 428.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#21 AR-1248-1

R.T.: 5.198 min
Delta R.T.: 0.000 min
Response: 342878
Conc: 438.90 ng/ml




#22 AR-1248-2

R.T.: 6.549 min
Delta R.T.: 0.000 min
Response: 128301
Conc: 182.35 ng/ml



#22 AR-1248-2


R.T.: 5.462 min
Delta R.T.: 0.000 min
Response: 202834
Conc: 201.57 ng/ml



#23 AR-1248-3


R.T.: 6.766 min
Delta R.T.: 0.000 min
Response: 150405
Conc: 184.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#23 AR-1248-3

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 245153
Conc: 239.68 ng/ml




#24 AR-1248-4

R.T.: 7.193 min
Delta R.T.: -0.002 min
Response: 26116
Conc: 27.68 ng/ml



#24 AR-1248-4


R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 254898
Conc: 204.53 ng/ml



#25 AR-1248-5


R.T.: 7.232 min
Delta R.T.: -0.002 min
Response: 22892
Conc: 25.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#25 AR-1248-5

R.T.: 6.112 min
Delta R.T.: 0.000 min
Response: 32291
Conc: 26.09 ng/ml




#26 AR-1254-1

R.T.: 7.162 min
Delta R.T.: -0.002 min
Response: 117634
Conc: 95.24 ng/ml



#26 AR-1254-1


R.T.: 6.069 min
Delta R.T.: -0.001 min
Response: 204558
Conc: 79.78 ng/ml



#27 AR-1254-2


R.T.: 7.391 min
Delta R.T.: -0.002 min
Response: 131322
Conc: 66.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#27 AR-1254-2

R.T.: 6.229 min
Delta R.T.: -0.001 min
Response: 199881
Conc: 87.64 ng/ml




#28 AR-1254-3

R.T.: 7.772 min
Delta R.T.: 0.000 min
Response: 76910
Conc: 37.06 ng/ml



#28 AR-1254-3


R.T.: 6.668 min
Delta R.T.: 0.019 min
Response: 354076
Conc: 96.98 ng/ml



#29 AR-1254-4


R.T.: 8.065 min
Delta R.T.: -0.003 min
Response: 51554
Conc: 31.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#29 AR-1254-4

R.T.: 6.887 min
Delta R.T.: -0.001 min
Response: 45429
Conc: 18.84 ng/ml




#30 AR-1254-5

R.T.: 8.492 min
Delta R.T.: -0.002 min
Response: 453811
Conc: 264.50 ng/ml



#30 AR-1254-5


R.T.: 7.316 min
Delta R.T.: -0.001 min
Response: 708430
Conc: 212.25 ng/ml



#31 AR-1260-1


R.T.: 7.938 min
Delta R.T.: -0.002 min
Response: 313216
Conc: 231.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#31 AR-1260-1

R.T.: 6.786 min
Delta R.T.: -0.002 min
Response: 491692
Conc: 234.54 ng/ml




#32 AR-1260-2

R.T.: 8.202 min
Delta R.T.: -0.002 min
Response: 380600
Conc: 225.45 ng/ml



#32 AR-1260-2


R.T.: 6.984 min
Delta R.T.: -0.002 min
Response: 599088
Conc: 233.31 ng/ml



#33 AR-1260-3


R.T.: 8.566 min
Delta R.T.: -0.001 min
Response: 239914
Conc: 184.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#33 AR-1260-3

R.T.: 7.137 min
Delta R.T.: -0.002 min
Response: 547574
Conc: 227.71 ng/ml




#34 AR-1260-4

R.T.: 8.796 min
Delta R.T.: -0.001 min
Response: 297310
Conc: 193.46 ng/ml



#34 AR-1260-4


R.T.: 7.620 min
Delta R.T.: -0.002 min
Response: 410952
Conc: 195.53 ng/ml



#35 AR-1260-5


R.T.: 9.119 min
Delta R.T.: -0.002 min
Response: 575494
Conc: 182.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#35 AR-1260-5

R.T.: 7.867 min
Delta R.T.: -0.002 min
Response: 974468
Conc: 192.49 ng/ml




#36 AR-1262-1

R.T.: 8.566 min
Delta R.T.: -0.002 min
Response: 239914
Conc: 143.11 ng/ml



#36 AR-1262-1


R.T.: 7.316 min
Delta R.T.: -0.002 min
Response: 708430
Conc: 511.23 ng/ml



#37 AR-1262-2


R.T.: 9.119 min
Delta R.T.: -0.002 min
Response: 575494
Conc: 200.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#37 AR-1262-2

R.T.: 7.867 min
Delta R.T.: -0.002 min
Response: 974468
Conc: 213.13 ng/ml




#38 AR-1262-3

R.T.: 9.448 min
Delta R.T.: 0.020 min
Response: 377816
Conc: 181.57 ng/ml



#38 AR-1262-3


R.T.: 8.154 min
Delta R.T.: -0.003 min
Response: 206397
Conc: 105.01 ng/ml



#39 AR-1262-4


R.T.: 9.501 min
Delta R.T.: -0.020 min
Response: 212542
Conc: 134.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#39 AR-1262-4

R.T.: 8.215 min
Delta R.T.: -0.004 min
Response: 700062
Conc: 204.22 ng/ml




#40 AR-1262-5

R.T.: 10.177 min
Delta R.T.: -0.003 min
Response: 158670
Conc: 134.57 ng/ml



#40 AR-1262-5


R.T.: 8.712 min
Delta R.T.: -0.003 min
Response: 221546
Conc: 130.33 ng/ml



#41 AR-1268-1


R.T.: 9.448 min
Delta R.T.: 0.021 min
Response: 377816
Conc: 102.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#41 AR-1268-1

R.T.: 8.154 min
Delta R.T.: -0.003 min
Response: 206397
Conc: 38.36 ng/ml




#42 AR-1268-2

R.T.: 9.501 min
Delta R.T.: -0.022 min
Response: 212542
Conc: 61.32 ng/ml



#42 AR-1268-2


R.T.: 8.215 min
Delta R.T.: -0.006 min
Response: 700062
Conc: 141.78 ng/ml



#43 AR-1268-3


R.T.: 9.743 min
Delta R.T.: -0.004 min
Response: 16278
Conc: 5.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD




#43 AR-1268-3

R.T.: 8.426 min
Delta R.T.: -0.003 min
Response: 18938
Conc: 4.52 ng/ml




#44 AR-1268-4

R.T.: 10.177 min
Delta R.T.: -0.003 min
Response: 158670
Conc: 114.30 ng/ml



#44 AR-1268-4


R.T.: 8.712 min
Delta R.T.: -0.003 min
Response: 221546
Conc: 114.50 ng/ml



#45 AR-1268-5

R.T.: 10.587 min
Delta R.T.: 0.000 min
Response: 62972
Conc: 6.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB128001BSD



#45 AR-1268-5

R.T.: 8.989 min
Delta R.T.: -0.004 min
Response: 75080
Conc: 5.82 ng/ml