

Data Path : P:\HPCHEM1\ECD 0\Data\P0040518\
 Data File : P0043634.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 05 Apr 2018 19:14
 Operator : SM/UA
 Sample : J2184-03
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-2-1(1-3)

Manual Integrations
 APPROVED

ugo
 4/6/2018 3:15:25 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 06 01:43:47 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0040418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 04 15:17:57 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.066	3.324	5462308	7415630	30.725	27.600
2) SA Decachlor...	9.405	8.084	3766327	4298548	20.900m	18.335

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO040518\
Data File : PO043634.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 05 Apr 2018 19:14
Operator : SM/UA
Sample : J2184-03
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
SB-2-1(1-3)


Manual Integrations
APPROVED

ugo
4/6/2018 3:15:25 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 06 01:43:47 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO040418.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 04 15:17:57 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.066 min
Delta R.T.: -0.004 min
Response: 5462308
Conc: 30.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-2-1(1-3)


Manual Integrations
APPROVED

ugo
4/6/2018 3:15:25 PM




#1 Tetrachloro-m-xylene

R.T.: 3.324 min
Delta R.T.: -0.002 min
Response: 7415630
Conc: 27.60 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.405 min
Delta R.T.: -0.012 min
Response: 3766327
Conc: 20.90 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.084 min
Delta R.T.: -0.006 min
Response: 4298548
Conc: 18.34 ng/ml