

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040720\
 Data File : P0067732.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Apr 2020 18:54
 Operator : DD\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 07 19:15:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.932	3.942	1284188	1799533	53.926	53.690
2) SA Decachlor...	10.913	9.229	1557878	1822484	45.761	44.079
Target Compounds						
3) L1 AR-1016-1	6.252	5.197	555271	774875	538.453	528.593
4) L1 AR-1016-2	6.276	5.218	764047	1050737	541.517	536.370
5) L1 AR-1016-3	6.342	5.408	471048	583917	535.901	558.309
6) L1 AR-1016-4	6.449	5.461	382075	477217	541.817	534.949
7) L1 AR-1016-5	6.763	5.688	384633	575728	534.926	508.766
31) L7 AR-1260-1	7.936	6.784	723434	1109930	535.310	529.437
32) L7 AR-1260-2	8.200	6.982	879223	1322826	520.813	515.171
33) L7 AR-1260-3	8.564	7.136	677001	1196846	519.378	497.716
34) L7 AR-1260-4	8.793	7.619	766462	1060748	498.739	504.710
35) L7 AR-1260-5	9.117	7.865	1522448	2477626	482.610	489.405



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040720\
Data File : PO067732.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Apr 2020 18:54
Operator : DD\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 07 19:15:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.932 min
Delta R.T.: -0.003 min
Response: 1284188
Conc: 53.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.942 min
Delta R.T.: -0.001 min
Response: 1799533
Conc: 53.69 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.913 min
Delta R.T.: -0.006 min
Response: 1557878
Conc: 45.76 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.229 min
Delta R.T.: -0.004 min
Response: 1822484
Conc: 44.08 ng/ml



#3 AR-1016-1


R.T.: 6.252 min
Delta R.T.: -0.004 min
Response: 555271
Conc: 538.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 5.197 min
Delta R.T.: -0.003 min
Response: 774875
Conc: 528.59 ng/ml




#4 AR-1016-2

R.T.: 6.276 min
Delta R.T.: -0.003 min
Response: 764047
Conc: 541.52 ng/ml



#4 AR-1016-2


R.T.: 5.218 min
Delta R.T.: -0.003 min
Response: 1050737
Conc: 536.37 ng/ml



#5 AR-1016-3


R.T.: 6.342 min
Delta R.T.: -0.004 min
Response: 471048
Conc: 535.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.408 min
Delta R.T.: -0.002 min
Response: 583917
Conc: 558.31 ng/ml




#6 AR-1016-4

R.T.: 6.449 min
Delta R.T.: -0.003 min
Response: 382075
Conc: 541.82 ng/ml



#6 AR-1016-4


R.T.: 5.461 min
Delta R.T.: -0.003 min
Response: 477217
Conc: 534.95 ng/ml



#7 AR-1016-5


R.T.: 6.763 min
Delta R.T.: -0.004 min
Response: 384633
Conc: 534.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.688 min
Delta R.T.: -0.003 min
Response: 575728
Conc: 508.77 ng/ml




#31 AR-1260-1

R.T.: 7.936 min
Delta R.T.: -0.004 min
Response: 723434
Conc: 535.31 ng/ml



#31 AR-1260-1


R.T.: 6.784 min
Delta R.T.: -0.004 min
Response: 1109930
Conc: 529.44 ng/ml



#32 AR-1260-2


R.T.: 8.200 min
Delta R.T.: -0.004 min
Response: 879223
Conc: 520.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.982 min
Delta R.T.: -0.003 min
Response: 1322826
Conc: 515.17 ng/ml




#33 AR-1260-3

R.T.: 8.564 min
Delta R.T.: -0.004 min
Response: 677001
Conc: 519.38 ng/ml



#33 AR-1260-3


R.T.: 7.136 min
Delta R.T.: -0.003 min
Response: 1196846
Conc: 497.72 ng/ml



#34 AR-1260-4


R.T.: 8.793 min
Delta R.T.: -0.004 min
Response: 766462
Conc: 498.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.619 min
Delta R.T.: -0.003 min
Response: 1060748
Conc: 504.71 ng/ml



#35 AR-1260-5

R.T.: 9.117 min
Delta R.T.: -0.004 min
Response: 1522448
Conc: 482.61 ng/ml



#35 AR-1260-5

R.T.: 7.865 min
Delta R.T.: -0.004 min
Response: 2477626
Conc: 489.41 ng/ml