

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040819\
 Data File : P0055044.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Apr 2019 15:48
 Operator : SM/SJ
 Sample : K1560-08
 Misc : AR1262 LOQ 100PPB
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOQ-WATER-02-QT1-2019

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 09 06:57:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.299	3.611	1402713	1313043	40.418	41.145
2)	SA Decachlor...	9.919	8.582	1604831	1141438	32.554	35.840
Target Compounds							
3)	L1 AR-1016-1	5.462	4.707	11904	12251	6.616	7.747
4)	L1 AR-1016-2	5.484	4.707	14875	12251	5.989	5.718
5)	L1 AR-1016-3	5.541	4.923	16897	9824	10.740	8.131
6)	L1 AR-1016-4	0.000	4.923	0	9824	N.D.	9.721 #
7)	L1 AR-1016-5	5.935	0.000	8973	0	6.725	N.D. #
9)	L2 AR-1221-2	4.602	3.909	21082	20556	62.168	68.569
12)	L3 AR-1232-2	0.000	4.707	0	12251	N.D.	14.227 #
13)	L3 AR-1232-3	5.484	4.923	14875	9824	15.221	20.274 #
14)	L3 AR-1232-4	0.000	4.923	0	9824	N.D.	22.357 #
15)	L3 AR-1232-5	5.733	0.000	11797	0	28.357	N.D. #
16)	L4 AR-1242-1	5.462	4.707	11904	12251	8.568	9.979
17)	L4 AR-1242-2	5.484	4.707	14875	12251	7.774	7.393
18)	L4 AR-1242-3	5.541	4.923	16897	9824	13.633	10.370
19)	L4 AR-1242-4	0.000	4.923	0	9824	N.D.	10.279 #
20)	L4 AR-1242-5	6.375	5.511	14782	30215	12.245	24.567 #
21)	L5 AR-1248-1	5.462	4.707	11904	12251	11.908	13.957
22)	L5 AR-1248-2	5.733	4.923	11797	9824	8.123	8.129
23)	L5 AR-1248-3	5.935	4.923	8973	9824	5.618	7.873 #
24)	L5 AR-1248-4	6.338	0.000	17812	0	9.482	N.D. #
25)	L5 AR-1248-5	6.375	5.511	14782	30215	8.245	20.266 #
26)	L6 AR-1254-1	6.311	5.511	125738	30215	69.347	13.825 #
27)	L6 AR-1254-2	6.527	5.630	94088	72171	32.959	37.222
28)	L6 AR-1254-3	6.895	6.048	66166	265189	21.041	82.724 #
29)	L6 AR-1254-4	7.177	6.288	54047	55280	22.093	28.234 #
30)	L6 AR-1254-5	7.592	6.674	439269	308880	163.888	107.448 #
31)	L7 AR-1260-1	7.056	6.161	434887	369233	173.079	160.671
32)	L7 AR-1260-2	7.312	6.348	496187	432411	159.878	157.083
33)	L7 AR-1260-3	7.670	6.504	641035	421252	261.182	163.409 #
34)	L7 AR-1260-4	7.894	6.971	596610	498654	215.693	234.312
35)	L7 AR-1260-5	8.204	7.211	1010910	887868	182.192	185.386
36)	L8 AR-1262-1	7.670	6.713	641035	570963	192.384	191.923
37)	L8 AR-1262-2	8.204	7.211	1010910	887868	170.635	181.155
38)	L8 AR-1262-3	8.501	7.493	721569	365670	174.174	178.076
39)	L8 AR-1262-4	8.587	7.557	539000	665930	167.289	183.851
40)	L8 AR-1262-5	9.212	8.049	364772	310299	166.847	192.269
41)	L9 AR-1268-1	8.501	7.493	721569	365670	101.380	64.673 #
42)	L9 AR-1268-2	8.587	7.557	539000	665930	82.265	128.767 #
43)	L9 AR-1268-3	8.804	7.764	36796	28300	6.755	6.576
44)	L9 AR-1268-4	9.212	8.049	364772	310299	152.894	167.381
45)	L9 AR-1268-5	9.601	8.333	144976	75955	8.800	6.673

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0040819\
Data File : P0055044.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Apr 2019 15:48
Operator : SM/SJ
Sample : K1560-08
Misc : AR1262 LOQ 100PPB
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 06:57:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO040819\
Data File : PO055044.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Apr 2019 15:48
Operator : SM/SJ
Sample : K1560-08
Misc : AR1262 LOQ 100PPB
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 09 06:57:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.299 min
Delta R.T.: -0.004 min
Response: 1402713
Conc: 40.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#1 Tetrachloro-m-xylene

R.T.: 3.611 min
Delta R.T.: -0.010 min
Response: 1313043
Conc: 41.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.919 min
Delta R.T.: -0.005 min
Response: 1604831
Conc: 32.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.582 min
Delta R.T.: -0.022 min
Response: 1141438
Conc: 35.84 ng/ml



#3 AR-1016-1


R.T.: 5.462 min
Delta R.T.: -0.003 min
Response: 11904
Conc: 6.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#3 AR-1016-1

R.T.: 4.707 min
Delta R.T.: 0.006 min
Response: 12251
Conc: 7.75 ng/ml




#4 AR-1016-2

R.T.: 5.484 min
Delta R.T.: -0.003 min
Response: 14875
Conc: 5.99 ng/ml



#4 AR-1016-2


R.T.: 4.707 min
Delta R.T.: -0.014 min
Response: 12251
Conc: 5.72 ng/ml



#5 AR-1016-3


R.T.: 5.541 min
Delta R.T.: -0.008 min
Response: 16897
Conc: 10.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#5 AR-1016-3

R.T.: 4.923 min
Delta R.T.: 0.027 min
Response: 9824
Conc: 8.13 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T. : 5.647 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 4.923 min
Delta R.T.: -0.013 min
Response: 9824
Conc: 9.72 ng/ml



#7 AR-1016-5


R.T.: 5.935 min
Delta R.T.: -0.005 min
Response: 8973
Conc: 6.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.602 min
Delta R.T.: 0.004 min
Response: 21082
Conc: 62.17 ng/ml



#9 AR-1221-2


R.T.: 3.909 min
Delta R.T.: -0.011 min
Response: 20556
Conc: 68.57 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T.: 5.200 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#12 AR-1232-2

R.T.: 4.707 min
Delta R.T.: -0.014 min
Response: 12251
Conc: 14.23 ng/ml




#13 AR-1232-3

R.T.: 5.484 min
Delta R.T.: -0.004 min
Response: 14875
Conc: 15.22 ng/ml



#13 AR-1232-3


R.T.: 4.923 min
Delta R.T.: 0.026 min
Response: 9824
Conc: 20.27 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 5.648 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#14 AR-1232-4

R.T.: 4.923 min
Delta R.T.: -0.056 min
Response: 9824
Conc: 22.36 ng/ml




#15 AR-1232-5

R.T.: 5.733 min
Delta R.T.: -0.006 min
Response: 11797
Conc: 28.36 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.150 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.462 min
Delta R.T.: -0.003 min
Response: 11904
Conc: 8.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#16 AR-1242-1

R.T.: 4.707 min
Delta R.T.: 0.006 min
Response: 12251
Conc: 9.98 ng/ml




#17 AR-1242-2

R.T.: 5.484 min
Delta R.T.: -0.003 min
Response: 14875
Conc: 7.77 ng/ml



#17 AR-1242-2


R.T.: 4.707 min
Delta R.T.: -0.014 min
Response: 12251
Conc: 7.39 ng/ml



#18 AR-1242-3


R.T.: 5.541 min
Delta R.T.: -0.008 min
Response: 16897
Conc: 13.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#18 AR-1242-3

R.T.: 4.923 min
Delta R.T.: 0.027 min
Response: 9824
Conc: 10.37 ng/ml




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.647 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 4.923 min
Delta R.T.: -0.055 min
Response: 9824
Conc: 10.28 ng/ml



#20 AR-1242-5


R.T.: 6.375 min
Delta R.T.: -0.004 min
Response: 14782
Conc: 12.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#20 AR-1242-5

R.T.: 5.511 min
Delta R.T.: 0.013 min
Response: 30215
Conc: 24.57 ng/ml




#21 AR-1248-1

R.T.: 5.462 min
Delta R.T.: -0.004 min
Response: 11904
Conc: 11.91 ng/ml



#21 AR-1248-1


R.T.: 4.707 min
Delta R.T.: 0.006 min
Response: 12251
Conc: 13.96 ng/ml



#22 AR-1248-2


R.T.: 5.733 min
Delta R.T.: -0.005 min
Response: 11797
Conc: 8.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#22 AR-1248-2

R.T.: 4.923 min
Delta R.T.: -0.013 min
Response: 9824
Conc: 8.13 ng/ml




#23 AR-1248-3

R.T.: 5.935 min
Delta R.T.: -0.005 min
Response: 8973
Conc: 5.62 ng/ml



#23 AR-1248-3


R.T.: 4.923 min
Delta R.T.: -0.055 min
Response: 9824
Conc: 7.87 ng/ml



#24 AR-1248-4


R.T.: 6.338 min
Delta R.T.: -0.004 min
Response: 17812
Conc: 9.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.375 min
Delta R.T.: -0.005 min
Response: 14782
Conc: 8.25 ng/ml



#25 AR-1248-5


R.T.: 5.511 min
Delta R.T.: -0.029 min
Response: 30215
Conc: 20.27 ng/ml



#26 AR-1254-1


R.T.: 6.311 min
Delta R.T.: -0.003 min
Response: 125738
Conc: 69.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#26 AR-1254-1

R.T.: 5.511 min
Delta R.T.: 0.011 min
Response: 30215
Conc: 13.83 ng/ml




#27 AR-1254-2

R.T.: 6.527 min
Delta R.T.: -0.006 min
Response: 94088
Conc: 32.96 ng/ml



#27 AR-1254-2


R.T.: 5.630 min
Delta R.T.: -0.016 min
Response: 72171
Conc: 37.22 ng/ml



#28 AR-1254-3


R.T.: 6.895 min
Delta R.T.: -0.002 min
Response: 66166
Conc: 21.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#28 AR-1254-3

R.T.: 6.048 min
Delta R.T.: -0.001 min
Response: 265189
Conc: 82.72 ng/ml




#29 AR-1254-4

R.T.: 7.177 min
Delta R.T.: -0.005 min
Response: 54047
Conc: 22.09 ng/ml



#29 AR-1254-4


R.T.: 6.288 min
Delta R.T.: 0.012 min
Response: 55280
Conc: 28.23 ng/ml



#30 AR-1254-5


R.T.: 7.592 min
Delta R.T.: -0.008 min
Response: 439269
Conc: 163.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#30 AR-1254-5

R.T.: 6.674 min
Delta R.T.: -0.018 min
Response: 308880
Conc: 107.45 ng/ml




#31 AR-1260-1

R.T.: 7.056 min
Delta R.T.: -0.004 min
Response: 434887
Conc: 173.08 ng/ml



#31 AR-1260-1


R.T.: 6.161 min
Delta R.T.: -0.016 min
Response: 369233
Conc: 160.67 ng/ml



#32 AR-1260-2


R.T.: 7.312 min
Delta R.T.: -0.004 min
Response: 496187
Conc: 159.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#32 AR-1260-2

R.T.: 6.348 min
Delta R.T.: -0.016 min
Response: 432411
Conc: 157.08 ng/ml




#33 AR-1260-3

R.T.: 7.670 min
Delta R.T.: -0.004 min
Response: 641035
Conc: 261.18 ng/ml



#33 AR-1260-3


R.T.: 6.504 min
Delta R.T.: -0.014 min
Response: 421252
Conc: 163.41 ng/ml



#34 AR-1260-4


R.T.: 7.894 min
Delta R.T.: -0.003 min
Response: 596610
Conc: 215.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#34 AR-1260-4

R.T.: 6.971 min
Delta R.T.: -0.016 min
Response: 498654
Conc: 234.31 ng/ml




#35 AR-1260-5

R.T.: 8.204 min
Delta R.T.: -0.003 min
Response: 1010910
Conc: 182.19 ng/ml



#35 AR-1260-5


R.T.: 7.211 min
Delta R.T.: -0.016 min
Response: 887868
Conc: 185.39 ng/ml



#36 AR-1262-1


R.T.: 7.670 min
Delta R.T.: -0.005 min
Response: 641035
Conc: 192.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#36 AR-1262-1

R.T.: 6.713 min
Delta R.T.: -0.017 min
Response: 570963
Conc: 191.92 ng/ml




#37 AR-1262-2

R.T.: 8.204 min
Delta R.T.: -0.004 min
Response: 1010910
Conc: 170.63 ng/ml



#37 AR-1262-2


R.T.: 7.211 min
Delta R.T.: -0.018 min
Response: 887868
Conc: 181.16 ng/ml



#38 AR-1262-3


R.T.: 8.501 min
Delta R.T.: -0.006 min
Response: 721569
Conc: 174.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#38 AR-1262-3

R.T.: 7.493 min
Delta R.T.: -0.018 min
Response: 365670
Conc: 178.08 ng/ml




#39 AR-1262-4

R.T.: 8.587 min
Delta R.T.: -0.006 min
Response: 539000
Conc: 167.29 ng/ml



#39 AR-1262-4


R.T.: 7.557 min
Delta R.T.: -0.018 min
Response: 665930
Conc: 183.85 ng/ml



#40 AR-1262-5


R.T.: 9.212 min
Delta R.T.: -0.006 min
Response: 364772
Conc: 166.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#40 AR-1262-5

R.T.: 8.049 min
Delta R.T.: -0.017 min
Response: 310299
Conc: 192.27 ng/ml




#41 AR-1268-1

R.T.: 8.501 min
Delta R.T.: -0.003 min
Response: 721569
Conc: 101.38 ng/ml



#41 AR-1268-1


R.T.: 7.493 min
Delta R.T.: -0.018 min
Response: 365670
Conc: 64.67 ng/ml



#42 AR-1268-2


R.T.: 8.587 min
Delta R.T.: -0.007 min
Response: 539000
Conc: 82.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#42 AR-1268-2

R.T.: 7.557 min
Delta R.T.: -0.018 min
Response: 665930
Conc: 128.77 ng/ml




#43 AR-1268-3

R.T.: 8.804 min
Delta R.T.: -0.006 min
Response: 36796
Conc: 6.75 ng/ml



#43 AR-1268-3


R.T.: 7.764 min
Delta R.T.: -0.019 min
Response: 28300
Conc: 6.58 ng/ml



#44 AR-1268-4


R.T.: 9.212 min
Delta R.T.: -0.005 min
Response: 364772
Conc: 152.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT1-2019




#44 AR-1268-4

R.T.: 8.049 min
Delta R.T.: -0.016 min
Response: 310299
Conc: 167.38 ng/ml



#45 AR-1268-5

R.T.: 9.601 min
Delta R.T.: -0.007 min
Response: 144976
Conc: 8.80 ng/ml



#45 AR-1268-5

R.T.: 8.333 min
Delta R.T.: -0.019 min
Response: 75955
Conc: 6.67 ng/ml