

Data Path : P:\HPCHEM1\ECD 0\Data\P0041018\
 Data File : P0043747.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Apr 2018 17:56
 Operator : SM/UA
 Sample : J2157-07
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CL-01-040618-D

Manual Integrations
 APPROVED

ugo
 4/16/2018 10:13:19 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 04:33:29 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0040418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 04 15:17:57 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.063	3.323	4316901	5949387	24.283	22.143
2) SA Decachlor...	9.396	8.080	2342769	2206959	13.000m	9.414m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO041018\
Data File : PO043747.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Apr 2018 17:56
Operator : SM/UA
Sample : J2157-07
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CL-01-040618-D


Manual Integrations
APPROVED

ugo
4/16/2018 10:13:19 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 04:33:29 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\PO040418.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 04 15:17:57 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.063 min
Delta R.T.: -0.006 min
Response: 4316901
Conc: 24.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
CL-01-040618-D


Manual Integrations
APPROVED

ugo
4/16/2018 10:13:19 AM




#1 Tetrachloro-m-xylene

R.T.: 3.323 min
Delta R.T.: -0.003 min
Response: 5949387
Conc: 22.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.396 min
Delta R.T.: -0.021 min
Response: 2342769
Conc: 13.00 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.080 min
Delta R.T.: -0.011 min
Response: 2206959
Conc: 9.41 ng/ml m