

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041119\
 Data File : P0055145.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Apr 2019 12:01
 Operator : SM/SJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 02:29:46 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.293	3.606	916663	896443	26.413	28.091
2)	SA Decachlor...	9.904	8.573	1155642	755416	23.442	23.719
Target Compounds							
7)	L1 AR-1016-5	5.924	0.000	3118	0	2.337	N.D. #
8)	L2 AR-1221-1	0.000	3.903	0	15459	N.D.	39.763 #
9)	L2 AR-1221-2	4.596	3.903	15191	15459	44.796	51.567
20)	L4 AR-1242-5	0.000	5.504	0	14273	N.D.	11.605 #
23)	L5 AR-1248-3	5.924	0.000	3118	0	1.952	N.D. #
24)	L5 AR-1248-4	6.304	0.000	32215	0	17.149	N.D. #
25)	L5 AR-1248-5	0.000	5.504	0	14273	N.D.	9.573 #
26)	L6 AR-1254-1	6.304	5.504	32215	14273	17.767	6.531 #
27)	L6 AR-1254-2	6.474	0.000	16477	0	5.772	N.D. #
28)	L6 AR-1254-3	6.894	0.000	21264	0	6.762	N.D. #
29)	L6 AR-1254-4	0.000	6.330	0	21154	N.D.	10.804 #
30)	L6 AR-1254-5	7.585	0.000	10452	0	3.900	N.D. #
32)	L7 AR-1260-2	7.371	6.330	35677	21154	11.496	7.685 #
33)	L7 AR-1260-3	7.666	6.550	4873	25323	1.986	9.823 #
35)	L7 AR-1260-5	8.196	0.000	11343	0	2.044	N.D. #
36)	L8 AR-1262-1	7.666	0.000	4873	0	1.463	N.D. #
37)	L8 AR-1262-2	8.196	0.000	11343	0	1.915	N.D. #
38)	L8 AR-1262-3	0.000	7.550	0	2496	N.D.	1.215 #
39)	L8 AR-1262-4	0.000	7.550	0	2496	N.D.	0.689 #
41)	L9 AR-1268-1	0.000	7.550	0	2496	N.D.	0.441 #
42)	L9 AR-1268-2	0.000	7.550	0	2496	N.D.	0.483 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041119\
Data File : PO055145.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Apr 2019 12:01
Operator : SM/SJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 02:29:46 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.293 min
Delta R.T.: -0.010 min
Response: 916663
Conc: 26.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: -0.015 min
Response: 896443
Conc: 28.09 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.904 min
Delta R.T.: -0.019 min
Response: 1155642
Conc: 23.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.573 min
Delta R.T.: -0.031 min
Response: 755416
Conc: 23.72 ng/ml



#7 AR-1016-5


R.T.: 5.924 min
Delta R.T.: -0.015 min
Response: 3118
Conc: 2.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.512 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.903 min
Delta R.T.: 0.068 min
Response: 15459
Conc: 39.76 ng/ml



#9 AR-1221-2


R.T.: 4.596 min
Delta R.T.: -0.002 min
Response: 15191
Conc: 44.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 3.903 min
Delta R.T.: -0.016 min
Response: 15459
Conc: 51.57 ng/ml




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 6.379 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 5.504 min
Delta R.T.: 0.006 min
Response: 14273
Conc: 11.60 ng/ml



#23 AR-1248-3


R.T.: 5.924 min
Delta R.T.: -0.016 min
Response: 3118
Conc: 1.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.977 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.304 min
Delta R.T.: -0.038 min
Response: 32215
Conc: 17.15 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T.: 6.380 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#25 AR-1248-5

R.T.: 5.504 min
Delta R.T.: -0.036 min
Response: 14273
Conc: 9.57 ng/ml




#26 AR-1254-1

R.T.: 6.304 min
Delta R.T.: -0.010 min
Response: 32215
Conc: 17.77 ng/ml



#26 AR-1254-1


R.T.: 5.504 min
Delta R.T.: 0.005 min
Response: 14273
Conc: 6.53 ng/ml



#27 AR-1254-2


R.T.: 6.474 min
Delta R.T.: -0.059 min
Response: 16477
Conc: 5.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#27 AR-1254-2

R.T.: 0.000 min
Exp R.T. : 5.646 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 6.894 min
Delta R.T.: -0.003 min
Response: 21264
Conc: 6.76 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.049 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T.: 7.182 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 6.330 min
Delta R.T.: 0.054 min
Response: 21154
Conc: 10.80 ng/ml




#30 AR-1254-5

R.T.: 7.585 min
Delta R.T.: -0.014 min
Response: 10452
Conc: 3.90 ng/ml



#30 AR-1254-5


R.T.: 0.000 min
Exp R.T.: 6.692 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 7.371 min
Delta R.T.: 0.054 min
Response: 35677
Conc: 11.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#32 AR-1260-2

R.T.: 6.330 min
Delta R.T.: -0.034 min
Response: 21154
Conc: 7.68 ng/ml




#33 AR-1260-3

R.T.: 7.666 min
Delta R.T.: -0.007 min
Response: 4873
Conc: 1.99 ng/ml



#33 AR-1260-3


R.T.: 6.550 min
Delta R.T.: 0.032 min
Response: 25323
Conc: 9.82 ng/ml



#35 AR-1260-5


R.T.: 8.196 min
Delta R.T.: -0.011 min
Response: 11343
Conc: 2.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.227 min
Response: 0
Conc: N.D.




#36 AR-1262-1

R.T.: 7.666 min
Delta R.T.: -0.008 min
Response: 4873
Conc: 1.46 ng/ml



#36 AR-1262-1


R.T.: 0.000 min
Exp R.T. : 6.730 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 8.196 min
Delta R.T.: -0.012 min
Response: 11343
Conc: 1.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.229 min
Response: 0
Conc: N.D.




#38 AR-1262-3

R.T.: 0.000 min
Exp R.T. : 8.507 min
Response: 0
Conc: N.D.



#38 AR-1262-3


R.T.: 7.550 min
Delta R.T.: 0.039 min
Response: 2496
Conc: 1.22 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 8.593 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK




#39 AR-1262-4

R.T.: 7.550 min
Delta R.T.: -0.025 min
Response: 2496
Conc: 0.69 ng/ml




#41 AR-1268-1

R.T.: 0.000 min
Exp R.T.: 8.504 min
Response: 0
Conc: N.D.



#41 AR-1268-1


R.T.: 7.550 min
Delta R.T.: 0.039 min
Response: 2496
Conc: 0.44 ng/ml



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T.: 8.594 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
I.BLK



#42 AR-1268-2

R.T.: 7.550 min
Delta R.T.: -0.025 min
Response: 2496
Conc: 0.48 ng/ml