

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041119\
 Data File : P0055177.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 Apr 2019 23:00
 Operator : SM/SJ
 Sample : K2311-07
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 0240-COMP-CS-5-ELTRT

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 12 02:47:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.293	3.606	524963	489892	15.126	15.351
2)	SA Decachlor...	9.904	8.571	525004	380164	10.650	11.937
Target Compounds							
7)	L1 AR-1016-5	5.917	0.000	33686	0	25.247	N.D. #
8)	L2 AR-1221-1	0.000	3.904	0	12804	N.D.	32.933 #
9)	L2 AR-1221-2	0.000	3.904	0	12804	N.D.	42.710 #
10)	L2 AR-1221-3	4.685	0.000	11802	0	11.361	N.D. #
11)	L3 AR-1232-1	4.685	0.000	11802	0	13.708	N.D. #
12)	L3 AR-1232-2	5.166	0.000	4095	0	8.371	N.D. #
23)	L5 AR-1248-3	5.917	0.000	33686	0	21.091	N.D. #
29)	L6 AR-1254-4	0.000	6.330	0	10375	N.D.	5.299 #
32)	L7 AR-1260-2	7.270	6.330	24323	10375	7.837	3.769 #
33)	L7 AR-1260-3	0.000	6.511	0	20874	N.D.	8.097 #
35)	L7 AR-1260-5	8.141	0.000	11953	0	2.154	N.D. #
37)	L8 AR-1262-2	8.141	0.000	11953	0	2.018	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041119\
Data File : PO055177.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 Apr 2019 23:00
Operator : SM/SJ
Sample : K2311-07
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 12 02:47:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.293 min
Delta R.T.: -0.010 min
Response: 524963
Conc: 15.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: -0.015 min
Response: 489892
Conc: 15.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.904 min
Delta R.T.: -0.019 min
Response: 525004
Conc: 10.65 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.571 min
Delta R.T.: -0.033 min
Response: 380164
Conc: 11.94 ng/ml



#7 AR-1016-5


R.T.: 5.917 min
Delta R.T.: -0.022 min
Response: 33686
Conc: 25.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 4.512 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.904 min
Delta R.T.: 0.068 min
Response: 12804
Conc: 32.93 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T.: 4.598 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#9 AR-1221-2

R.T.: 3.904 min
Delta R.T.: -0.016 min
Response: 12804
Conc: 42.71 ng/ml




#10 AR-1221-3

R.T.: 4.685 min
Delta R.T.: 0.012 min
Response: 11802
Conc: 11.36 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T.: 3.996 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.685 min
Delta R.T.: 0.011 min
Response: 11802
Conc: 13.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.997 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.166 min
Delta R.T.: -0.034 min
Response: 4095
Conc: 8.37 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.721 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 5.917 min
Delta R.T.: -0.023 min
Response: 33686
Conc: 21.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.977 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 7.182 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.330 min
Delta R.T.: 0.054 min
Response: 10375
Conc: 5.30 ng/ml



#32 AR-1260-2


R.T.: 7.270 min
Delta R.T.: -0.046 min
Response: 24323
Conc: 7.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#32 AR-1260-2

R.T.: 6.330 min
Delta R.T.: -0.034 min
Response: 10375
Conc: 3.77 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 7.674 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 6.511 min
Delta R.T.: -0.007 min
Response: 20874
Conc: 8.10 ng/ml



#35 AR-1260-5


R.T.: 8.141 min
Delta R.T.: -0.067 min
Response: 11953
Conc: 2.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
0240-COMP-CS-5-ELTRT




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.227 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 8.141 min
Delta R.T.: -0.068 min
Response: 11953
Conc: 2.02 ng/ml



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.229 min
Response: 0
Conc: N.D.