

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041219\
 Data File : P0055257.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Apr 2019 23:38
 Operator : SM/SJ
 Sample : PB118841BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB118841BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 00:50:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.291	3.604	631179	595023	18.187	18.646
2)	SA Decachlor...	9.900	8.567	859830	642879	17.442	20.186
Target Compounds							
3)	L1 AR-1016-1	5.475	4.699	429540	376281	238.709	237.942
4)	L1 AR-1016-2	5.475	4.729	429540	27121	172.945	12.657 #
5)	L1 AR-1016-3	5.536	4.914	278289	154056	176.888	127.506 #
6)	L1 AR-1016-4	5.634	4.955	227545	198648	174.621	196.577
7)	L1 AR-1016-5	5.926	5.166	232874	66366	174.534	50.918 #
8)	L2 AR-1221-1	4.497	3.815	24846	21984	56.911	56.545
9)	L2 AR-1221-2	4.586	3.901	37267	32166	109.893	107.294
10)	L2 AR-1221-3	4.663	3.976	142538	122612	137.217	133.016
11)	L3 AR-1232-1	4.663	3.976	142538	122612	165.557	163.506
12)	L3 AR-1232-2	5.187	4.729	210092	27121	429.515	31.496 #
13)	L3 AR-1232-3	5.475	4.914	429540	154056	439.515	317.926 #
14)	L3 AR-1232-4	5.634	4.986	227545	65917	444.551	150.015 #
15)	L3 AR-1232-5	5.724	5.166	212652	66366	511.186	136.455 #
16)	L4 AR-1242-1	5.475	4.699	429540	376281	309.179	306.490
17)	L4 AR-1242-2	5.475	4.729	429540	27121	224.482	16.366 #
18)	L4 AR-1242-3	5.536	4.914	278289	154056	224.526	162.621 #
19)	L4 AR-1242-4	5.634	4.986	227545	65917	223.182	68.968 #
20)	L4 AR-1242-5	6.366	5.515	35953	28796	29.783	23.413
21)	L5 AR-1248-1	5.475	4.699	429540	376281	429.679	428.675
22)	L5 AR-1248-2	5.724	4.955	212652	198648	146.424	164.369
23)	L5 AR-1248-3	5.926	4.986	232874	65917	145.804	52.827 #
24)	L5 AR-1248-4	6.327	5.166	38880	66366	20.697	43.577 #
25)	L5 AR-1248-5	6.366	5.515	35953	28796	20.054	19.315
26)	L6 AR-1254-1	6.327	5.515	38880	28796	21.443	13.176 #
27)	L6 AR-1254-2	6.518	5.621	210097	178235	73.596	91.923
28)	L6 AR-1254-3	6.883	6.037	122084	320828	38.822	100.081 #
29)	L6 AR-1254-4	7.167	6.278	77750	109884	31.782	56.122 #
30)	L6 AR-1254-5	7.614	6.701	86073	399650	32.113	139.023 #
31)	L7 AR-1260-1	7.045	6.151	496807	449080	197.723	195.416
32)	L7 AR-1260-2	7.302	6.338	593128	546063	191.114	198.369
33)	L7 AR-1260-3	7.659	6.491	388897	507563	158.451	196.890
34)	L7 AR-1260-4	7.883	6.999	472624	51785	170.868	24.333 #
35)	L7 AR-1260-5	8.193	7.200	841118	820713	151.591	171.364
36)	L8 AR-1262-1	7.659	6.754	388897	208928	116.713	70.229 #
37)	L8 AR-1262-2	8.193	7.256	841118	17201	141.975	3.510 #
38)	L8 AR-1262-3	8.504	7.482	568660	166284	137.265	80.978 #
39)	L8 AR-1262-4	8.556	7.546	305304	583535	94.757	161.103 #
40)	L8 AR-1262-5	9.198	8.081	204031	9810	93.324	6.079 #
41)	L9 AR-1268-1	8.504	7.482	568660	166284	79.896	29.409 #
42)	L9 AR-1268-2	8.556	7.546	305304	583535	46.597	112.835 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041219\
 Data File : P0055257.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Apr 2019 23:38
 Operator : SM/SJ
 Sample : PB118841BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB118841BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 00:50:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0032519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Mar 25 17:22:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	8.790	7.800	17214	12102	3.160	2.812
44) L9	AR-1268-4	9.198	8.081	204031	9810	85.519	5.292 #
45) L9	AR-1268-5	9.586	0.000	71019	0	4.311	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041219\
Data File : PO055257.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Apr 2019 23:38
Operator : SM/SJ
Sample : PB118841BS
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB118841BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 00:50:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO032519.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Mar 25 17:22:26 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.291 min
Delta R.T.: -0.012 min
Response: 631179
Conc: 18.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#1 Tetrachloro-m-xylene

R.T.: 3.604 min
Delta R.T.: -0.017 min
Response: 595023
Conc: 18.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.900 min
Delta R.T.: -0.023 min
Response: 859830
Conc: 17.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.567 min
Delta R.T.: -0.037 min
Response: 642879
Conc: 20.19 ng/ml



#3 AR-1016-1


R.T.: 5.475 min
Delta R.T.: 0.009 min
Response: 429540
Conc: 238.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#3 AR-1016-1

R.T.: 4.699 min
Delta R.T.: -0.002 min
Response: 376281
Conc: 237.94 ng/ml




#4 AR-1016-2

R.T.: 5.475 min
Delta R.T.: -0.012 min
Response: 429540
Conc: 172.95 ng/ml



#4 AR-1016-2


R.T.: 4.729 min
Delta R.T.: 0.008 min
Response: 27121
Conc: 12.66 ng/ml



#5 AR-1016-3


R.T.: 5.536 min
Delta R.T.: -0.012 min
Response: 278289
Conc: 176.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#5 AR-1016-3

R.T.: 4.914 min
Delta R.T.: 0.018 min
Response: 154056
Conc: 127.51 ng/ml




#6 AR-1016-4

R.T.: 5.634 min
Delta R.T.: -0.013 min
Response: 227545
Conc: 174.62 ng/ml



#6 AR-1016-4


R.T.: 4.955 min
Delta R.T.: 0.020 min
Response: 198648
Conc: 196.58 ng/ml



#7 AR-1016-5


R.T.: 5.926 min
Delta R.T.: -0.014 min
Response: 232874
Conc: 174.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#7 AR-1016-5

R.T.: 5.166 min
Delta R.T.: 0.017 min
Response: 66366
Conc: 50.92 ng/ml




#8 AR-1221-1

R.T.: 4.497 min
Delta R.T.: -0.015 min
Response: 24846
Conc: 56.91 ng/ml



#8 AR-1221-1


R.T.: 3.815 min
Delta R.T.: -0.021 min
Response: 21984
Conc: 56.54 ng/ml



#9 AR-1221-2


R.T.: 4.586 min
Delta R.T.: -0.012 min
Response: 37267
Conc: 109.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#9 AR-1221-2

R.T.: 3.901 min
Delta R.T.: -0.018 min
Response: 32166
Conc: 107.29 ng/ml




#10 AR-1221-3

R.T.: 4.663 min
Delta R.T.: -0.011 min
Response: 142538
Conc: 137.22 ng/ml



#10 AR-1221-3


R.T.: 3.976 min
Delta R.T.: -0.020 min
Response: 122612
Conc: 133.02 ng/ml



#11 AR-1232-1


R.T.: 4.663 min
Delta R.T.: -0.012 min
Response: 142538
Conc: 165.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#11 AR-1232-1

R.T.: 3.976 min
Delta R.T.: -0.021 min
Response: 122612
Conc: 163.51 ng/ml




#12 AR-1232-2

R.T.: 5.187 min
Delta R.T.: -0.013 min
Response: 210092
Conc: 429.52 ng/ml



#12 AR-1232-2


R.T.: 4.729 min
Delta R.T.: 0.008 min
Response: 27121
Conc: 31.50 ng/ml



#13 AR-1232-3


R.T.: 5.475 min
Delta R.T.: -0.014 min
Response: 429540
Conc: 439.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#13 AR-1232-3

R.T.: 4.914 min
Delta R.T.: 0.017 min
Response: 154056
Conc: 317.93 ng/ml




#14 AR-1232-4

R.T.: 5.634 min
Delta R.T.: -0.014 min
Response: 227545
Conc: 444.55 ng/ml



#14 AR-1232-4


R.T.: 4.986 min
Delta R.T.: 0.007 min
Response: 65917
Conc: 150.01 ng/ml



#15 AR-1232-5


R.T.: 5.724 min
Delta R.T.: -0.015 min
Response: 212652
Conc: 511.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#15 AR-1232-5

R.T.: 5.166 min
Delta R.T.: 0.016 min
Response: 66366
Conc: 136.45 ng/ml




#16 AR-1242-1

R.T.: 5.475 min
Delta R.T.: 0.010 min
Response: 429540
Conc: 309.18 ng/ml



#16 AR-1242-1


R.T.: 4.699 min
Delta R.T.: -0.002 min
Response: 376281
Conc: 306.49 ng/ml



#17 AR-1242-2


R.T.: 5.475 min
Delta R.T.: -0.012 min
Response: 429540
Conc: 224.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#17 AR-1242-2

R.T.: 4.729 min
Delta R.T.: 0.009 min
Response: 27121
Conc: 16.37 ng/ml




#18 AR-1242-3

R.T.: 5.536 min
Delta R.T.: -0.013 min
Response: 278289
Conc: 224.53 ng/ml



#18 AR-1242-3


R.T.: 4.914 min
Delta R.T.: 0.018 min
Response: 154056
Conc: 162.62 ng/ml



#19 AR-1242-4


R.T.: 5.634 min
Delta R.T.: -0.013 min
Response: 227545
Conc: 223.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#19 AR-1242-4

R.T.: 4.986 min
Delta R.T.: 0.009 min
Response: 65917
Conc: 68.97 ng/ml




#20 AR-1242-5

R.T.: 6.366 min
Delta R.T.: -0.013 min
Response: 35953
Conc: 29.78 ng/ml



#20 AR-1242-5


R.T.: 5.515 min
Delta R.T.: 0.017 min
Response: 28796
Conc: 23.41 ng/ml



#21 AR-1248-1


R.T.: 5.475 min
Delta R.T.: 0.009 min
Response: 429540
Conc: 429.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#21 AR-1248-1

R.T.: 4.699 min
Delta R.T.: -0.003 min
Response: 376281
Conc: 428.68 ng/ml




#22 AR-1248-2

R.T.: 5.724 min
Delta R.T.: -0.014 min
Response: 212652
Conc: 146.42 ng/ml



#22 AR-1248-2


R.T.: 4.955 min
Delta R.T.: 0.019 min
Response: 198648
Conc: 164.37 ng/ml



#23 AR-1248-3


R.T.: 5.926 min
Delta R.T.: -0.014 min
Response: 232874
Conc: 145.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#23 AR-1248-3

R.T.: 4.986 min
Delta R.T.: 0.009 min
Response: 65917
Conc: 52.83 ng/ml




#24 AR-1248-4

R.T.: 6.327 min
Delta R.T.: -0.015 min
Response: 38880
Conc: 20.70 ng/ml



#24 AR-1248-4


R.T.: 5.166 min
Delta R.T.: 0.016 min
Response: 66366
Conc: 43.58 ng/ml



#25 AR-1248-5


R.T.: 6.366 min
Delta R.T.: -0.014 min
Response: 35953
Conc: 20.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#25 AR-1248-5

R.T.: 5.515 min
Delta R.T.: -0.025 min
Response: 28796
Conc: 19.31 ng/ml




#26 AR-1254-1

R.T.: 6.327 min
Delta R.T.: 0.013 min
Response: 38880
Conc: 21.44 ng/ml



#26 AR-1254-1


R.T.: 5.515 min
Delta R.T.: 0.015 min
Response: 28796
Conc: 13.18 ng/ml



#27 AR-1254-2


R.T.: 6.518 min
Delta R.T.: -0.015 min
Response: 210097
Conc: 73.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#27 AR-1254-2

R.T.: 5.621 min
Delta R.T.: -0.025 min
Response: 178235
Conc: 91.92 ng/ml




#28 AR-1254-3

R.T.: 6.883 min
Delta R.T.: -0.014 min
Response: 122084
Conc: 38.82 ng/ml



#28 AR-1254-3


R.T.: 6.037 min
Delta R.T.: -0.012 min
Response: 320828
Conc: 100.08 ng/ml



#29 AR-1254-4


R.T.: 7.167 min
Delta R.T.: -0.015 min
Response: 77750
Conc: 31.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#29 AR-1254-4

R.T.: 6.278 min
Delta R.T.: 0.002 min
Response: 109884
Conc: 56.12 ng/ml




#30 AR-1254-5

R.T.: 7.614 min
Delta R.T.: 0.015 min
Response: 86073
Conc: 32.11 ng/ml



#30 AR-1254-5


R.T.: 6.701 min
Delta R.T.: 0.009 min
Response: 399650
Conc: 139.02 ng/ml



#31 AR-1260-1


R.T.: 7.045 min
Delta R.T.: -0.014 min
Response: 496807
Conc: 197.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#31 AR-1260-1

R.T.: 6.151 min
Delta R.T.: -0.026 min
Response: 449080
Conc: 195.42 ng/ml




#32 AR-1260-2

R.T.: 7.302 min
Delta R.T.: -0.015 min
Response: 593128
Conc: 191.11 ng/ml



#32 AR-1260-2


R.T.: 6.338 min
Delta R.T.: -0.026 min
Response: 546063
Conc: 198.37 ng/ml



#33 AR-1260-3


R.T.: 7.659 min
Delta R.T.: -0.015 min
Response: 388897
Conc: 158.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#33 AR-1260-3

R.T.: 6.491 min
Delta R.T.: -0.027 min
Response: 507563
Conc: 196.89 ng/ml




#34 AR-1260-4

R.T.: 7.883 min
Delta R.T.: -0.014 min
Response: 472624
Conc: 170.87 ng/ml



#34 AR-1260-4


R.T.: 6.999 min
Delta R.T.: 0.012 min
Response: 51785
Conc: 24.33 ng/ml



#35 AR-1260-5


R.T.: 8.193 min
Delta R.T.: -0.014 min
Response: 841118
Conc: 151.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#35 AR-1260-5

R.T.: 7.200 min
Delta R.T.: -0.027 min
Response: 820713
Conc: 171.36 ng/ml




#36 AR-1262-1

R.T.: 7.659 min
Delta R.T.: -0.016 min
Response: 388897
Conc: 116.71 ng/ml



#36 AR-1262-1


R.T.: 6.754 min
Delta R.T.: 0.025 min
Response: 208928
Conc: 70.23 ng/ml



#37 AR-1262-2


R.T.: 8.193 min
Delta R.T.: -0.015 min
Response: 841118
Conc: 141.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#37 AR-1262-2

R.T.: 7.256 min
Delta R.T.: 0.027 min
Response: 17201
Conc: 3.51 ng/ml




#38 AR-1262-3

R.T.: 8.504 min
Delta R.T.: -0.003 min
Response: 568660
Conc: 137.26 ng/ml



#38 AR-1262-3


R.T.: 7.482 min
Delta R.T.: -0.030 min
Response: 166284
Conc: 80.98 ng/ml



#39 AR-1262-4


R.T.: 8.556 min
Delta R.T.: -0.037 min
Response: 305304
Conc: 94.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#39 AR-1262-4

R.T.: 7.546 min
Delta R.T.: -0.029 min
Response: 583535
Conc: 161.10 ng/ml




#40 AR-1262-5

R.T.: 9.198 min
Delta R.T.: -0.020 min
Response: 204031
Conc: 93.32 ng/ml



#40 AR-1262-5


R.T.: 8.081 min
Delta R.T.: 0.014 min
Response: 9810
Conc: 6.08 ng/ml



#41 AR-1268-1


R.T.: 8.504 min
Delta R.T.: 0.000 min
Response: 568660
Conc: 79.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#41 AR-1268-1

R.T.: 7.482 min
Delta R.T.: -0.030 min
Response: 166284
Conc: 29.41 ng/ml




#42 AR-1268-2

R.T.: 8.556 min
Delta R.T.: -0.038 min
Response: 305304
Conc: 46.60 ng/ml



#42 AR-1268-2


R.T.: 7.546 min
Delta R.T.: -0.030 min
Response: 583535
Conc: 112.83 ng/ml



#43 AR-1268-3


R.T.: 8.790 min
Delta R.T.: -0.020 min
Response: 17214
Conc: 3.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS




#43 AR-1268-3

R.T.: 7.800 min
Delta R.T.: 0.017 min
Response: 12102
Conc: 2.81 ng/ml




#44 AR-1268-4

R.T.: 9.198 min
Delta R.T.: -0.019 min
Response: 204031
Conc: 85.52 ng/ml



#44 AR-1268-4


R.T.: 8.081 min
Delta R.T.: 0.015 min
Response: 9810
Conc: 5.29 ng/ml



#45 AR-1268-5

R.T.: 9.586 min
Delta R.T.: -0.022 min
Response: 71019
Conc: 4.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB118841BS



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T. : 8.352 min
Response: 0
Conc: N.D.