

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041320\
 Data File : P0067808.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Apr 2020 10:22
 Operator : DD\AJ
 Sample : PB128181BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 11:25:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.945	3.945	572018	746861	24.020	22.283
2)	SA Decachlor...	10.925	9.233	764170	855699	22.447	20.696
Target Compounds							
28)	L6 AR-1254-3	7.797	0.000	43041	0	20.742	N.D. #
29)	L6 AR-1254-4	0.000	6.948	0	52082	N.D.	21.605 #
32)	L7 AR-1260-2	0.000	6.948	0	52082	N.D.	20.283 #
35)	L7 AR-1260-5	9.129	0.000	1623	0	0.515	N.D. #
37)	L8 AR-1262-2	9.129	0.000	1623	0	0.566	N.D. #
39)	L8 AR-1262-4	9.558	0.000	50002	0	31.600	N.D. #
42)	L9 AR-1268-2	9.558	0.000	50002	0	14.427	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041320\
Data File : P0067808.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 13 Apr 2020 10:22
Operator : DD\AJ
Sample : PB128181BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 11:25:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.945 min
Delta R.T.: 0.010 min
Response: 572018
Conc: 24.02 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.945 min
Delta R.T.: 0.002 min
Response: 746861
Conc: 22.28 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.925 min
Delta R.T.: 0.006 min
Response: 764170
Conc: 22.45 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.233 min
Delta R.T.: 0.000 min
Response: 855699
Conc: 20.70 ng/ml



#28 AR-1254-3


R.T.: 7.797 min
Delta R.T.: 0.024 min
Response: 43041
Conc: 20.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.649 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 8.068 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.948 min
Delta R.T.: 0.059 min
Response: 52082
Conc: 21.60 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 8.205 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.948 min
Delta R.T.: -0.037 min
Response: 52082
Conc: 20.28 ng/ml




#35 AR-1260-5

R.T.: 9.129 min
Delta R.T.: 0.008 min
Response: 1623
Conc: 0.51 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.869 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 9.129 min
Delta R.T.: 0.008 min
Response: 1623
Conc: 0.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.869 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 9.558 min
Delta R.T.: 0.037 min
Response: 50002
Conc: 31.60 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 8.219 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 9.558 min
Delta R.T.: 0.035 min
Response: 50002
Conc: 14.43 ng/ml

Instrument :
ECD_O
ClientSampleId :



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 8.221 min
Response: 0
Conc: N.D.