

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042220\
 Data File : P0067962.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Apr 2020 8:28
 Operator : DD\AJ
 Sample : L2343-04RE
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SB-4RE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 10:38:34 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.932	3.941	409708	657775	17.204	19.625
2)	SA Decachlor...	10.905	9.224	384344	440954	11.290	10.665
Target Compounds							
3)	L1 AR-1016-1	6.177	0.000	21406	0	20.758	N.D. #
10)	L2 AR-1221-3	0.000	4.445	0	347006	N.D.	353.460 #
11)	L3 AR-1232-1	0.000	4.445	0	347006	N.D.	501.406 #
20)	L4 AR-1242-5	7.193	6.090	33940	41787	63.531	44.657 #
24)	L5 AR-1248-4	7.193	0.000	33940	0	35.974	N.D. #
25)	L5 AR-1248-5	7.193	6.090	33940	41787	37.661	33.768
26)	L6 AR-1254-1	7.193	6.090	33940	41787	27.479	16.297 #
27)	L6 AR-1254-2	7.444	6.298	90401	267748	45.966	117.399 #
28)	L6 AR-1254-3	7.784	0.000	38353	0	18.482	N.D. #
29)	L6 AR-1254-4	0.000	6.941	0	47732	N.D.	19.800 #
32)	L7 AR-1260-2	8.194	7.007	32091	46154	19.009	17.975
35)	L7 AR-1260-5	9.113	0.000	11477	0	3.638	N.D. #
37)	L8 AR-1262-2	9.113	0.000	11477	0	4.000	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042220\
Data File : P0067962.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Apr 2020 8:28
Operator : DD\AJ
Sample : L2343-04RE
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SB-4RE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 10:38:34 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.932 min
Delta R.T.: -0.003 min
Response: 409708
Conc: 17.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#1 Tetrachloro-m-xylene

R.T.: 3.941 min
Delta R.T.: -0.003 min
Response: 657775
Conc: 19.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.905 min
Delta R.T.: -0.014 min
Response: 384344
Conc: 11.29 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.224 min
Delta R.T.: -0.009 min
Response: 440954
Conc: 10.67 ng/ml



#3 AR-1016-1


R.T.: 6.177 min
Delta R.T.: -0.063 min
Response: 21406
Conc: 20.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 5.200 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.373 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.445 min
Delta R.T.: 0.055 min
Response: 347006
Conc: 353.46 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.372 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#11 AR-1232-1

R.T.: 4.445 min
Delta R.T.: 0.055 min
Response: 347006
Conc: 501.41 ng/ml




#20 AR-1242-5

R.T.: 7.193 min
Delta R.T.: -0.042 min
Response: 33940
Conc: 63.53 ng/ml



#20 AR-1242-5


R.T.: 6.090 min
Delta R.T.: 0.019 min
Response: 41787
Conc: 44.66 ng/ml



#24 AR-1248-4


R.T.: 7.193 min
Delta R.T.: -0.001 min
Response: 33940
Conc: 35.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.690 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 7.193 min
Delta R.T.: -0.041 min
Response: 33940
Conc: 37.66 ng/ml



#25 AR-1248-5


R.T.: 6.090 min
Delta R.T.: -0.023 min
Response: 41787
Conc: 33.77 ng/ml



#26 AR-1254-1


R.T.: 7.193 min
Delta R.T.: 0.029 min
Response: 33940
Conc: 27.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#26 AR-1254-1

R.T.: 6.090 min
Delta R.T.: 0.020 min
Response: 41787
Conc: 16.30 ng/ml




#27 AR-1254-2

R.T.: 7.444 min
Delta R.T.: 0.051 min
Response: 90401
Conc: 45.97 ng/ml



#27 AR-1254-2


R.T.: 6.298 min
Delta R.T.: 0.068 min
Response: 267748
Conc: 117.40 ng/ml



#28 AR-1254-3


R.T.: 7.784 min
Delta R.T.: 0.011 min
Response: 38353
Conc: 18.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.649 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 8.068 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.941 min
Delta R.T.: 0.052 min
Response: 47732
Conc: 19.80 ng/ml



#32 AR-1260-2


R.T.: 8.194 min
Delta R.T.: -0.011 min
Response: 32091
Conc: 19.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE




#32 AR-1260-2

R.T.: 7.007 min
Delta R.T.: 0.022 min
Response: 46154
Conc: 17.97 ng/ml




#35 AR-1260-5

R.T.: 9.113 min
Delta R.T.: -0.009 min
Response: 11477
Conc: 3.64 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.869 min
Response: 0
Conc: N.D.



#37 AR-1262-2

R.T.: 9.113 min
Delta R.T.: -0.008 min
Response: 11477
Conc: 4.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
SB-4RE



#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.869 min
Response: 0
Conc: N.D.