

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042220\
 Data File : P0067968.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Apr 2020 10:10
 Operator : DD\AJ
 Sample : L2327-32
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-3-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 10:41:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.919	3.937	469708	749830	19.724	22.372
2) SA Decachlor...	10.891	9.220	750731	855036	22.052	20.680

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042220\
Data File : P0067968.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Apr 2020 10:10
Operator : DD\AJ
Sample : L2327-32
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-3-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 10:41:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.919 min
Delta R.T.: -0.016 min
Response: 469708
Conc: 19.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-3-S




#1 Tetrachloro-m-xylene

R.T.: 3.937 min
Delta R.T.: -0.007 min
Response: 749830
Conc: 22.37 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.891 min
Delta R.T.: -0.027 min
Response: 750731
Conc: 22.05 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.220 min
Delta R.T.: -0.013 min
Response: 855036
Conc: 20.68 ng/ml