

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0042418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 25 01:34:16 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044038.D 1000 =P0044036.D 750 =P0044037.D
 250 =P0044039.D 50 =P0044040.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		4.719	4.453	4.568	4.666	3.695	4.420	E5	9.46
2) SA	Decachlorobiphenyl		4.846	4.445	4.721	4.978	4.105	4.619	E5	7.54
3) L1	AR-1016-1		0.993	0.909	0.945	1.011	0.805	0.933	E4	8.80
4) L1	AR-1016-2		1.292	1.165	1.225	1.361	1.052	1.219	E4	9.74
5) L1	AR-1016-3		1.048	0.961	1.003	1.081	0.809	0.980	E4	10.84
6) L1	AR-1016-4		1.055	0.969	1.013	1.056	0.831	0.985	E4	9.45
7) L1	AR-1016-5		8.733	7.961	8.358	9.029	7.329	8.282	E3	8.05
8) L2	AR-1221-1							0.000		-1.00
9) L2	AR-1221-2							0.000		-1.00
10) L2	AR-1221-3							0.000		-1.00
11) L3	AR-1232-1							0.000		-1.00
12) L3	AR-1232-2							0.000		-1.00
13) L3	AR-1232-3							0.000		-1.00
14) L3	AR-1232-4							0.000		-1.00
15) L3	AR-1232-5							0.000		-1.00
16) L4	AR-1242-1							0.000		-1.00
17) L4	AR-1242-2							0.000		-1.00
18) L4	AR-1242-3							0.000		-1.00
19) L4	AR-1242-4							0.000		-1.00
20) L4	AR-1242-5							0.000		-1.00
21) L5	AR-1248-1							0.000		-1.00
22) L5	AR-1248-2							0.000		-1.00
23) L5	AR-1248-3							0.000		-1.00
24) L5	AR-1248-4							0.000		-1.00
25) L5	AR-1248-5							0.000		-1.00
26) L6	AR-1254-1							0.000		-1.00
27) L6	AR-1254-2							0.000		-1.00
28) L6	AR-1254-3							0.000		-1.00
29) L6	AR-1254-4							0.000		-1.00
30) L6	AR-1254-5							0.000		-1.00
31) L7	AR-1260-1		1.924	1.760	1.853	1.986	1.670	1.839	E4	6.86
32) L7	AR-1260-2		2.594	2.394	2.542	2.686	2.260	2.495	E4	6.76
33) L7	AR-1260-3		2.179	1.990	2.121	2.251	1.853	2.079	E4	7.63
34) L7	AR-1260-4		5.186	4.878	5.159	5.258	4.117	4.919	E4	9.58
35) L7	AR-1260-5		3.161	2.982	3.156	3.288	2.664	3.050	E4	7.93
36) L8	AR-1262-1							0.000		-1.00
37) L8	AR-1262-2							0.000		-1.00
38) L8	AR-1262-3							0.000		-1.00
39) L8	AR-1262-4							0.000		-1.00
40) L8	AR-1262-5							0.000		-1.00
41) L9	AR-1268-1							0.000		-1.00
42) L9	AR-1268-2							0.000		-1.00
43) L9	AR-1268-3							0.000		-1.00
44) L9	AR-1268-4							0.000		-1.00
45) L9	AR-1268-5							0.000		-1.00

Signal #2 Calibration Files

500 =P0044038.D 1000 =P0044036.D 750 =P0044037.D
 250 =P0044039.D 50 =P0044040.D

Compound			500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene		6.187	5.693	5.917	6.195	4.877	5.774	E5	9.40
2) SA	Decachlorobiphenyl		6.474	5.993	6.333	6.477	4.973	6.050	E5	10.47

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0042418.M
 Title : GC EXTRACTABLES
 Last Update : Wed Apr 25 01:34:16 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044038.D 1000 =P0044036.D 750 =P0044037.D
 250 =P0044039.D 50 =P0044040.D

Compound			500	1000	750	250	50	Avg	%RSD

3)	L1	AR-1016-1	1.385	1.236	1.301	1.432	1.111	1.293 E4	9.79
4)	L1	AR-1016-2	1.086	0.960	1.011	1.125	0.950	1.026 E4	7.50
5)	L1	AR-1016-3	1.310	1.159	1.225	1.353	1.117	1.233 E4	8.04
6)	L1	AR-1016-4	1.401	1.248	1.310	1.429	1.193	1.316 E4	7.59
7)	L1	AR-1016-5	1.440	1.272	1.363	1.500	1.158	1.347 E4	10.08
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	2.828	2.531	2.726	2.956	2.522	2.713 E4	6.94
32)	L7	AR-1260-2	3.766	3.368	3.649	3.936	3.399	3.624 E4	6.68
33)	L7	AR-1260-3	3.878	3.527	3.798	4.008	3.297	3.702 E4	7.74
34)	L7	AR-1260-4	7.935	7.317	7.807	8.040	6.340	7.488 E4	9.33
35)	L7	AR-1260-5	5.593	5.149	5.474	5.673	4.481	5.274 E4	9.22
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range