

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042420\
 Data File : P0068039.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 24 Apr 2020 13:16
 Operator : DD\AJ
 Sample : PB128469BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 13:38:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 03 16:04:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.918	3.938	575707	783953	24.175	23.390
2)	SA Decachlor...	10.889	9.218	632367	639164	18.575	15.459
Target Compounds							
28)	L6 AR-1254-3	7.774	0.000	26258	0	12.654	N.D. #
29)	L6 AR-1254-4	0.000	6.936	0	35374	N.D.	14.674 #
32)	L7 AR-1260-2	0.000	6.936	0	35374	N.D.	13.776 #
35)	L7 AR-1260-5	9.142	0.000	10610	0	3.363	N.D. #
37)	L8 AR-1262-2	9.142	0.000	10610	0	3.698	N.D. #
43)	L9 AR-1268-3	9.725	0.000	10308	0	3.550	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042420\
Data File : P0068039.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 24 Apr 2020 13:16
Operator : DD\AJ
Sample : PB128469BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 13:38:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040120.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 03 16:04:09 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.918 min
Delta R.T.: -0.017 min
Response: 575707
Conc: 24.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.938 min
Delta R.T.: -0.006 min
Response: 783953
Conc: 23.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.889 min
Delta R.T.: -0.030 min
Response: 632367
Conc: 18.58 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.218 min
Delta R.T.: -0.015 min
Response: 639164
Conc: 15.46 ng/ml



#28 AR-1254-3


R.T.: 7.774 min
Delta R.T.: 0.001 min
Response: 26258
Conc: 12.65 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.649 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 8.068 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 6.936 min
Delta R.T.: 0.047 min
Response: 35374
Conc: 14.67 ng/ml



#32 AR-1260-2


R.T.: 0.000 min
Exp R.T.: 8.205 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.936 min
Delta R.T.: -0.050 min
Response: 35374
Conc: 13.78 ng/ml




#35 AR-1260-5

R.T.: 9.142 min
Delta R.T.: 0.021 min
Response: 10610
Conc: 3.36 ng/ml



#35 AR-1260-5


R.T.: 0.000 min
Exp R.T.: 7.869 min
Response: 0
Conc: N.D.



#37 AR-1262-2


R.T.: 9.142 min
Delta R.T.: 0.021 min
Response: 10610
Conc: 3.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 0.000 min
Exp R.T. : 7.869 min
Response: 0
Conc: N.D.



#43 AR-1268-3

R.T.: 9.725 min
Delta R.T.: -0.021 min
Response: 10308
Conc: 3.55 ng/ml



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.430 min
Response: 0
Conc: N.D.