

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042720\
 Data File : P0068066.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 27 Apr 2020 14:30
 Operator : DD\AJ
 Sample : L2408-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HP-627-EB

Manual Integrations
 APPROVED

Sohil
 4/28/2020 3:43:25 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 18:10:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Apr 27 18:08:26 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.921	3.938	595391	832324	24.881	26.047m
2) SA Decachlor...	10.889	9.216	552875	576683	14.650	14.942

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042720\
Data File : P0068066.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Apr 2020 14:30
Operator : DD\AJ
Sample : L2408-01
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
HP-627-EB


Manual Integrations
APPROVED

Sohil
4/28/2020 3:43:25 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 18:10:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Apr 27 18:08:26 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.921 min
Delta R.T.: 0.004 min
Response: 595391
Conc: 24.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
HP-627-EB


Manual Integrations
APPROVED

Sohil
4/28/2020 3:43:25 PM




#1 Tetrachloro-m-xylene

R.T.: 3.938 min
Delta R.T.: 0.001 min
Response: 832324
Conc: 26.05 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.889 min
Delta R.T.: 0.004 min
Response: 552875
Conc: 14.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.216 min
Delta R.T.: 0.001 min
Response: 576683
Conc: 14.94 ng/ml