

Data Path : P:\HPCHEM1\ECD_0\Data\P0042818\
Data File : P0044184.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Apr 2018 23:53
Operator : SM/UA
Sample : J2574-11
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BLDG-FND-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 04:11:07 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.040	3.294	14132894	17824711	19.743	18.426
2) SA Decachlor...	9.389	8.060	10582592	12733511	17.181	15.344

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0042818\
Data File : P0044184.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 27 Apr 2018 23:53
Operator : SM/UA
Sample : J2574-11
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BLDG-FND-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 28 04:11:07 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.040 min
Delta R.T.: -0.004 min
Response: 14132894
Conc: 19.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
BLDG-FND-3




#1 Tetrachloro-m-xylene

R.T.: 3.294 min
Delta R.T.: -0.004 min
Response: 17824711
Conc: 18.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.389 min
Delta R.T.: -0.007 min
Response: 10582592
Conc: 17.18 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.060 min
Delta R.T.: -0.007 min
Response: 12733511
Conc: 15.34 ng/ml