

Data Path : P:\HPCHEM1\ECD_0\Data\P0043018\
Data File : P0044242.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Apr 2018 20:25
Operator : SM/UA
Sample : J2553-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BPS1-SB4017-1520

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 03:24:24 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.041	3.297	13261031	17152318	18.525	17.731
2) SA Decachlor...	9.387	8.059	9432299	10375591	15.313	12.503

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : P:\HPCHEM1\ECD_0\Data\P0043018\
Data File : P0044242.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Apr 2018 20:25
Operator : SM/UA
Sample : J2553-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
BPS1-SB4017-1520

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 03:24:24 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.041 min
Delta R.T.: -0.003 min
Response: 13261031
Conc: 18.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
BPS1-SB4017-1520




#1 Tetrachloro-m-xylene

R.T.: 3.297 min
Delta R.T.: -0.001 min
Response: 17152318
Conc: 17.73 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.387 min
Delta R.T.: -0.009 min
Response: 9432299
Conc: 15.31 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.059 min
Delta R.T.: -0.008 min
Response: 10375591
Conc: 12.50 ng/ml