

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050119\
 Data File : P0055761.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 01 May 2019 23:04
 Operator : SM/SJ
 Sample : K2626-04
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RI2-4(0-6)

Manual Integrations
 APPROVED

Ankita
 5/2/2019 4:08:09 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 04:24:03 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 30 05:21:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.282	3.597	804218	857362	28.593	29.390
2)	SA Decachlor...	9.882	8.548	569519	380959	13.335	11.992
Target Compounds							
26)	L6 AR-1254-1	6.291	5.461	129768	151757	68.701	58.117
27)	L6 AR-1254-2	6.509	5.607	160727	85364	54.275	36.170 #
28)	L6 AR-1254-3	6.875	6.008	360806	326202	108.448	84.518
29)	L6 AR-1254-4	7.157	6.235	76650	66966	29.858	25.461
30)	L6 AR-1254-5	7.574	6.650	217006	219965	76.327	62.872
31)	L7 AR-1260-1	7.035	6.137	127319	145318	55.416	60.486
32)	L7 AR-1260-2	7.290	6.321	158021	201622	55.136	60.766
33)	L7 AR-1260-3	7.648	6.476	137940	133256	61.457	47.775
34)	L7 AR-1260-4	7.866	6.944	177656	54490	69.077m	23.268 #
35)	L7 AR-1260-5	8.181	7.183	116125	107205	23.954	19.129m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050119\
Data File : PO055761.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 01 May 2019 23:04
Operator : SM/SJ
Sample : K2626-04
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 04:24:03 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042919.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 30 05:21:50 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.282 min
Delta R.T.: -0.002 min
Response: 804218
Conc: 28.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#1 Tetrachloro-m-xylene

R.T.: 3.597 min
Delta R.T.: 0.000 min
Response: 857362
Conc: 29.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.882 min
Delta R.T.: -0.004 min
Response: 569519
Conc: 13.33 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.548 min
Delta R.T.: -0.003 min
Response: 380959
Conc: 11.99 ng/ml




#26 AR-1254-1

R.T.: 6.291 min
Delta R.T.: -0.002 min
Response: 129768
Conc: 68.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#26 AR-1254-1

R.T.: 5.461 min
Delta R.T.: -0.003 min
Response: 151757
Conc: 58.12 ng/ml




#27 AR-1254-2

R.T.: 6.509 min
Delta R.T.: -0.001 min
Response: 160727
Conc: 54.27 ng/ml



#27 AR-1254-2

R.T.: 5.607 min
Delta R.T.: -0.003 min
Response: 85364
Conc: 36.17 ng/ml




#28 AR-1254-3

R.T.: 6.875 min
Delta R.T.: 0.000 min
Response: 360806
Conc: 108.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#28 AR-1254-3

R.T.: 6.008 min
Delta R.T.: -0.003 min
Response: 326202
Conc: 84.52 ng/ml




#29 AR-1254-4

R.T.: 7.157 min
Delta R.T.: -0.002 min
Response: 76650
Conc: 29.86 ng/ml



#29 AR-1254-4

R.T.: 6.235 min
Delta R.T.: -0.003 min
Response: 66966
Conc: 25.46 ng/ml




#30 AR-1254-5

R.T.: 7.574 min
Delta R.T.: -0.003 min
Response: 217006
Conc: 76.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#30 AR-1254-5

R.T.: 6.650 min
Delta R.T.: -0.003 min
Response: 219965
Conc: 62.87 ng/ml




#31 AR-1260-1

R.T.: 7.035 min
Delta R.T.: -0.002 min
Response: 127319
Conc: 55.42 ng/ml



#31 AR-1260-1

R.T.: 6.137 min
Delta R.T.: -0.003 min
Response: 145318
Conc: 60.49 ng/ml




#32 AR-1260-2

R.T.: 7.290 min
Delta R.T.: -0.004 min
Response: 158021
Conc: 55.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#32 AR-1260-2

R.T.: 6.321 min
Delta R.T.: -0.005 min
Response: 201622
Conc: 60.77 ng/ml




#33 AR-1260-3

R.T.: 7.648 min
Delta R.T.: -0.003 min
Response: 137940
Conc: 61.46 ng/ml



#33 AR-1260-3

R.T.: 6.476 min
Delta R.T.: -0.003 min
Response: 133256
Conc: 47.78 ng/ml




#34 AR-1260-4

R.T.: 7.866 min
Delta R.T.: -0.009 min
Response: 177656
Conc: 69.08 ng/ml m

Instrument :
ECD_O
ClientSampleId :
RI2-4(0-6)


Manual Integrations
APPROVED

Ankita
5/2/2019 4:08:09 PM




#34 AR-1260-4

R.T.: 6.944 min
Delta R.T.: -0.002 min
Response: 54490
Conc: 23.27 ng/ml



#35 AR-1260-5

R.T.: 8.181 min
Delta R.T.: -0.004 min
Response: 116125
Conc: 23.95 ng/ml



#35 AR-1260-5

R.T.: 7.183 min
Delta R.T.: -0.004 min
Response: 107205
Conc: 19.13 ng/ml m