

Data Path : P:\HPCHEM1\ECD_0\Data\P0050318\
Data File : P0044340.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 May 2018 20:27
Operator : SM/UA
Sample : J2703-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-1

Manual Integrations
APPROVED

UGO
5/4/2018 4:02:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 02:25:24 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.035	3.294	10383745	12026542	14.505	12.432
2) SA Decachlor...	9.371	8.050	8670966	9541706	14.077m	11.498m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD_0\Data\P0050318\
Data File : P0044340.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 May 2018 20:27
Operator : SM/UA
Sample : J2703-01
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

UGO
5/4/2018 4:02:41 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 02:25:24 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0042518.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 26 04:36:05 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.035 min
Delta R.T.: -0.009 min
Response: 10383745
Conc: 14.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-1


Manual Integrations
APPROVED

UGO
5/4/2018 4:02:41 PM




#1 Tetrachloro-m-xylene

R.T.: 3.294 min
Delta R.T.: -0.004 min
Response: 12026542
Conc: 12.43 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.371 min
Delta R.T.: -0.024 min
Response: 8670966
Conc: 14.08 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.050 min
Delta R.T.: -0.017 min
Response: 9541706
Conc: 11.50 ng/ml m