

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050420\
 Data File : P0068145.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 May 2020 12:38
 Operator : DD\AJ
 Sample : L2457-01 5X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 1360-1361

Manual Integrations
 APPROVED

Sohil
 5/6/2020 12:18:12 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 13:25:57 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 28 07:12:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.914	3.939	117071	136332	4.892	4.266
2) SA Decachlor...	10.880	9.206	356710	316975	9.452	8.213m
Target Compounds						
31) L7 AR-1260-1	7.916	6.774	61678	95635	48.869	47.573
32) L7 AR-1260-2	8.181	6.972	88727	117404	56.677	47.463
33) L7 AR-1260-3	8.543	7.125	56350	107859	44.774	46.891
34) L7 AR-1260-4	8.772	7.606	68292	76523	49.106	38.197
35) L7 AR-1260-5	9.095	7.853	117261	180164	40.214	38.567

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050420\
Data File : PO068145.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 May 2020 12:38
Operator : DD\AJ
Sample : L2457-01 5X
Misc :
ALS Vial : 17 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
1360-1361


Manual Integrations
APPROVED

Sohil
5/6/2020 12:18:12 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 13:25:57 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042720.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 28 07:12:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.914 min
Delta R.T.: -0.002 min
Response: 117071
Conc: 4.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
1360-1361


Manual Integrations
APPROVED

Sohil
5/6/2020 12:18:12 PM




#1 Tetrachloro-m-xylene

R.T.: 3.939 min
Delta R.T.: 0.003 min
Response: 136332
Conc: 4.27 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.880 min
Delta R.T.: -0.005 min
Response: 356710
Conc: 9.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.206 min
Delta R.T.: -0.009 min
Response: 316975
Conc: 8.21 ng/ml m




#31 AR-1260-1

R.T.: 7.916 min
Delta R.T.: -0.004 min
Response: 61678
Conc: 48.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
1360-1361


Manual Integrations
APPROVED

Sohil
5/6/2020 12:18:12 PM




#31 AR-1260-1

R.T.: 6.774 min
Delta R.T.: -0.001 min
Response: 95635
Conc: 47.57 ng/ml




#32 AR-1260-2

R.T.: 8.181 min
Delta R.T.: -0.004 min
Response: 88727
Conc: 56.68 ng/ml



#32 AR-1260-2

R.T.: 6.972 min
Delta R.T.: -0.001 min
Response: 117404
Conc: 47.46 ng/ml




#33 AR-1260-3

R.T.: 8.543 min
Delta R.T.: -0.005 min
Response: 56350
Conc: 44.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
1360-1361


Manual Integrations
APPROVED

Sohil
5/6/2020 12:18:12 PM




#33 AR-1260-3

R.T.: 7.125 min
Delta R.T.: -0.002 min
Response: 107859
Conc: 46.89 ng/ml




#34 AR-1260-4

R.T.: 8.772 min
Delta R.T.: -0.005 min
Response: 68292
Conc: 49.11 ng/ml



#34 AR-1260-4

R.T.: 7.606 min
Delta R.T.: -0.002 min
Response: 76523
Conc: 38.20 ng/ml




#35 AR-1260-5

R.T.: 9.095 min
Delta R.T.: -0.005 min
Response: 117261
Conc: 40.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
1360-1361

Manual Integrations
APPROVED

Sohil
5/6/2020 12:18:12 PM



#35 AR-1260-5

R.T.: 7.853 min
Delta R.T.: -0.003 min
Response: 180164
Conc: 38.57 ng/ml