

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050620\
 Data File : P0068200.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 06 May 2020 15:08
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 15:26:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 28 07:12:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.910	3.934	536718	687382	22.429	21.511
2)	SA Decachlor...	10.875	9.210	828706	796019	21.959	20.625
Target Compounds							
9)	L2 AR-1221-2	5.263	4.283	6908	9885	26.356	26.665
20)	L4 AR-1242-5	7.219	6.080	39491	31797	59.351	30.828 #
24)	L5 AR-1248-4	7.145	5.732	79022	-9340	70.713	N.D. #
25)	L5 AR-1248-5	7.219	6.080	39491	31797	37.013	23.484 #
26)	L6 AR-1254-1	7.145	6.080	79022	31797	71.439	15.499 #
27)	L6 AR-1254-2	7.368	0.000	24358	0	14.060	N.D. #
28)	L6 AR-1254-3	7.767	6.609	51666	41049	28.031	13.776 #
29)	L6 AR-1254-4	8.053	6.931	43197	13404	29.938	6.700 #
30)	L6 AR-1254-5	8.471	0.000	42756	0	28.555	N.D. #
31)	L7 AR-1260-1	7.941	0.000	38735	0	30.690	N.D. #
32)	L7 AR-1260-2	8.250	6.931	88268	13404	56.384	5.419 #
33)	L7 AR-1260-3	0.000	7.193	0	10890	N.D.	4.734 #
35)	L7 AR-1260-5	9.093	0.000	11543	0	3.959	N.D. #
37)	L8 AR-1262-2	9.093	0.000	11543	0	3.490	N.D. #
40)	L8 AR-1262-5	0.000	8.724	0	1091	N.D.	0.616 #
43)	L9 AR-1268-3	9.717	8.410	118048	8444	34.292	1.946 #
44)	L9 AR-1268-4	0.000	8.724	0	1091	N.D.	0.531 #
45)	L9 AR-1268-5	10.559	8.972	51956	8169	4.395	0.611 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050620\
Data File : PO068200.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 06 May 2020 15:08
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 15:26:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042720.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 28 07:12:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.910 min
Delta R.T.: -0.006 min
Response: 536718
Conc: 22.43 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.934 min
Delta R.T.: -0.003 min
Response: 687382
Conc: 21.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.875 min
Delta R.T.: -0.010 min
Response: 828706
Conc: 21.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.210 min
Delta R.T.: -0.005 min
Response: 796019
Conc: 20.62 ng/ml



#9 AR-1221-2


R.T.: 5.263 min
Delta R.T.: -0.004 min
Response: 6908
Conc: 26.36 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 4.283 min
Delta R.T.: -0.009 min
Response: 9885
Conc: 26.66 ng/ml




#20 AR-1242-5

R.T.: 7.219 min
Delta R.T.: 0.004 min
Response: 39491
Conc: 59.35 ng/ml



#20 AR-1242-5


R.T.: 6.080 min
Delta R.T.: 0.020 min
Response: 31797
Conc: 30.83 ng/ml



#24 AR-1248-4


R.T.: 7.145 min
Delta R.T.: -0.030 min
Response: 79022
Conc: 70.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 5.732 min
Delta R.T.: 0.052 min
Response: -9340
Conc: N.D.




#25 AR-1248-5

R.T.: 7.219 min
Delta R.T.: 0.005 min
Response: 39491
Conc: 37.01 ng/ml



#25 AR-1248-5


R.T.: 6.080 min
Delta R.T.: -0.022 min
Response: 31797
Conc: 23.48 ng/ml



#26 AR-1254-1


R.T.: 7.145 min
Delta R.T.: 0.001 min
Response: 79022
Conc: 71.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 6.080 min
Delta R.T.: 0.021 min
Response: 31797
Conc: 15.50 ng/ml




#27 AR-1254-2

R.T.: 7.368 min
Delta R.T.: -0.005 min
Response: 24358
Conc: 14.06 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 6.220 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.767 min
Delta R.T.: 0.014 min
Response: 51666
Conc: 28.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 6.609 min
Delta R.T.: -0.028 min
Response: 41049
Conc: 13.78 ng/ml




#29 AR-1254-4

R.T.: 8.053 min
Delta R.T.: 0.005 min
Response: 43197
Conc: 29.94 ng/ml



#29 AR-1254-4


R.T.: 6.931 min
Delta R.T.: 0.054 min
Response: 13404
Conc: 6.70 ng/ml



#30 AR-1254-5


R.T.: 8.471 min
Delta R.T.: -0.003 min
Response: 42756
Conc: 28.55 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.305 min
Response: 0
Conc: N.D.




#31 AR-1260-1

R.T.: 7.941 min
Delta R.T.: 0.022 min
Response: 38735
Conc: 30.69 ng/ml



#31 AR-1260-1


R.T.: 0.000 min
Exp R.T. : 6.776 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 8.250 min
Delta R.T.: 0.065 min
Response: 88268
Conc: 56.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#32 AR-1260-2

R.T.: 6.931 min
Delta R.T.: -0.042 min
Response: 13404
Conc: 5.42 ng/ml




#33 AR-1260-3

R.T.: 0.000 min
Exp R.T. : 8.548 min
Response: 0
Conc: N.D.



#33 AR-1260-3


R.T.: 7.193 min
Delta R.T.: 0.066 min
Response: 10890
Conc: 4.73 ng/ml



#35 AR-1260-5


R.T.: 9.093 min
Delta R.T.: -0.007 min
Response: 11543
Conc: 3.96 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 0.000 min
Exp R.T. : 7.856 min
Response: 0
Conc: N.D.




#37 AR-1262-2

R.T.: 9.093 min
Delta R.T.: -0.005 min
Response: 11543
Conc: 3.49 ng/ml



#37 AR-1262-2


R.T.: 0.000 min
Exp R.T. : 7.855 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 0.000 min
Exp R.T.: 10.151 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#40 AR-1262-5

R.T.: 8.724 min
Delta R.T.: 0.024 min
Response: 1091
Conc: 0.62 ng/ml




#43 AR-1268-3

R.T.: 9.717 min
Delta R.T.: -0.003 min
Response: 118048
Conc: 34.29 ng/ml



#43 AR-1268-3


R.T.: 8.410 min
Delta R.T.: -0.003 min
Response: 8444
Conc: 1.95 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 10.152 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#44 AR-1268-4

R.T.: 8.724 min
Delta R.T.: 0.025 min
Response: 1091
Conc: 0.53 ng/ml



#45 AR-1268-5

R.T.: 10.559 min
Delta R.T.: 0.002 min
Response: 51956
Conc: 4.39 ng/ml



#45 AR-1268-5

R.T.: 8.972 min
Delta R.T.: -0.003 min
Response: 8169
Conc: 0.61 ng/ml