

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050720\
 Data File : P0068208.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 May 2020 9:44
 Operator : DD\AJ
 Sample : L2492-13
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 OILY-DEBRIS-COMP

Manual Integrations
 APPROVED

Sohil
 5/8/2020 7:55:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 11:10:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 28 07:12:58 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.910	3.934	633034	797754	26.454	24.965m
2) SA Decachlor...	10.873	9.210	814996	779302	21.596m	20.192m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050720\
Data File : P0068208.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 May 2020 9:44
Operator : DD\AJ
Sample : L2492-13
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
OILY-DEBRIS-COMP


Manual Integrations
APPROVED

Sohil
5/8/2020 7:55:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 11:10:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042720.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 28 07:12:58 2020
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.910 min
Delta R.T.: -0.006 min
Response: 633034
Conc: 26.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
OILY-DEBRIS-COMP


Manual Integrations
APPROVED

Sohil
5/8/2020 7:55:14 AM




#1 Tetrachloro-m-xylene

R.T.: 3.934 min
Delta R.T.: -0.003 min
Response: 797754
Conc: 24.96 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.873 min
Delta R.T.: -0.012 min
Response: 814996
Conc: 21.60 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.210 min
Delta R.T.: -0.006 min
Response: 779302
Conc: 20.19 ng/ml m