

Data Path : P:\HPCHEM1\ECD 0\Data\P0050818\  
 Data File : P0044459.D  
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
 Acq On : 08 May 2018 11:09  
 Operator : SM/UA  
 Sample : J2722-02  
 Misc :  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 ECD\_O  
 ClientSampleId :  
 OILY-STONES

Manual Integrations  
 APPROVED

UGO  
 5/9/2018 12:41:20 PM

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: May 08 15:58:43 2018  
 Quant Method : P:\HPCHEM1\ECD\_0\methods\P0042518.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Thu Apr 26 04:36:05 2018  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl  
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1   | Resp#2   | nq/ml   | nq/ml   |
|-----------------------------|-------|-------|----------|----------|---------|---------|
| -----                       |       |       |          |          |         |         |
| System Monitoring Compounds |       |       |          |          |         |         |
| 1) SA Tetrachlo...          | 4.034 | 3.297 | 7103646  | 7706609  | 9.923   | 7.967   |
| 2) SA Decachlor...          | 9.361 | 8.044 | 10194703 | 11737022 | 16.551m | 14.143m |

Target Compounds

-----

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\PO050818\  
Data File : PO044459.D  
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH  
Acq On : 08 May 2018 11:09  
Operator : SM/UA  
Sample : J2722-02  
Misc :  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
ECD\_O  
ClientSampleId :  
OILY-STONES

Manual Integrations  
APPROVED

UGO  
5/9/2018 12:41:20 PM

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: May 08 15:58:43 2018  
Quant Method : P:\HPCHEM1\ECD\_0\methods\PO042518.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Thu Apr 26 04:36:05 2018  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl  
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2  
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

