

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0050918.M
 Title : GC EXTRACTABLES
 Last Update : Wed May 09 17:00:02 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044503.D 1000 =P0044501.D 750 =P0044502.D
 250 =P0044504.D 50 =P0044505.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		1.984	2.011	2.022	1.944	2.090	2.010	E5	2.67
2) SA	Decachlorobiphenyl		1.466	1.379	1.391	1.439	1.607	1.456	E5	6.25
3) L1	AR-1016-1		4.834	4.698	4.782	4.742	5.095	4.830	E3	3.24
4) L1	AR-1016-2		6.281	5.639	6.083	6.337	5.996	6.067	E3	4.57
5) L1	AR-1016-3		5.235	4.518	4.886	5.059	5.660	5.072	E3	8.34
6) L1	AR-1016-4		4.627	4.474	4.452	4.694	4.801	4.610	E3	3.21
7) L1	AR-1016-5		3.946	3.715	3.755	3.988	4.179	3.917	E3	4.80
8) L2	AR-1221-1		3.489					3.489	E3	0.00
9) L2	AR-1221-2		2.545					2.545	E3	0.00
10) L2	AR-1221-3		8.255					8.255	E3	0.00
11) L3	AR-1232-1		3.784					3.784	E3	0.00
12) L3	AR-1232-2		2.943					2.943	E3	0.00
13) L3	AR-1232-3		4.967					4.967	E3	0.00
14) L3	AR-1232-4		2.941					2.941	E3	0.00
15) L3	AR-1232-5		2.318					2.318	E3	0.00
16) L4	AR-1242-1		3.832					3.832	E3	0.00
17) L4	AR-1242-2		8.570					8.570	E3	0.00
18) L4	AR-1242-3		4.123					4.123	E3	0.00
19) L4	AR-1242-4		3.999					3.999	E3	0.00
20) L4	AR-1242-5		4.484					4.484	E3	0.00
21) L5	AR-1248-1		6.525					6.525	E3	0.00
22) L5	AR-1248-2		5.679					5.679	E3	0.00
23) L5	AR-1248-3		7.795					7.795	E3	0.00
24) L5	AR-1248-4		7.585					7.585	E3	0.00
25) L5	AR-1248-5		4.827					4.827	E3	0.00
26) L6	AR-1254-1		1.202					1.202	E4	0.00
27) L6	AR-1254-2		7.076					7.076	E3	0.00
28) L6	AR-1254-3		1.209					1.209	E4	0.00
29) L6	AR-1254-4		9.660					9.660	E3	0.00
30) L6	AR-1254-5		6.644					6.644	E3	0.00
31) L7	AR-1260-1		8.002	7.539	7.471	7.972	9.146	8.026	E3	8.37
32) L7	AR-1260-2		1.039	0.974	0.980	1.029	1.130	1.030	E4	6.09
33) L7	AR-1260-3		8.081	7.779	7.482	7.952	8.773	8.013	E3	5.99
34) L7	AR-1260-4		1.724	1.681	1.682	1.675	1.846	1.722	E4	4.19
35) L7	AR-1260-5		1.034	1.003	1.006	1.016	0.965	1.005	E4	2.53
36) L8	AR-1262-1		7.769					7.769	E3	0.00
37) L8	AR-1262-2		1.002					1.002	E4	0.00
38) L8	AR-1262-3		8.907					8.907	E3	0.00
39) L8	AR-1262-4		1.184					1.184	E4	0.00
40) L8	AR-1262-5		2.467					2.467	E4	0.00
41) L9	AR-1268-1		2.616					2.616	E4	0.00
42) L9	AR-1268-2		2.303					2.303	E4	0.00
43) L9	AR-1268-3		2.010					2.010	E4	0.00
44) L9	AR-1268-4		8.347					8.347	E3	0.00
45) L9	AR-1268-5		5.999					5.999	E4	0.00

Signal #2 Calibration Files

500 =P0044503.D 1000 =P0044501.D 750 =P0044502.D
 250 =P0044504.D 50 =P0044505.D

Compound			500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene		2.683	2.620	2.666	2.682	2.781	2.686	E5	2.19
2) SA	Decachlorobiphenyl		1.738	1.625	1.573	1.808	2.033	1.755	E5	10.28

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0050918.M
 Title : GC EXTRACTABLES
 Last Update : Wed May 09 17:00:02 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044503.D 1000 =P0044501.D 750 =P0044502.D
 250 =P0044504.D 50 =P0044505.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	6.443	6.307	6.262	6.882	6.777	6.534	E3	4.29
4)	L1	AR-1016-2	5.743	5.301	5.438	5.873	4.562	5.383	E3	9.53
5)	L1	AR-1016-3	6.912	6.483	6.614	7.002	4.412	6.285	E3	16.99
6)	L1	AR-1016-4	7.477	7.037	7.184	7.563	6.808	7.214	E3	4.32
7)	L1	AR-1016-5	7.789	7.352	7.480	7.639	6.838	7.420	E3	4.91
8)	L2	AR-1221-1	4.614					4.614	E3	0.00
9)	L2	AR-1221-2	3.360					3.360	E3	0.00
10)	L2	AR-1221-3	1.110					1.110	E4	0.00
11)	L3	AR-1232-1	5.039					5.039	E3	0.00
12)	L3	AR-1232-2	1.144					1.144	E4	0.00
13)	L3	AR-1232-3	4.614					4.614	E3	0.00
14)	L3	AR-1232-4	3.443					3.443	E3	0.00
15)	L3	AR-1232-5	2.996					2.996	E3	0.00
16)	L4	AR-1242-1	5.348					5.348	E3	0.00
17)	L4	AR-1242-2	1.991					1.991	E4	0.00
18)	L4	AR-1242-3	5.972					5.972	E3	0.00
19)	L4	AR-1242-4	5.717					5.717	E3	0.00
20)	L4	AR-1242-5	6.201					6.201	E3	0.00
21)	L5	AR-1248-1	1.005					1.005	E4	0.00
22)	L5	AR-1248-2	8.985					8.985	E3	0.00
23)	L5	AR-1248-3	1.507					1.507	E4	0.00
24)	L5	AR-1248-4	1.214					1.214	E4	0.00
25)	L5	AR-1248-5	6.546					6.546	E3	0.00
26)	L6	AR-1254-1	1.701					1.701	E4	0.00
27)	L6	AR-1254-2	1.462					1.462	E4	0.00
28)	L6	AR-1254-3	2.348					2.348	E4	0.00
29)	L6	AR-1254-4	1.813					1.813	E4	0.00
30)	L6	AR-1254-5	1.039					1.039	E4	0.00
31)	L7	AR-1260-1	1.533	1.226	1.369	1.532	1.768	1.486	E4	13.69
32)	L7	AR-1260-2	1.806	1.515	1.599	1.939	2.392	1.850	E4	18.71
33)	L7	AR-1260-3	1.648	1.428	1.456	1.726	1.982	1.648	E4	13.66
34)	L7	AR-1260-4	2.738	2.423	2.528	2.697	3.117	2.701	E4	9.83
35)	L7	AR-1260-5	1.747	1.701	1.608	1.756	2.044	1.771	E4	9.22
36)	L8	AR-1262-1	1.713					1.713	E4	0.00
37)	L8	AR-1262-2	1.632					1.632	E4	0.00
38)	L8	AR-1262-3	2.738					2.738	E4	0.00
39)	L8	AR-1262-4	2.177					2.177	E4	0.00
40)	L8	AR-1262-5	4.542					4.542	E4	0.00
41)	L9	AR-1268-1	4.546					4.546	E4	0.00
42)	L9	AR-1268-2	4.073					4.073	E4	0.00
43)	L9	AR-1268-3	3.446					3.446	E4	0.00
44)	L9	AR-1268-4	1.275					1.275	E4	0.00
45)	L9	AR-1268-5	9.086					9.086	E4	0.00

(#) = Out of Range