

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050919\
 Data File : P0056004.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 May 2019 12:14
 Operator : SM/SJ
 Sample : K2626-07
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 RI2-4(18-24)

Manual Integrations
 APPROVED

Ankita
 5/13/2019 4:14:02 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 05:00:06 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 15:20:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.272	3.587	931178	940346	31.116	33.294
2)	SA Decachlor...	9.860	8.532	608038	455201	16.250m	16.832

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050919\
Data File : PO056004.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 May 2019 12:14
Operator : SM/SJ
Sample : K2626-07
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
RI2-4(18-24)


Manual Integrations
APPROVED

Ankita
5/13/2019 4:14:02 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 05:00:06 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 15:20:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.272 min
Delta R.T.: -0.007 min
Response: 931178
Conc: 31.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
RI2-4(18-24)


Manual Integrations
APPROVED

Ankita
5/13/2019 4:14:02 PM




#1 Tetrachloro-m-xylene

R.T.: 3.587 min
Delta R.T.: -0.006 min
Response: 940346
Conc: 33.29 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.860 min
Delta R.T.: -0.019 min
Response: 608038
Conc: 16.25 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.532 min
Delta R.T.: -0.012 min
Response: 455201
Conc: 16.83 ng/ml