

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051019\
 Data File : P0056082.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 May 2019 20:19
 Operator : SM/SJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 5/14/2019 2:02:26 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 02:42:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 15:20:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.268	3.584	1972646	1840497	65.918	65.166
2)	SA Decachlor...	9.858	8.529	1879228	1386853	50.224	51.283
Target Compounds							
3)	L1 AR-1016-1	5.428	4.653	841506	764374	592.144m	568.768m
4)	L1 AR-1016-2	5.450	4.673	1165255	1039629	588.848m	569.364
5)	L1 AR-1016-3	5.511	4.846	727447	565029	567.547m	563.829m
6)	L1 AR-1016-4	5.610	4.886	619251	446893	592.456	549.255m
7)	L1 AR-1016-5	5.901	5.098	639089	571443	567.313	519.088m
31)	L7 AR-1260-1	7.019	6.120	1050818	1011698	486.604m	459.198m
32)	L7 AR-1260-2	7.275	6.307	1305749	1370907	498.590m	448.952m
33)	L7 AR-1260-3	7.631	6.460	1000447	1148759	500.226m	460.974m
34)	L7 AR-1260-4	7.856	6.928	1205894	948879	521.444	459.846
35)	L7 AR-1260-5	8.166	7.169	2107955	2305319	496.888	459.019

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051019\
Data File : PO056082.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 May 2019 20:19
Operator : SM/SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Ankita
5/14/2019 2:02:26 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 02:42:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 15:20:04 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

