

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051124\
 Data File : P0103631.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2024 10:15
 Operator : YP/AJ
 Sample : P2437-08 10X
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 WAREHOUSE-TOTE-WATER

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 22:29:24 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 07 04:55:35 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.448	3.586	6374534	3992792	0.828	1.585 #
2) SA Decachlor...	10.210	8.550	1045559	2195776	0.221	1.688 #

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051124\
Data File : PO103631.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2024 10:15
Operator : YP/AJ
Sample : P2437-08 10X
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
WAREHOUSE-TOTE-WATER

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 22:29:24 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 07 04:55:35 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.448 min
Delta R.T.: -0.003 min
Response: 6374534
Conc: 0.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
WAREHOUSE-TOTE-WATER




#1 Tetrachloro-m-xylene

R.T.: 3.586 min
Delta R.T.: -0.005 min
Response: 3992792
Conc: 1.58 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.210 min
Delta R.T.: 0.008 min
Response: 1045559
Conc: 0.22 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.550 min
Delta R.T.: -0.005 min
Response: 2195776
Conc: 1.69 ng/ml