

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0051518.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 15 15:44:07 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044676.D 1000 =P0044674.D 750 =P0044675.D
 250 =P0044677.D 50 =P0044678.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.548	1.412	1.537	1.481	1.684	1.532	E5	6.55
2) SA	Decachlorobiphenyl	2.061	2.013	2.025	1.997	2.184	2.056	E5	3.67
3) L1	AR-1016-1	4.042	3.711	3.998	3.608	4.158	3.903	E3	5.97
4) L1	AR-1016-2	5.600	5.181	5.533	5.139	5.219	5.334	E3	4.03
5) L1	AR-1016-3	4.460	4.189	4.446	3.933	3.993	4.204	E3	5.85
6) L1	AR-1016-4	4.571	4.355	4.629	4.150	4.370	4.415	E3	4.32
7) L1	AR-1016-5	4.006	3.786	4.024	3.687	4.000	3.901	E3	3.95
8) L2	AR-1221-1	3.052					3.052	E3	0.00
9) L2	AR-1221-2	2.224					2.224	E3	0.00
10) L2	AR-1221-3	7.351					7.351	E3	0.00
11) L3	AR-1232-1	3.296					3.296	E3	0.00
12) L3	AR-1232-2	2.352					2.352	E3	0.00
13) L3	AR-1232-3	4.345					4.345	E3	0.00
14) L3	AR-1232-4	2.701					2.701	E3	0.00
15) L3	AR-1232-5	2.069					2.069	E3	0.00
16) L4	AR-1242-1	2.948					2.948	E3	0.00
17) L4	AR-1242-2	6.963					6.963	E3	0.00
18) L4	AR-1242-3	3.412					3.412	E3	0.00
19) L4	AR-1242-4	3.280					3.280	E3	0.00
20) L4	AR-1242-5	3.634					3.634	E3	0.00
21) L5	AR-1248-1	5.441					5.441	E3	0.00
22) L5	AR-1248-2	4.764					4.764	E3	0.00
23) L5	AR-1248-3	6.722					6.722	E3	0.00
24) L5	AR-1248-4	7.020					7.020	E3	0.00
25) L5	AR-1248-5	4.392					4.392	E3	0.00
26) L6	AR-1254-1	1.080					1.080	E4	0.00
27) L6	AR-1254-2	7.274					7.274	E3	0.00
28) L6	AR-1254-3	1.240					1.240	E4	0.00
29) L6	AR-1254-4	1.032					1.032	E4	0.00
30) L6	AR-1254-5	7.718					7.718	E3	0.00
31) L7	AR-1260-1	0.951	0.891	0.931	0.892	1.014	0.936	E4	5.41
32) L7	AR-1260-2	1.275	1.219	1.253	1.211	1.380	1.268	E4	5.34
33) L7	AR-1260-3	1.068	1.026	1.046	1.012	1.111	1.052	E4	3.69
34) L7	AR-1260-4	2.419	2.395	2.399	2.262	2.381	2.371	E4	2.64
35) L7	AR-1260-5	1.481	1.455	1.467	1.394	1.500	1.459	E4	2.76
36) L8	AR-1262-1	7.862					7.862	E3	0.00
37) L8	AR-1262-2	1.053					1.053	E4	0.00
38) L8	AR-1262-3	1.004					1.004	E4	0.00
39) L8	AR-1262-4	1.397					1.397	E4	0.00
40) L8	AR-1262-5	3.046					3.046	E4	0.00
41) L9	AR-1268-1	3.302					3.302	E4	0.00
42) L9	AR-1268-2	2.987					2.987	E4	0.00
43) L9	AR-1268-3	2.625					2.625	E4	0.00
44) L9	AR-1268-4	1.110					1.110	E4	0.00
45) L9	AR-1268-5	7.709					7.709	E4	0.00

Signal #2 Calibration Files

500 =P0044676.D 1000 =P0044674.D 750 =P0044675.D
 250 =P0044677.D 50 =P0044678.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.478	2.344	2.469	2.261	2.435	2.397	E5	3.87
2) SA	Decachlorobiphenyl	2.798	2.800	2.783	2.775	2.912	2.814	E5	1.99

Method Path : P:\HPCHEM1\ECD_0\methods\
 Method File : P0051518.M
 Title : GC EXTRACTABLES
 Last Update : Tue May 15 15:44:07 2018
 Response Via : Initial Calibration

Calibration Files

500 =P0044676.D 1000 =P0044674.D 750 =P0044675.D
 250 =P0044677.D 50 =P0044678.D

Compound			500	1000	750	250	50	Avg		%RSD

3)	L1	AR-1016-1	6.160	6.156	6.129	6.041	6.253	6.148	E3	1.24
4)	L1	AR-1016-2	5.568	5.245	5.470	5.024	5.642	5.390	E3	4.70
5)	L1	AR-1016-3	6.773	6.410	6.676	6.043	6.011	6.382	E3	5.50
6)	L1	AR-1016-4	7.207	6.908	7.164	6.559	7.436	7.055	E3	4.74
7)	L1	AR-1016-5	7.988	7.497	7.782	7.205	6.426	7.380	E3	8.26
8)	L2	AR-1221-1	4.521					4.521	E3	0.00
9)	L2	AR-1221-2	3.253					3.253	E3	0.00
10)	L2	AR-1221-3	1.088					1.088	E4	0.00
11)	L3	AR-1232-1	4.811					4.811	E3	0.00
12)	L3	AR-1232-2	1.065					1.065	E4	0.00
13)	L3	AR-1232-3	4.357					4.357	E3	0.00
14)	L3	AR-1232-4	5.703					5.703	E3	0.00
15)	L3	AR-1232-5	2.839					2.839	E3	0.00
16)	L4	AR-1242-1	5.732					5.732	E3	0.00
17)	L4	AR-1242-2	1.834					1.834	E4	0.00
18)	L4	AR-1242-3	6.228					6.228	E3	0.00
19)	L4	AR-1242-4	5.910					5.910	E3	0.00
20)	L4	AR-1242-5	6.848					6.848	E3	0.00
21)	L5	AR-1248-1	9.304					9.304	E3	0.00
22)	L5	AR-1248-2	8.391					8.391	E3	0.00
23)	L5	AR-1248-3	1.430					1.430	E4	0.00
24)	L5	AR-1248-4	1.312					1.312	E4	0.00
25)	L5	AR-1248-5	7.120					7.120	E3	0.00
26)	L6	AR-1254-1	1.618					1.618	E4	0.00
27)	L6	AR-1254-2	1.409					1.409	E4	0.00
28)	L6	AR-1254-3	2.339					2.339	E4	0.00
29)	L6	AR-1254-4	1.867					1.867	E4	0.00
30)	L6	AR-1254-5	1.098					1.098	E4	0.00
31)	L7	AR-1260-1	1.638	1.548	1.587	1.508	1.652	1.586	E4	3.80
32)	L7	AR-1260-2	2.178	2.078	2.115	2.031	2.452	2.171	E4	7.65
33)	L7	AR-1260-3	2.175	2.125	2.151	1.998	2.292	2.148	E4	4.90
34)	L7	AR-1260-4	4.280	4.272	4.261	3.989	4.418	4.244	E4	3.69
35)	L7	AR-1260-5	2.939	2.889	2.899	2.840	2.975	2.909	E4	1.77
36)	L8	AR-1262-1	1.794					1.794	E4	0.00
37)	L8	AR-1262-2	1.685					1.685	E4	0.00
38)	L8	AR-1262-3	2.966					2.966	E4	0.00
39)	L8	AR-1262-4	2.443					2.443	E4	0.00
40)	L8	AR-1262-5	5.461					5.461	E4	0.00
41)	L9	AR-1268-1	5.829					5.829	E4	0.00
42)	L9	AR-1268-2	5.393					5.393	E4	0.00
43)	L9	AR-1268-3	4.720					4.720	E4	0.00
44)	L9	AR-1268-4	1.875					1.875	E4	0.00
45)	L9	AR-1268-5	1.320					1.320	E5	0.00

(#) = Out of Range