

Data Path : P:\HPCHEM1\ECD 0\Data\P0051518\
 Data File : P0044692.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 May 2018 17:05
 Operator : SM/UA
 Sample : J2868-05
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PS-RO-327

Manual Integrations
 APPROVED

ugo
 5/17/2018 10:19:29 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 15 18:13:39 2018
 Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 15 15:44:07 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	nq/ml	nq/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.031	3.296	3061030	4314841	19.975m	17.999m
2) SA Decachlor...	9.347	8.037	2153918	2694951	10.476	9.577

Target Compounds

 (f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : P:\HPCHEM1\ECD 0\Data\P0051518\
Data File : P0044692.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 May 2018 17:05
Operator : SM/UA
Sample : J2868-05
Misc :
ALS Vial : 22 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
PS-RO-327


Manual Integrations
APPROVED

ugo
5/17/2018 10:19:29 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 15 18:13:39 2018
Quant Method : P:\HPCHEM1\ECD_0\methods\P0051518.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 15 15:44:07 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.031 min
Delta R.T.: -0.002 min
Response: 3061030
Conc: 19.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
PS-RO-327


Manual Integrations
APPROVED

ugo
5/17/2018 10:19:29 AM




#1 Tetrachloro-m-xylene

R.T.: 3.296 min
Delta R.T.: -0.002 min
Response: 4314841
Conc: 18.00 ng/ml m



#2 Decachlorobiphenyl

R.T.: 9.347 min
Delta R.T.: -0.003 min
Response: 2153918
Conc: 10.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.037 min
Delta R.T.: -0.001 min
Response: 2694951
Conc: 9.58 ng/ml