

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\
 Method File : P0051719.M
 Title : GC EXTRACTABLES
 Last Update : Fri May 17 23:42:43 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0056307.D 1000 =P0056305.D 750 =P0056306.D
 250 =P0056308.D 50 =P0056309.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.790	4.345	4.719	4.606	4.747	4.641	E4	3.85
2) SA	Decachlorobiphenyl	4.733	4.401	4.510	4.907	5.464	4.803	E4	8.71
3) L1	AR-1016-1	2.038	1.793	1.934	2.111	2.428	2.061	E3	11.53
4) L1	AR-1016-2	2.922	2.604	2.767	2.961	3.321	2.915	E3	9.17
5) L1	AR-1016-3	1.829	1.618	1.720	1.880	2.128	1.835	E3	10.49
6) L1	AR-1016-4	1.511	1.349	1.428	1.541	1.715	1.509	E3	9.10
7) L1	AR-1016-5	1.545	1.362	1.450	1.594	1.823	1.555	E3	11.22
8) L2	AR-1221-1	5.868					5.868	E2	0.00
9) L2	AR-1221-2	4.506					4.506	E2	0.00
10) L2	AR-1221-3	1.374					1.374	E3	0.00
11) L3	AR-1232-1	1.712					1.712	E3	0.00
12) L3	AR-1232-2	9.701					9.701	E2	0.00
13) L3	AR-1232-3	1.973					1.973	E3	0.00
14) L3	AR-1232-4	1.013					1.013	E3	0.00
15) L3	AR-1232-5	8.012					8.012	E2	0.00
16) L4	AR-1242-1	1.614					1.614	E3	0.00
17) L4	AR-1242-2	2.268					2.268	E3	0.00
18) L4	AR-1242-3	1.451					1.451	E3	0.00
19) L4	AR-1242-4	1.194					1.194	E3	0.00
20) L4	AR-1242-5	1.363					1.363	E3	0.00
21) L5	AR-1248-1	1.256					1.256	E3	0.00
22) L5	AR-1248-2	1.767					1.767	E3	0.00
23) L5	AR-1248-3	1.944					1.944	E3	0.00
24) L5	AR-1248-4	2.189					2.189	E3	0.00
25) L5	AR-1248-5	2.101					2.101	E3	0.00
26) L6	AR-1254-1	2.299					2.299	E3	0.00
27) L6	AR-1254-2	3.586					3.586	E3	0.00
28) L6	AR-1254-3	3.822					3.822	E3	0.00
29) L6	AR-1254-4	2.819					2.819	E3	0.00
30) L6	AR-1254-5	3.030					3.030	E3	0.00
31) L7	AR-1260-1	2.720	2.518	2.607	2.831	3.277	2.791	E3	10.63
32) L7	AR-1260-2	3.218	3.005	3.091	3.328	3.831	3.295	E3	9.83
33) L7	AR-1260-3	2.496	2.331	2.385	2.583	2.849	2.529	E3	8.07
34) L7	AR-1260-4	2.790	2.648	2.688	2.852	3.132	2.822	E3	6.78
35) L7	AR-1260-5	5.113	5.007	5.012	5.119	5.461	5.142	E3	3.61
36) L8	AR-1262-1	3.625					3.625	E3	0.00
37) L8	AR-1262-2	5.885					5.885	E3	0.00
38) L8	AR-1262-3	4.054					4.054	E3	0.00
39) L8	AR-1262-4	3.066					3.066	E3	0.00
40) L8	AR-1262-5	2.022					2.022	E3	0.00
41) L9	AR-1268-1	6.738					6.738	E3	0.00
42) L9	AR-1268-2	6.148					6.148	E3	0.00
43) L9	AR-1268-3	5.215					5.215	E3	0.00
44) L9	AR-1268-4	2.183					2.183	E3	0.00
45) L9	AR-1268-5	1.523					1.523	E4	0.00

Signal #2 Calibration Files

500 =P0056307.D 1000 =P0056305.D 750 =P0056306.D
 250 =P0056308.D 50 =P0056309.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.323	3.865	4.234	4.148	4.442	4.202	E4	5.18
2) SA	Decachlorobiphenyl	4.030	3.924	3.924	4.024	4.403	4.061	E4	4.88
3) L1	AR-1016-1	1.744	1.543	1.652	1.803	2.007	1.750	E3	9.95
4) L1	AR-1016-2	2.444	2.159	2.330	2.472	2.732	2.427	E3	8.66
5) L1	AR-1016-3	1.347	1.186	1.277	1.375	1.473	1.332	E3	8.06

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0051719.M
 Title : GC EXTRACTABLES
 Last Update : Fri May 17 23:42:43 2019
 Response Via : Initial Calibration

Calibration Files

500 =P0056307.D 1000 =P0056305.D 750 =P0056306.D
 250 =P0056308.D 50 =P0056309.D

	Compound	500	1000	750	250	50	Avg		%RSD

6)	L1 AR-1016-4	1.091	0.951	1.023	1.121	1.252	1.088	E3	10.40
7)	L1 AR-1016-5	1.428	1.265	1.352	1.469	1.567	1.416	E3	8.08
8)	L2 AR-1221-1	5.098					5.098	E2	0.00
9)	L2 AR-1221-2	3.789					3.789	E2	0.00
10)	L2 AR-1221-3	1.181					1.181	E3	0.00
11)	L3 AR-1232-1	1.473					1.473	E3	0.00
12)	L3 AR-1232-2	1.667					1.667	E3	0.00
13)	L3 AR-1232-3	9.051					9.051	E2	0.00
14)	L3 AR-1232-4	7.886					7.886	E2	0.00
15)	L3 AR-1232-5	8.900					8.900	E2	0.00
16)	L4 AR-1242-1	1.347					1.347	E3	0.00
17)	L4 AR-1242-2	1.888					1.888	E3	0.00
18)	L4 AR-1242-3	1.048					1.048	E3	0.00
19)	L4 AR-1242-4	1.014					1.014	E3	0.00
20)	L4 AR-1242-5	1.263					1.263	E3	0.00
21)	L5 AR-1248-1	1.073					1.073	E3	0.00
22)	L5 AR-1248-2	1.404					1.404	E3	0.00
23)	L5 AR-1248-3	1.436					1.436	E3	0.00
24)	L5 AR-1248-4	1.777					1.777	E3	0.00
25)	L5 AR-1248-5	1.706					1.706	E3	0.00
26)	L6 AR-1254-1	2.826					2.826	E3	0.00
27)	L6 AR-1254-2	2.461					2.461	E3	0.00
28)	L6 AR-1254-3	4.029					4.029	E3	0.00
29)	L6 AR-1254-4	2.495					2.495	E3	0.00
30)	L6 AR-1254-5	3.581					3.581	E3	0.00
31)	L7 AR-1260-1	2.583	2.392	2.466	2.649	3.001	2.618	E3	9.02
32)	L7 AR-1260-2	3.202	3.026	3.095	3.282	3.837	3.288	E3	9.79
33)	L7 AR-1260-3	2.963	2.792	2.850	2.995	3.306	2.981	E3	6.70
34)	L7 AR-1260-4	2.479	2.351	2.396	2.504	2.657	2.477	E3	4.75
35)	L7 AR-1260-5	5.845	5.737	5.755	5.664	6.111	5.822	E3	2.98
36)	L8 AR-1262-1	3.952					3.952	E3	0.00
37)	L8 AR-1262-2	6.470					6.470	E3	0.00
38)	L8 AR-1262-3	2.737					2.737	E3	0.00
39)	L8 AR-1262-4	4.649					4.649	E3	0.00
40)	L8 AR-1262-5	2.081					2.081	E3	0.00
41)	L9 AR-1268-1	7.635					7.635	E3	0.00
42)	L9 AR-1268-2	6.686					6.686	E3	0.00
43)	L9 AR-1268-3	5.824					5.824	E3	0.00
44)	L9 AR-1268-4	2.240					2.240	E3	0.00
45)	L9 AR-1268-5	1.549					1.549	E4	0.00

 (#) = Out of Range